

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle p582,575,780,q499,488,678,c622

Collected By: NAKHLA/ODOMIRO

Comment:

Sample HILLVIEW BX

Sample 3

ID of Source:

Collection 11/9/2001 AT 9:52:00

Submitted 11/9/2001 AT 10:49:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD27544

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
<u>Inorganics</u>						
EPA94	Silver	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	12/10/2001	PAD
SM18	Arsenic	< MDL	ug/L	20	11/12/2001	PAD
EPA94	Barium	< MDL	ug/L	200.0	12/10/2001	PAD
EPA94	Beryllium	< MDL	ug/L	2	12/10/2001	PAD
EPA94	Calcium	5900	ug/L	1000.0	12/10/2001	PAD
EPA94	Cadmium	< MDL	ug/L	2.0	12/10/2001	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94	Chromium	< MDL	ug/L	10.0	12/10/2001	PAD
EPA94	Copper	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94	Iron	28	ug/L	20.0	12/10/2001	PAD
EPA79	Mercury	< MDL	ug/L	0.40	11/14/2001	PAD
SM18	Potassium	520	ug/L	100.0	11/19/2001	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 1/22/2002

Environmental Laboratories

emailed

Original 1/22/2002

NYS ELAP # 10108

Page 1 of 4

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle p582,575,780,q499,488,678,c622

Collected By: NAKHLA/ODOMIRO

Comment:

Sample HILLVIEW BX

Sample 3

ID of Source:

Collection 11/9/2001 AT 9:52:00

Submitted 11/9/2001 AT 10:49:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD27544

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Magnesium	1300	ug/L	1000.0	12/10/2001	PAD
EPA94	Manganese	25	ug/L	20.0	12/10/2001	PAD
EPA94	Molybdenum	< MDL	ug/L	50.0	12/10/2001	PAD
EPA94	Sodium	7200	ug/L	1000.0	12/10/2001	PAD
EPA94	Nickel	< MDL	ug/L	20.0	12/10/2001	PAD
SM18	Lead	< MDL	ug/L	1.0	11/28/2001	PAD
EPA'79	Potable Metals Digestion	Completed			11/15/2001	PAD
EPA94	Antimony	< MDL	ug/L	0.1	12/13/2001	PAD
SM18	Selenium	< MDL	ug/L	2	12/12/2001	PAD
SM18	Tin	< MDL	ug/L	100	12/21/2001	PAD
SM18	Titanium	< MDL	ug/L	100	1/18/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	12/14/2001	PAD
EPA94	Vanadium	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94	Zinc	< MDL	ug/L	20.0	12/10/2001	PAD

Approved Pam Dilisizian

QA Officer

Date Approved: 1/22/2002

Environmental Laboratories

emailed

Original 1/22/2002

NYS ELAP # 10108

Page 2 of 4

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle p582,575,780,q499,488,678,c622

Collected By: NAKHLA/ODOMIRO

Comment:

Sample HILLVIEW BX

Sample 3

ID of Source:

Collection 11/9/2001 AT 9:52:00

Submitted 11/9/2001 AT 10:49:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD27544

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Organics						
Chlorine Pesticides - 508						
	4,4-DDE	< MDL	ug/L	0.080	11/15/2001	DLV
	Aldrin	< MDL	ug/L	0.010	11/15/2001	DLV
	Chlordane	< MDL	ug/L	0.40	11/15/2001	DLV
	Dieldrin	< MDL	ug/L	0.016	11/15/2001	DLV
	Endrin	< MDL	ug/L	0.020	11/15/2001	DLV
	Heptachlor	< MDL	ug/L	0.020	11/15/2001	DLV
	Heptachlor Epoxide	< MDL	ug/L	0.030	11/15/2001	DLV
	Lindane	< MDL	ug/L	0.030	11/15/2001	DLV
	Methoxychlor	< MDL	ug/L	0.039	11/15/2001	DLV
	PCB's	< MDL	ug/L	0.20	11/15/2001	DLV
	Propachlor	< MDL	ug/L	0.040	11/15/2001	DLV
	Toxaphene	< MDL	ug/L	0.40	11/15/2001	DLV
The recoveries for 4 of the 11 compounds in the LCS Standard were below their acceptance ranges.						
Organic Chemicals - 525.2						
	2,4-Dinitrotoluene	< MDL	ug/L	2.0	11/19/2001	DLV
	2,6-Dinitrotoluene	< MDL	ug/L	2.0	11/19/2001	DLV
	4,4-DDE	< MDL	ug/L	0.80	11/19/2001	DLV
	Acetochlor	< MDL	ug/L	2.0	11/19/2001	DLV
	Alachlor	< MDL	ug/L	0.20	11/19/2001	DLV
	Atrazine	< MDL	ug/L	0.20	11/19/2001	DLV

Approved Pam Dilsizian

QA Officer

Date Approved: 1/22/2002

Environmental Laboratories

emailed

Original 1/22/2002

NYS ELAP # 10108

Page 3 of 4

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle p582,575,780,q499,488,678,c622

Collected By: NAKHLA/ODOMIRO

Comment:

Sample HILLVIEW BX

Sample 3

ID of Source:

Collection 11/9/2001 AT 9:52:00

Submitted 11/9/2001 AT 10:49:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD27544

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
	Benzo(a)pyrene	< MDL	ug/L	0.20	11/19/2001	DLV
	bis(2-Ethylhexyl)adipate	< MDL	ug/L	0.60	11/19/2001	DLV
	bis(2-Ethylhexyl)phthalate	2.7	ug/L	0.60	11/19/2001	DLV
	Butachlor	< MDL	ug/L	0.40	11/19/2001	DLV
	EPTC	< MDL	ug/L	1.0	11/19/2001	DLV
	Hexachlorobenzene	< MDL	ug/L	0.10	11/19/2001	DLV
	Hexachlorocyclopentadiene	< MDL	ug/L	0.10	11/19/2001	DLV
	Metolachlor	< MDL	ug/L	0.75	11/19/2001	DLV
	Metribuzin	< MDL	ug/L	0.15	11/19/2001	DLV
	Molinate	< MDL	ug/L	0.90	11/19/2001	DLV
	Simazine	< MDL	ug/L	0.40	11/19/2001	DLV
	Terbacil	< MDL	ug/L	2.0	11/19/2001	DLV

The recoveries for 6 of the 18 compounds in the LCS Standard were outside of their acceptance ranges.

GCMS Scan For Unknowns

Other unknown compounds	.	ug/L	12/18/2001	DLV
-------------------------	---	------	------------	-----

A scan of the extract prepared for the 508 analysis, from 35-550 amu using a mass spectrometer, does not reveal the presence of non-target compounds.

Approved Pam Dilsizian

QA Officer

Date Approved: 1/22/2002

Environmental Laboratories

emailed

Original 1/22/2002

NYS ELAP # 10108

Page 4 of 4

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle c607,p776,767,348,q117,653,654

Collected By: NAKHLA/ODOMIRO

Comment:

Sample HILLVIEW BX

Sample 58

ID of Source:

Collection 11/9/2001 AT 9:30:00

Submitted 11/9/2001 AT 10:49:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD27545

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
<u>Inorganics</u>						
EPA94	Silver	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	12/10/2001	PAD
SM18	Arsenic	< MDL	ug/L	20	11/12/2001	PAD
EPA94	Barium	< MDL	ug/L	200.0	12/10/2001	PAD
EPA94	Beryllium	< MDL	ug/L	2	12/10/2001	PAD
EPA94	Calcium	5900	ug/L	1000.0	12/10/2001	PAD
EPA94	Cadmium	< MDL	ug/L	2.0	12/10/2001	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94	Chromium	< MDL	ug/L	10.0	12/10/2001	PAD
EPA94	Copper	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94	Iron	25	ug/L	20.0	12/10/2001	PAD
EPA79	Mercury	< MDL	ug/L	0.40	11/14/2001	PAD
SM18	Potassium	520	ug/L	100.0	11/19/2001	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 1/22/2002

Environmental Laboratories

emailed

Original 1/22/2002

NYS ELAP # 10108

Page 1 of 4

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle c607,p776,767,348,q117,653,654

Collected By: NAKHLA/ODOMIRO

Comment:

Sample HILLVIEW BX

Sample 58

ID of Source:

Collection 11/9/2001 AT 9:30:00

Submitted 11/9/2001 AT 10:49:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD27545

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Magnesium	1200	ug/L	1000.0	12/10/2001	PAD
EPA94	Manganese	24	ug/L	20.0	12/10/2001	PAD
EPA94	Molybdenum	< MDL	ug/L	50.0	12/10/2001	PAD
EPA94	Sodium	6800	ug/L	1000.0	12/10/2001	PAD
EPA94	Nickel	< MDL	ug/L	20.0	12/10/2001	PAD
SM18	Lead	< MDL	ug/L	1.0	11/28/2001	PAD
EPA79	Potable Metals Digestion	Completed			11/15/2001	PAD
EPA94	Antimony	< MDL	ug/L	0.1	12/13/2001	PAD
SM18	Selenium	< MDL	ug/L	2	12/12/2001	PAD
SM18	Tin	< MDL	ug/L	100	12/21/2001	PAD
SM18	Titanium	< MDL	ug/L	100	1/18/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	12/14/2001	PAD
EPA94	Vanadium	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94	Zinc	< MDL	ug/L	20.0	12/10/2001	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 1/22/2002

Environmental Laboratories

emailed

Original 1/22/2002

NYS ELAP # 10108

Page 2 of 4

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle c607,p776,767,348,q117,653,654

Collected By: NAKHLA/ODOMIRO

Comment:

Sample HILLVIEW BX

Sample 58

ID of Source:

Collection 11/9/2001 AT 9:30:00

Submitted 11/9/2001 AT 10:49:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD27545

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Organics						
Chlorine Pesticides - 508						
	4,4-DDE	< MDL	ug/L	0.080	11/15/2001	DLV
	Aldrin	< MDL	ug/L	0.010	11/15/2001	DLV
	Chlordane	< MDL	ug/L	0.40	11/15/2001	DLV
	Dieldrin	< MDL	ug/L	0.016	11/15/2001	DLV
	Endrin	< MDL	ug/L	0.020	11/15/2001	DLV
	Heptachlor	< MDL	ug/L	0.020	11/15/2001	DLV
	Heptachlor Epoxide	< MDL	ug/L	0.030	11/15/2001	DLV
	Lindane	< MDL	ug/L	0.030	11/15/2001	DLV
	Methoxychlor	< MDL	ug/L	0.039	11/15/2001	DLV
	PCB's	< MDL	ug/L	0.20	11/15/2001	DLV
	Propachlor	< MDL	ug/L	0.040	11/15/2001	DLV
	Toxaphene	< MDL	ug/L	0.40	11/15/2001	DLV
The recoveries for 4 of the 11 compounds in the LCS Standard were below their acceptance ranges.						
Organic Chemicals - 525.2						
	2,4-Dinitrotoluene	< MDL	ug/L	2.0	11/19/2001	DLV
	2,6-Dinitrotoluene	< MDL	ug/L	2.0	11/19/2001	DLV
	4,4-DDE	< MDL	ug/L	0.80	11/19/2001	DLV
	Acetochlor	< MDL	ug/L	2.0	11/19/2001	DLV
	Alachlor	< MDL	ug/L	0.20	11/19/2001	DLV
	Atrazine	< MDL	ug/L	0.20	11/19/2001	DLV

Approved Pam Dilsizian

QA Officer

Date Approved: 1/22/2002

Environmental Laboratories

emailed

Original 1/22/2002

NYS ELAP # 10108

Page 3 of 4

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle c607,p776,767,348,q117,653,654

Collected By: NAKHLA/ODOMIRO

Comment:

Sample HILLVIEW BX

Sample 58

ID of Source:

Collection 11/9/2001 AT 9:30:00

Submitted 11/9/2001 AT 10:49:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD27545

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
	Benzo(a)pyrene	< MDL	ug/L	0.20	11/19/2001	DLV
	bis(2-Ethylhexyl)adipate	< MDL	ug/L	0.60	11/19/2001	DLV
	bis(2-Ethylhexyl)phthalate	0.68	ug/L	0.60	11/19/2001	DLV
	Butachlor	< MDL	ug/L	0.40	11/19/2001	DLV
	EPTC	< MDL	ug/L	1.0	11/19/2001	DLV
	Hexachlorobenzene	< MDL	ug/L	0.10	11/19/2001	DLV
	Hexachlorocyclopentadiene	< MDL	ug/L	0.10	11/19/2001	DLV
	Metolachlor	< MDL	ug/L	0.75	11/19/2001	DLV
	Metribuzin	< MDL	ug/L	0.15	11/19/2001	DLV
	Molinate	< MDL	ug/L	0.90	11/19/2001	DLV
	Simazine	< MDL	ug/L	0.40	11/19/2001	DLV
	Terbacil	< MDL	ug/L	2.0	11/19/2001	DLV

The recoveries for 6 of the 18 compounds in the LCS Standard were outside of their acceptance ranges.

GCMS Scan For Unknowns

Other unknown compounds	.	ug/L	12/18/2001	DLV
-------------------------	---	------	------------	-----

A scan of the extract prepared for the 508 analysis, from 35-550 amu using a mass spectrometer, does not reveal the presence of non-target compounds.

Approved Pam Dilsizian

QA Officer

Date Approved: 1/22/2002

Environmental Laboratories

emailed

Original 1/22/2002

NYS ELAP # 10108

Page 4 of 4

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
Received EP
Bottle c488,p727,699,711,q626,614,613
Collected By: PATEL
Comment:

Sample BRONX

Sample 17750
ID of Source:
Collection 11/8/2001 AT 9:00:00
Submitted 11/9/2001 AT 10:49:00
Sample Type: POT_DW
PWS No.: **Source Code:**

Type Descriptor:
Sample chilled on arrival? : yes

Report

Sample No. AD27546

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
<u>Inorganics</u>						
EPA94	Silver	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	12/10/2001	PAD
SM18	Arsenic	< MDL	ug/L	20	11/12/2001	PAD
EPA94	Barium	< MDL	ug/L	200.0	12/10/2001	PAD
EPA94	Beryllium	< MDL	ug/L	2	12/10/2001	PAD
EPA94	Calcium	5900	ug/L	1000.0	12/10/2001	PAD
EPA94	Cadmium	< MDL	ug/L	2.0	12/10/2001	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94	Chromium	< MDL	ug/L	10.0	12/10/2001	PAD
EPA94	Copper	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94	Iron	32	ug/L	20.0	12/10/2001	PAD
EPA79	Mercury	< MDL	ug/L	0.40	11/14/2001	PAD
SM18	Potassium	520	ug/L	100.0	11/19/2001	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 1/22/2002

Environmental Laboratories

emailed

Original 1/22/2002

NYS ELAP # 10108

Page 1 of 4

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle c488,p727,699,711,q626,614,613

Collected By: PATEL

Comment:

Sample BRONX

Sample 17750

ID of Source:

Collection 11/8/2001 AT 9:00:00

Submitted 11/9/2001 AT 10:49:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD27546

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Magnesium	1300	ug/L	1000.0	12/10/2001	PAD
EPA94	Manganese	27	ug/L	20.0	12/10/2001	PAD
EPA94	Molybdenum	< MDL	ug/L	50.0	12/10/2001	PAD
EPA94	Sodium	6300	ug/L	1000.0	12/10/2001	PAD
EPA94	Nickel	< MDL	ug/L	20.0	12/10/2001	PAD
SM18	Lead	< MDL	ug/L	1.0	11/28/2001	PAD
EPA79	Potable Metals Digestion	Completed			11/15/2001	PAD
EPA94	Antimony	< MDL	ug/L	0.1	12/13/2001	PAD
SM18	Selenium	< MDL	ug/L	2	12/12/2001	PAD
SM18	Tin	< MDL	ug/L	100	12/21/2001	PAD
SM18	Titanium	< MDL	ug/L	100	1/18/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	12/14/2001	PAD
EPA94	Vanadium	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94	Zinc	< MDL	ug/L	20.0	12/10/2001	PAD

Approved Pam Dilisizian

QA Officer

Date Approved: 1/22/2002

Environmental Laboratories

emailed

Original 1/22/2002

NYS ELAP # 10108

Page 2 of 4

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Sample BRONX

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle c488,p727,699,711,q626,614,613

Collected By: PATEL

Comment:

Sample 17750

ID of Source:

Collection 11/8/2001 AT 9:00:00

Submitted 11/9/2001 AT 10:49:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD27546

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
<u>Organics</u>						
<u>Chlorine Pesticides - 508</u>						
	4,4-DDE	< MDL	ug/L	0.080	11/15/2001	DLV
	Aldrin	< MDL	ug/L	0.010	11/15/2001	DLV
	Chlordane	< MDL	ug/L	0.40	11/15/2001	DLV
	Dieldrin	< MDL	ug/L	0.016	11/15/2001	DLV
	Endrin	< MDL	ug/L	0.020	11/15/2001	DLV
	Heptachlor	< MDL	ug/L	0.020	11/15/2001	DLV
	Heptachlor Epoxide	< MDL	ug/L	0.030	11/15/2001	DLV
	Lindane	< MDL	ug/L	0.030	11/15/2001	DLV
	Methoxychlor	< MDL	ug/L	0.039	11/15/2001	DLV
	PCB's	< MDL	ug/L	0.20	11/15/2001	DLV
	Propachlor	< MDL	ug/L	0.040	11/15/2001	DLV
	Toxaphene	< MDL	ug/L	0.40	11/15/2001	DLV
The recoveries for 4 of the 11 compounds in the LCS Standard were below their acceptance ranges.						
<u>Organic Chemicals - 525.2</u>						
	2,4-Dinitrotoluene	< MDL	ug/L	2.0	11/19/2001	DLV
	2,6-Dinitrotoluene	< MDL	ug/L	2.0	11/19/2001	DLV
	4,4-DDE	< MDL	ug/L	0.80	11/19/2001	DLV
	Acetochlor	< MDL	ug/L	2.0	11/19/2001	DLV
	Alachlor	< MDL	ug/L	0.20	11/19/2001	DLV
	Atrazine	< MDL	ug/L	0.20	11/19/2001	DLV

Approved Pam Dilsizian

QA Officer

Date Approved: 1/22/2002

Environmental Laboratories

emailed

Original 1/22/2002

NYS ELAP # 10108

Page 3 of 4

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Sample BRONX

Distribution Laboratory
96-05 Horace Harding Ex.
Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle c488,p727,699,711,q626,614,613

Collected By: PATEL

Comment:

Sample 17750

ID of Source:

Collection 11/8/2001 AT 9:00:00

Submitted 11/9/2001 AT 10:49:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD27546

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
	Benzo(a)pyrene	< MDL	ug/L	0.20	11/19/2001	DLV
	bis(2-Ethylhexyl)adipate	< MDL	ug/L	0.60	11/19/2001	DLV
	bis(2-Ethylhexyl)phthalate	0.89	ug/L	0.60	11/19/2001	DLV
	Butachlor	< MDL	ug/L	0.40	11/19/2001	DLV
	EPTC	< MDL	ug/L	1.0	11/19/2001	DLV
	Hexachlorobenzene	< MDL	ug/L	0.10	11/19/2001	DLV
	Hexachlorocyclopentadiene	< MDL	ug/L	0.10	11/19/2001	DLV
	Metolachlor	< MDL	ug/L	0.75	11/19/2001	DLV
	Metribuzin	< MDL	ug/L	0.15	11/19/2001	DLV
	Molinate	< MDL	ug/L	0.90	11/19/2001	DLV
	Simazine	< MDL	ug/L	0.40	11/19/2001	DLV
	Terbacil	< MDL	ug/L	2.0	11/19/2001	DLV

The recoveries for 6 of the 18 compounds in the LCS Standard were outside of their acceptance ranges.

GCMS Scan For Unknowns

Other unknown compounds	.	ug/L	12/18/2001	DLV
-------------------------	---	------	------------	-----

A scan of the extract prepared for the 508 analysis, from 35-550 amu using a mass spectrometer, does not reveal the presence of non-target compounds.

Approved Pam Dilsizian

QA Officer

Date Approved: 1/22/2002

Environmental Laboratories

emailed

Original 1/22/2002

NYS ELAP # 10108

Page 4 of 4

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle c490,p721,623,710,q708,611,535

Collected By: SPERANZA

Comment:

Sample BROOKLYN

Sample 27550

ID of Source:

Collection 11/8/2001 AT 9:12:00

Submitted 11/9/2001 AT 10:49:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD27547

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
<u>Inorganics</u>						
EPA94	Silver	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	12/10/2001	PAD
SM18	Arsenic	< MDL	ug/L	20	11/12/2001	PAD
EPA94	Barium	< MDL	ug/L	200.0	12/10/2001	PAD
EPA94	Beryllium	< MDL	ug/L	2	12/10/2001	PAD
EPA94	Calcium	6000	ug/L	1000.0	12/10/2001	PAD
EPA94	Cadmium	< MDL	ug/L	2.0	12/10/2001	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94	Chromium	< MDL	ug/L	10.0	12/10/2001	PAD
EPA94	Copper	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94	Iron	30	ug/L	20.0	12/10/2001	PAD
EPA79	Mercury	< MDL	ug/L	0.40	11/14/2001	PAD
SM18	Potassium	530	ug/L	100.0	11/19/2001	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 1/22/2002

Environmental Laboratories

emailed

Original 1/22/2002

NYS ELAP # 10108

Page 1 of 4

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Sample BROOKLYN

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle c490,p721,623,710,q708,611,535

Collected By: SPERANZA

Comment:

Sample 27550

ID of Source:

Collection 11/8/2001 AT 9:12:00

Submitted 11/9/2001 AT 10:49:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD27547

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Magnesium	1300	ug/L	1000.0	12/10/2001	PAD
EPA94	Manganese	21	ug/L	20.0	12/10/2001	PAD
EPA94	Molybdenum	< MDL	ug/L	50.0	12/10/2001	PAD
EPA94	Sodium	6900	ug/L	1000.0	12/10/2001	PAD
EPA94	Nickel	< MDL	ug/L	20.0	12/10/2001	PAD
SM18	Lead	< MDL	ug/L	1.0	11/28/2001	PAD
EPA'79	Potable Metals Digestion	Completed			11/15/2001	PAD
EPA94	Antimony	< MDL	ug/L	0.1	12/13/2001	PAD
SM18	Selenium	< MDL	ug/L	2	12/12/2001	PAD
SM18	Tin	< MDL	ug/L	100	12/21/2001	PAD
SM18	Titanium	< MDL	ug/L	100	1/18/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	12/14/2001	PAD
EPA94	Vanadium	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94	Zinc	< MDL	ug/L	20.0	12/10/2001	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 1/22/2002

Environmental Laboratories

emailed

Original 1/22/2002

NYS ELAP # 10108

Page 2 of 4

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle c490,p721,623,710,q708,611,535

Collected By: SPERANZA

Comment:

Sample BROOKLYN

Sample 27550

ID of Source:

Collection 11/8/2001 AT 9:12:00

Submitted 11/9/2001 AT 10:49:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD27547

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Organics						
Chlorine Pesticides - 508						
	4,4-DDE	< MDL	ug/L	0.080	11/15/2001	DLV
	Aldrin	< MDL	ug/L	0.010	11/15/2001	DLV
	Chlordane	< MDL	ug/L	0.40	11/15/2001	DLV
	Dieldrin	< MDL	ug/L	0.016	11/15/2001	DLV
	Endrin	< MDL	ug/L	0.020	11/15/2001	DLV
	Heptachlor	< MDL	ug/L	0.020	11/15/2001	DLV
	Heptachlor Epoxide	< MDL	ug/L	0.030	11/15/2001	DLV
	Lindane	< MDL	ug/L	0.030	11/15/2001	DLV
	Methoxychlor	< MDL	ug/L	0.039	11/15/2001	DLV
	PCB's	< MDL	ug/L	0.20	11/15/2001	DLV
	Propachlor	< MDL	ug/L	0.040	11/15/2001	DLV
	Toxaphene	< MDL	ug/L	0.40	11/15/2001	DLV
The recoveries for 4 of the 11 compounds in the LCS Standard were below their acceptance ranges.						
Organic Chemicals - 525.2						
	2,4-Dinitrotoluene	< MDL	ug/L	2.0	11/16/2001	DLV
	2,6-Dinitrotoluene	< MDL	ug/L	2.0	11/16/2001	DLV
	4,4-DDE	< MDL	ug/L	0.80	11/16/2001	DLV
	Acetochlor	< MDL	ug/L	2.0	11/16/2001	DLV
	Alachlor	< MDL	ug/L	0.20	11/16/2001	DLV
	Atrazine	< MDL	ug/L	0.20	11/16/2001	DLV

Approved Pam Dilsizian

QA Officer

Date Approved: 1/22/2002

Environmental Laboratories

emailed

Original 1/22/2002

NYS ELAP # 10108

Page 3 of 4

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Sample BROOKLYN

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle c490,p721,623,710,q708,611,535

Collected By: SPERANZA

Comment:

Sample 27550

ID of Source:

Collection 11/8/2001 AT 9:12:00

Submitted 11/9/2001 AT 10:49:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD27547

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
	Benzo(a)pyrene	< MDL	ug/L	0.20	11/16/2001	DLV
	bis(2-Ethylhexyl)adipate	< MDL	ug/L	0.60	11/16/2001	DLV
	bis(2-Ethylhexyl)phthalate	< MDL	ug/L	0.60	11/16/2001	DLV
	Butachlor	< MDL	ug/L	0.40	11/16/2001	DLV
	EPTC	< MDL	ug/L	1.0	11/16/2001	DLV
	Hexachlorobenzene	< MDL	ug/L	0.10	11/16/2001	DLV
	Hexachlorocyclopentadiene	< MDL	ug/L	0.10	11/16/2001	DLV
	Metolachlor	< MDL	ug/L	0.75	11/16/2001	DLV
	Metribuzin	< MDL	ug/L	0.15	11/16/2001	DLV
	Molinate	< MDL	ug/L	0.90	11/16/2001	DLV
	Simazine	< MDL	ug/L	0.40	11/16/2001	DLV
	Terbacil	< MDL	ug/L	2.0	11/16/2001	DLV

Recoveries for 4 of the 18 compounds in the MDL and LCS Standard and the Spiked Sample were outside of their acceptance ranges.

GCMS Scan For Unknowns

Other unknown compounds	.	ug/L	12/18/2001	DLV
-------------------------	---	------	------------	-----

A scan of the extract prepared for the 508 analysis, from 35-550 amu using a mass spectrometer, does not reveal the presence of non-target compounds.

Approved Pam Dilsizian

QA Officer

Date Approved: 1/22/2002

Environmental Laboratories

emailed

Original 1/22/2002

NYS ELAP # 10108

Page 4 of 4

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
Received EP
Bottle p,q,c 695,693,484,47,628
Collected By: HOYT
Comment:

Sample MANHATTAN

Sample 39750

ID of Source:
Collection 11/8/2001 AT 11:08:00
Submitted 11/9/2001 AT 10:49:00
Sample Type: POT_DW
PWS No.:

Source Code:

Type Descriptor:
Sample chilled on arrival? : yes

Report

Sample No. AD27548

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
<u>Inorganics</u>						
EPA94	Silver	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	12/10/2001	PAD
SM18	Arsenic	< MDL	ug/L	20	11/12/2001	PAD
EPA94	Barium	< MDL	ug/L	200.0	12/10/2001	PAD
EPA94	Beryllium	< MDL	ug/L	2	12/10/2001	PAD
EPA94	Calcium	6200	ug/L	1000.0	12/10/2001	PAD
EPA94	Cadmium	< MDL	ug/L	2.0	12/10/2001	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94	Chromium	< MDL	ug/L	10.0	12/10/2001	PAD
EPA94	Copper	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94	Iron	42	ug/L	20.0	12/10/2001	PAD
EPA79	Mercury	< MDL	ug/L	0.40	11/14/2001	PAD
SM18	Potassium	530	ug/L	100.0	11/19/2001	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 1/22/2002

Environmental Laboratories

emailed

Original 1/22/2002

NYS ELAP # 10108

Page 1 of 4

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
Received EP
Bottle p,q,c 695,693,484,47,628
Collected By: HOYT
Comment:

Sample MANHATTAN

Sample 39750

ID of Source:
Collection 11/8/2001 AT 11:08:00
Submitted 11/9/2001 AT 10:49:00
Sample Type: POT_DW
PWS No.:

Source Code:

Type Descriptor:
Sample chilled on arrival? : yes

Report

Sample No. AD27548

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Magnesium	1300	ug/L	1000.0	12/10/2001	PAD
EPA94	Manganese	25	ug/L	20.0	12/10/2001	PAD
EPA94	Molybdenum	< MDL	ug/L	50.0	12/10/2001	PAD
EPA94	Sodium	7600	ug/L	1000.0	12/10/2001	PAD
EPA94	Nickel	< MDL	ug/L	20.0	12/10/2001	PAD
SM18	Lead	< MDL	ug/L	1.0	11/28/2001	PAD
EPA79	Potable Metals Digestion	Completed			11/15/2001	PAD
EPA94	Antimony	< MDL	ug/L	0.1	12/13/2001	PAD
SM18	Selenium	< MDL	ug/L	2	12/12/2001	PAD
SM18	Tin	< MDL	ug/L	100	12/21/2001	PAD
SM18	Titanium	< MDL	ug/L	100	1/18/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	12/14/2001	PAD
EPA94	Vanadium	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94	Zinc	< MDL	ug/L	20.0	12/10/2001	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 1/22/2002

Environmental Laboratories

emailed

Original 1/22/2002

NYS ELAP # 10108

Page 2 of 4

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle p,q,c 695,693,484,47,628

Collected By: HOYT

Comment:

Sample MANHATTAN

Sample 39750

ID of Source:

Collection 11/8/2001 AT 11:08:00

Submitted 11/9/2001 AT 10:49:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD27548

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Organics						
Chlorine Pesticides - 508						
	4,4-DDE	< MDL	ug/L	0.080	11/15/2001	DLV
	Aldrin	< MDL	ug/L	0.010	11/15/2001	DLV
	Chlordane	< MDL	ug/L	0.40	11/15/2001	DLV
	Dieldrin	< MDL	ug/L	0.016	11/15/2001	DLV
	Endrin	< MDL	ug/L	0.020	11/15/2001	DLV
	Heptachlor	< MDL	ug/L	0.020	11/15/2001	DLV
	Heptachlor Epoxide	< MDL	ug/L	0.030	11/15/2001	DLV
	Lindane	< MDL	ug/L	0.030	11/15/2001	DLV
	Methoxychlor	< MDL	ug/L	0.039	11/15/2001	DLV
	PCB's	< MDL	ug/L	0.20	11/15/2001	DLV
	Propachlor	< MDL	ug/L	0.040	11/15/2001	DLV
	Toxaphene	< MDL	ug/L	0.40	11/15/2001	DLV

The recoveries for 4 of the 11 compounds in the LCS Standard were below their acceptance ranges.

Organic Chemicals - 525.2

2,4-Dinitrotoluene	< MDL	ug/L	2.0	11/19/2001	DLV
2,6-Dinitrotoluene	< MDL	ug/L	2.0	11/19/2001	DLV
4,4-DDE	< MDL	ug/L	0.80	11/19/2001	DLV
Acetochlor	< MDL	ug/L	2.0	11/19/2001	DLV
Alachlor	< MDL	ug/L	0.20	11/19/2001	DLV
Atrazine	< MDL	ug/L	0.20	11/19/2001	DLV

Approved Pam Dilsizian

QA Officer

Date Approved: 1/22/2002

Environmental Laboratories

emailed

Original 1/22/2002

NYS ELAP # 10108

Page 3 of 4

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
Received EP
Bottle p,q,c 695,693,484,47,628
Collected By: HOYT
Comment:

Sample MANHATTAN

Sample 39750
ID of Source:
Collection 11/8/2001 AT 11:08:00
Submitted 11/9/2001 AT 10:49:00
Sample Type: POT_DW
PWS No.: **Source Code:**

Type Descriptor:
Sample chilled on arrival? : yes

Report

Sample No. AD27548

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
	Benzo(a)pyrene	< MDL	ug/L	0.20	11/19/2001	DLV
	bis(2-Ethylhexyl)adipate	< MDL	ug/L	0.60	11/19/2001	DLV
	bis(2-Ethylhexyl)phthalate	1.3	ug/L	0.60	11/19/2001	DLV
	Butachlor	< MDL	ug/L	0.40	11/19/2001	DLV
	EPTC	< MDL	ug/L	1.0	11/19/2001	DLV
	Hexachlorobenzene	< MDL	ug/L	0.10	11/19/2001	DLV
	Hexachlorocyclopentadiene	< MDL	ug/L	0.10	11/19/2001	DLV
	Metolachlor	< MDL	ug/L	0.75	11/19/2001	DLV
	Metribuzin	< MDL	ug/L	0.15	11/19/2001	DLV
	Molinate	< MDL	ug/L	0.90	11/19/2001	DLV
	Simazine	< MDL	ug/L	0.40	11/19/2001	DLV
	Terbacil	< MDL	ug/L	2.0	11/19/2001	DLV

The recoveries for 6 of the 18 compounds in the LCS Standard were outside of their acceptance ranges.

GCMS Scan For Unknowns

Other unknown compounds	.	ug/L	12/18/2001	DLV
-------------------------	---	------	------------	-----

A scan of the extract prepared for the 508 analysis, from 35-550 amu using a mass spectrometer, does not reveal the presence of non-target compounds.

Approved Pam Dilsizian

QA Officer

Date Approved: 1/22/2002

Environmental Laboratories

emailed

Original 1/22/2002

NYS ELAP # 10108

Page 4 of 4

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle c499,p715,723,714,q609,635,608

Collected By: DESSOURCES

Comment:

Sample QUEENS

Sample 43550

ID of Source:

Collection 11/8/2001 AT 11:56:00

Submitted 11/9/2001 AT 10:49:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD27549

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
<u>Inorganics</u>						
EPA94	Silver	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	12/10/2001	PAD
SM18	Arsenic	< MDL	ug/L	20	11/28/2001	PAD
EPA94	Barium	< MDL	ug/L	200.0	12/10/2001	PAD
EPA94	Beryllium	< MDL	ug/L	2	12/10/2001	PAD
EPA94	Calcium	6300	ug/L	1000.0	12/10/2001	PAD
EPA94	Cadmium	PENDING	ug/L	2.0	12/10/2001	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94	Chromium	< MDL	ug/L	10.0	12/10/2001	PAD
EPA94	Copper	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94	Iron	39	ug/L	20.0	12/10/2001	PAD
EPA79	Mercury	< MDL	ug/L	0.40	11/14/2001	PAD
SM18	Potassium	520	ug/L	100.0	11/19/2001	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 1/22/2002

Environmental Laboratories

emailed

Original 1/22/2002

NYS ELAP # 10108

Page 1 of 3

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle c499,p715,723,714,q609,635,608

Collected By: DESSOURCES

Comment:

Sample QUEENS

Sample 43550

ID of Source:

Collection 11/8/2001 AT 11:56:00

Submitted 11/9/2001 AT 10:49:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD27549

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Magnesium	1200	ug/L	1000.0	12/10/2001	PAD
EPA94	Manganese	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94	Molybdenum	< MDL	ug/L	50.0	12/10/2001	PAD
EPA94	Sodium	6700	ug/L	1000.0	12/10/2001	PAD
EPA94	Nickel	< MDL	ug/L	20.0	12/10/2001	PAD
SM18	Lead	< MDL	ug/L	1.0	11/28/2001	PAD
EPA'79	Potable Metals Digestion	Completed			11/15/2001	PAD
EPA94	Antimony	< MDL	ug/L	0.1	12/13/2001	PAD
SM18	Selenium	< MDL	ug/L	2	12/12/2001	PAD
SM18	Tin	< MDL	ug/L	100	12/21/2001	PAD
SM18	Titanium	< MDL	ug/L	100	1/18/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	12/14/2001	PAD
EPA94	Vanadium	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94	Zinc	< MDL	ug/L	20.0	12/10/2001	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 1/22/2002

Environmental Laboratories

emailed

Original 1/22/2002

NYS ELAP # 10108

Page 2 of 3

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Sample QUEENS

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle c499,p715,723,714,q609,635,608

Collected By: DESSOURCES

Comment:

Sample 43550

ID of Source:

Collection 11/8/2001 AT 11:56:00

Submitted 11/9/2001 AT 10:49:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD27549

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Organics						
Chlorine Pesticides - 508						
	4,4-DDE	< MDL	ug/L	0.080	11/15/2001	DLV
	Aldrin	< MDL	ug/L	0.010	11/15/2001	DLV
	Chlordane	< MDL	ug/L	0.40	11/15/2001	DLV
	Dieldrin	< MDL	ug/L	0.016	11/15/2001	DLV
	Endrin	< MDL	ug/L	0.020	11/15/2001	DLV
	Heptachlor	< MDL	ug/L	0.020	11/15/2001	DLV
	Heptachlor Epoxide	< MDL	ug/L	0.030	11/15/2001	DLV
	Lindane	< MDL	ug/L	0.030	11/15/2001	DLV
	Methoxychlor	< MDL	ug/L	0.039	11/15/2001	DLV
	PCB's	< MDL	ug/L	0.20	11/15/2001	DLV
	Propachlor	< MDL	ug/L	0.040	11/15/2001	DLV
	Toxaphene	< MDL	ug/L	0.40	11/15/2001	DLV

The recoveries for 4 of the 11 compounds in the LCS Standard were below their acceptance ranges.

GCMS Scan For Unknowns

Other unknown compounds	.	ug/L	12/18/2001	DLV
-------------------------	---	------	------------	-----

A scan of the extract prepared for the 508 analysis, from 35-550 amu using a mass spectrometer, does not reveal the presence of non-target compounds.

Approved Pam Dilsizian

QA Officer

Date Approved: 1/22/2002

Environmental Laboratories

emailed

Original 1/22/2002

NYS ELAP # 10108

Page 3 of 3

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Sample STATEN ISLAND

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle c491,p692,694,718,q612,617,625

Collected By: MOHAN

Comment:

Sample 50650

ID of Source:

Collection 11/8/2001 AT 11:55:00

Submitted 11/9/2001 AT 10:49:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD27550

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
<u>Inorganics</u>						
EPA94	Silver	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	12/10/2001	PAD
SM18	Arsenic	< MDL	ug/L	20	11/28/2001	PAD
EPA94	Barium	< MDL	ug/L	200.0	12/10/2001	PAD
EPA94	Beryllium	< MDL	ug/L	2	12/10/2001	PAD
EPA94	Calcium	6400	ug/L	1000.0	12/10/2001	PAD
EPA94	Cadmium	< MDL	ug/L	2.0	12/10/2001	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94	Chromium	< MDL	ug/L	10.0	12/10/2001	PAD
EPA94	Copper	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94	Iron	54	ug/L	20.0	12/10/2001	PAD
EPA79	Mercury	< MDL	ug/L	0.40	11/14/2001	PAD
SM18	Potassium	590	ug/L	100.0	11/19/2001	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 1/22/2002

Environmental Laboratories

emailed

Original 1/22/2002

NYS ELAP # 10108

Page 1 of 3

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle c491,p692,694,718,q612,617,625

Collected By: MOHAN

Comment:

Sample STATEN ISLAND

Sample 50650

ID of Source:

Collection 11/8/2001 AT 11:55:00

Submitted 11/9/2001 AT 10:49:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD27550

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Magnesium	1200	ug/L	1000.0	12/10/2001	PAD
EPA94	Manganese	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94	Molybdenum	< MDL	ug/L	50.0	12/10/2001	PAD
EPA94	Sodium	7000	ug/L	1000.0	12/10/2001	PAD
EPA94	Nickel	< MDL	ug/L	20.0	12/10/2001	PAD
SM18	Lead	< MDL	ug/L	1.0	11/28/2001	PAD
EPA79	Potable Metals Digestion	Completed			11/15/2001	PAD
EPA94	Antimony	< MDL	ug/L	0.1	12/13/2001	PAD
SM18	Selenium	< MDL	ug/L	2	12/12/2001	PAD
SM18	Tin	< MDL	ug/L	100	12/21/2001	PAD
SM18	Titanium	< MDL	ug/L	100	1/18/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	12/14/2001	PAD
EPA94	Vanadium	< MDL	ug/L	20.0	12/10/2001	PAD
EPA94	Zinc	< MDL	ug/L	20.0	12/10/2001	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 1/22/2002

Environmental Laboratories

emailed

Original 1/22/2002

NYS ELAP # 10108

Page 2 of 3

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Sample STATEN ISLAND

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle c491,p692,694,718,q612,617,625

Collected By: MOHAN

Comment:

Sample 50650

ID of Source:

Collection 11/8/2001 AT 11:55:00

Submitted 11/9/2001 AT 10:49:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD27550

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Organics						
Chlorine Pesticides - 508						
	4,4-DDE	< MDL	ug/L	0.080	11/15/2001	DLV
	Aldrin	< MDL	ug/L	0.010	11/15/2001	DLV
	Chlordane	< MDL	ug/L	0.40	11/15/2001	DLV
	Dieldrin	< MDL	ug/L	0.016	11/15/2001	DLV
	Endrin	< MDL	ug/L	0.020	11/15/2001	DLV
	Heptachlor	< MDL	ug/L	0.020	11/15/2001	DLV
	Heptachlor Epoxide	< MDL	ug/L	0.030	11/15/2001	DLV
	Lindane	< MDL	ug/L	0.030	11/15/2001	DLV
	Methoxychlor	< MDL	ug/L	0.039	11/15/2001	DLV
	PCB's	< MDL	ug/L	0.20	11/15/2001	DLV
	Propachlor	< MDL	ug/L	0.040	11/15/2001	DLV
	Toxaphene	< MDL	ug/L	0.40	11/15/2001	DLV

The recoveries for 4 of the 11 compounds in the LCS Standard were below their acceptance ranges. The recovery for the Surrogate Standard in this sample was below its acceptance range. The EPA 525.2 analysis was cancelled

GCMS Scan For Unknowns

Other unknown compounds	.	ug/L	12/18/2001	DLV
-------------------------	---	------	------------	-----

A scan of the extract prepared for the 508 analysis, from 35-550 amu using a mass spectrometer, does not reveal the presence of non-target compounds.

Approved Pam Dilsizian

QA Officer

Date Approved: 1/22/2002

Environmental Laboratories

emailed

Original 1/22/2002

NYS ELAP # 10108

Page 3 of 3