

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
Received BH
Bottle p1145,q414
Collected By: LAB
Comment:

Sample FIELD BLANK

Sample FBD
ID of Source:
Collection 11/22/2001 AT
Submitted 11/22/2001 AT 9:40:00
Sample Type: POT_DW
PWS No.: **Source Code:**

Type Descriptor:
Sample chilled on arrival? : YES

Report

Sample No. AD28662

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Organics						
GCMS Scan For Unknowns						
	Other unknown compounds	.	ug/L		12/7/2001	DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds.

The recovery for the Surrogate Standard in this sample was 89%. Recoveries for 9 of the 44 compounds in the LCS Standard were outside of their acceptance ranges. The chromatogram reveals contamination which has been identified to be from the TurboVap concentrator.

Approved Pam Dilsizian

QA Officer

Date Approved: 1/10/2002

Environmental Laboratories

emailed

Copy 1/10/2002

NYS ELAP # 10108

Page 1 of 1

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
Received BH
Bottle p1145,q414
Collected By: LAB
Comment:

Sample FIELD BLANK

Sample FB-11/22/01
ID of Source:
Collection 11/22/2001 AT
Submitted 11/22/2001 AT 9:40:00
Sample Type: POT_DW
PWS No.:

Source Code:

Type Descriptor:
Sample chilled on arrival? : YES

Report

Sample No. AD28662

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Organics						
GCMS Scan For Unknowns						
	Other unknown compounds	.	ug/L		12/7/2001	DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds.

The recovery for the Surrogate Standard in this sample was 89%. Recoveries for 9 of the 44 compounds in the LCS Standard were outside of their acceptance ranges. The chromatogram reveals contamination which has been identified to be from the TurboVap concentrator.

Approved Pam Dilsizian

QA Officer

Date Approved: 12/10/2001

Environmental Laboratories

emailed

Original 12/10/2001

NYS ELAP # 10108

Page 1 of 1

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
Received BH
Bottle c759,p1128,1130,q602,380,348,3
Collected By: YACOB
Comment:

Sample 3

Sample 3
ID of Source:
Collection 11/22/2001 AT 9:02:00
Submitted 11/22/2001 AT 9:40:00
Sample Type: POT_DW
PWS No.: **Source Code:**

Type Descriptor:
Sample chilled on arrival? : YES

Report

Sample No. AD28663

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
<u>Inorganics</u>						
<u>ICPMS Metals for NEWQ</u>						
EPA94	Antimony	< MDL	ug/L	0.1	12/19/2001	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	12/19/2001	PAD
EPA94	Barium	14.0	ug/L	0.1	12/19/2001	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	12/19/2001	PAD
EPA94	Chromium	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Copper	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Lead	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Manganese	24.2	ug/L	1.0	12/19/2001	PAD
EPA94	Nickel	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Selenium	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Silver	< MDL	ug/L	2.0	12/19/2001	PAD
EPA94	Thallium	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Zinc	< MDL	ug/L	2.0	12/19/2001	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	12/20/2001	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	12/20/2001	PAD
EPA94	Iron	24	ug/L	20.0	12/20/2001	PAD
EPA79	Mercury	< MDL	ug/L	0.40	11/27/2001	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 1/2/2002

Environmental Laboratories

emailed

Original 1/2/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
Received BH
Bottle c759,p1128,1130,q602,380,348,3
Collected By: YACoub
Comment:

Sample 3

Sample 3
ID of Source:
Collection 11/22/2001 AT 9:02:00
Submitted 11/22/2001 AT 9:40:00
Sample Type: POT_DW
PWS No.: **Source Code:**

Type Descriptor:
Sample chilled on arrival? : YES

Report

Sample No. AD28663

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	12/20/2001	PAD
EPA'79	Potable Metals Digestion	Completed			11/26/2001	PAD
SM18	Tin	< MDL	ug/L	100	1/2/2002	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds	.	ug/L	12/7/2001	DLV
-------------------------	---	------	-----------	-----

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds.

The recovery for the Surrogate Standard in this sample was 103%. Recoveries for 9 of the 44 compounds in the LCS Standard were outside of their acceptance ranges. The chromatogram reveals contamination which has been identified to be from the TurboVap concentrator.

Approved Pam Dilsizian **QA Officer**

Environmental Laboratories

NYS ELAP # 10108

Date Approved: 1/2/2002

emailed Original 1/2/2002

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Sample 58

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received BH

Bottle c701,p1150,1151,q983,984

Collected By: YACUB

Comment:

Sample 58

ID of Source:

Collection 11/22/2001 AT 8:30:00

Submitted 11/22/2001 AT 9:40:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : YES

Report

Sample No. AD28664

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
ICPMS Metals for NEWQ						
EPA94	Antimony	< MDL	ug/L	0.1	12/19/2001	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	12/19/2001	PAD
EPA94	Barium	13.4	ug/L	0.1	12/19/2001	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	12/19/2001	PAD
EPA94	Chromium	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Copper	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Lead	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Manganese	22.3	ug/L	1.0	12/19/2001	PAD
EPA94	Nickel	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Selenium	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Silver	< MDL	ug/L	2.0	12/19/2001	PAD
EPA94	Thallium	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Zinc	< MDL	ug/L	2.0	12/19/2001	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	12/20/2001	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	12/20/2001	PAD
EPA94	Iron	25	ug/L	20.0	12/20/2001	PAD
EPA79	Mercury	< MDL	ug/L	0.40	11/27/2001	PAD

Approved Pam Dilsizian

QA Officer

Date Approved:

1/2/2002

Environmental Laboratories

emailed

Original 1/2/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
 Received BH
 Bottle c701,p1150,1151,q983,984
 Collected By: YACOB
 Comment:

Sample 58
 Sample 58
 ID of Source:
 Collection 11/22/2001 AT 8:30:00
 Submitted 11/22/2001 AT 9:40:00
 Sample Type: POT_DW
 PWS No.: Source Code:
 Type Descriptor:
 Sample chilled on arrival? : YES

Report

Sample No. AD28664						
Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	12/20/2001	PAD
EPA79	Potable Metals Digestion	Completed			11/26/2001	PAD
SM18	Tin	< MDL	ug/L	100	1/2/2002	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds	.	ug/L	12/7/2001	DLV
-------------------------	---	------	-----------	-----

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds.
 The recovery for the Surrogate Standard in this sample was 95%. Recoveries for 9 of the 44 compounds in the LCS Standard were outside of their acceptance ranges. The chromatogram reveals contamination which has been identified to be from the TurboVap concentrator.

Approved Pam Dilsizian	QA Officer	Date Approved:	1/2/2002
Environmental Laboratories		emailed	Original 1/2/2002
NYS ELAP # 10108			Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received BH

Bottle c545,p924,604,q168,382

Collected By: PATEL

Comment:

Sample 36045

Sample 11550

ID of Source:

Collection 11/21/2001 AT 9:48:00

Submitted 11/22/2001 AT 9:40:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : YES

Report

Sample No. AD28665

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
<u>Inorganics</u>						
<u>ICPMS Metals for NEWQ</u>						
EPA94	Antimony	< MDL	ug/L	0.1	2/1/2002	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	2/1/2002	PAD
EPA94	Barium	14.3	ug/L	0.1	2/1/2002	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	2/1/2002	PAD
EPA94	Chromium	< MDL	ug/L	1.0	2/1/2002	PAD
EPA94	Copper	10.2	ug/L	1.0	2/1/2002	PAD
EPA94	Lead	< MDL	ug/L	1.0	2/1/2002	PAD
EPA94	Manganese	24.6	ug/L	1.0	2/1/2002	PAD
EPA94	Nickel	< MDL	ug/L	1.0	2/1/2002	PAD
EPA94	Selenium	< MDL	ug/L	1.0	2/1/2002	PAD
EPA94	Silver	< MDL	ug/L	2.0	2/1/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	2/1/2002	PAD
EPA94	Zinc	2.7	ug/L	2.0	2/1/2002	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	12/20/2001	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	12/20/2001	PAD
EPA94	Iron	28	ug/L	20.0	12/20/2001	PAD
EPA79	Mercury	< MDL	ug/L	0.40	11/27/2001	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 2/4/2002

Environmental Laboratories

emailed

Copy 2/4/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
Received BH
Bottle c545,p924,604,q168,382
Collected By: PATEL
Comment:

Sample 36045

Sample 11550
ID of Source:
Collection 11/21/2001 AT 9:48:00
Submitted 11/22/2001 AT 9:40:00
Sample Type: POT_DW
PWS No.: **Source Code:**

Type Descriptor:
Sample chilled on arrival? : YES

Report

Sample No. AD28665

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	12/20/2001	PAD
EPA79	Potable Metals Digestion	Completed			11/26/2001	PAD
SM18	Tin	< MDL	ug/L	100	1/2/2002	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds	.	ug/L	12/7/2001	DLV
-------------------------	---	------	-----------	-----

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds.

The recovery for the Surrogate Standard in this sample was 85%. Recoveries for 16 of the 44 compounds in the LCS Standard were outside of their acceptance ranges.

Approved Pam Dilsizian

QA Officer

Date Approved: 2/4/2002

Environmental Laboratories

emailed

Copy 2/4/2002

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
Received BH
Bottle c536,p728,1118,q953,965
Collected By: MOHAN
Comment:

Sample 36060

Sample 25150
ID of Source:
Collection 11/21/2001 AT
Submitted 11/22/2001 AT 9:40:00
Sample Type: POT_DW
PWS No.: **Source Code:**

Type Descriptor:
Sample chilled on arrival? : YES

Report

Sample No. AD28666

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
<u>ICPMS Metals for NEWQ</u>						
EPA94	Antimony	< MDL	ug/L	0.1	12/19/2001	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	12/19/2001	PAD
EPA94	Barium	15.2	ug/L	0.1	12/19/2001	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	12/19/2001	PAD
EPA94	Chromium	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Copper	13.7	ug/L	1.0	12/19/2001	PAD
EPA94	Lead	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Manganese	33.8	ug/L	1.0	12/19/2001	PAD
EPA94	Nickel	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Selenium	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Silver	< MDL	ug/L	2.0	12/19/2001	PAD
EPA94	Thallium	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Zinc	< MDL	ug/L	2.0	12/19/2001	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	12/20/2001	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	12/20/2001	PAD
EPA94	Iron	36	ug/L	20.0	12/20/2001	PAD
EPA79	Mercury	< MDL	ug/L	0.40	11/27/2001	PAD

Approved Pam Dilsizian **QA Officer**

Date Approved: 1/2/2002

Environmental Laboratories

emailed Original 1/2/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
 Received BH
 Bottle c536,p728,1118,q953,965
 Collected By: MOHAN
 Comment:

Sample 36060

 Sample 25150
 ID of Source:
 Collection 11/21/2001 AT
 Submitted 11/22/2001 AT 9:40:00
 Sample Type: POT_DW
 PWS No.: Source Code:

 Type Descriptor:
 Sample chilled on arrival? : YES

Report

		Sample No.	AD28666				
Method	Test Description		Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum		< MDL	ug/L	50.0	12/20/2001	PAD
EPA79	Potable Metals Digestion		Completed			11/26/2001	PAD
SM18	Tin		< MDL	ug/L	100	1/2/2002	PAD
Organics							
GCMS Scan For Unknowns							
	Other unknown compounds		.	ug/L		12/7/2001	DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds.
 The recovery for the Surrogate Standard in this sample was 85%. Recoveries for 16 of the 44 compounds in the LCS Standard were outside of their acceptance ranges.

Approved Pam Dilsizian		QA Officer	Date Approved: 1/2/2002	
Environmental Laboratories			emailed	Original 1/2/2002
NYS ELAP # 10108			Page 2 of 2	

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
Received BH
Bottle c543,p461,q618,381
Collected By: SPERANZA
Comment:

Sample 36071

Sample 38650

ID of Source:
Collection 11/21/2001 AT 9:14:00
Submitted 11/22/2001 AT 9:40:00
Sample Type: POT_DW
PWS No.:

Source Code:

Type Descriptor:
Sample chilled on arrival? : YES

Report

Sample No. AD28667

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
<u>Inorganics</u>						
<u>ICPMS Metals for NEWQ</u>						
EPA94	Antimony	< MDL	ug/L	0.1	12/19/2001	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	12/19/2001	PAD
EPA94	Barium	14.4	ug/L	0.1	12/19/2001	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	12/19/2001	PAD
EPA94	Chromium	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Copper	7.5	ug/L	1.0	12/19/2001	PAD
EPA94	Lead	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Manganese	30.5	ug/L	1.0	12/19/2001	PAD
EPA94	Nickel	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Selenium	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Silver	< MDL	ug/L	2.0	12/19/2001	PAD
EPA94	Thallium	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Zinc	< MDL	ug/L	2.0	12/19/2001	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	12/20/2001	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	12/20/2001	PAD
EPA94	Iron	35	ug/L	20.0	12/20/2001	PAD
EPA79	Mercury	< MDL	ug/L	0.40	11/27/2001	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 1/2/2002

Environmental Laboratories

emailed

Original 1/2/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
 Received BH
 Bottle c543,p461,q618,381
 Collected By: SPERANZA
 Comment:

Sample 36071

Sample 38650
 ID of Source:
 Collection 11/21/2001 AT 9:14:00
 Submitted 11/22/2001 AT 9:40:00
 Sample Type: POT_DW
 PWS No.: Source Code:

Type Descriptor:
 Sample chilled on arrival? : YES

Report

Sample No. AD28667

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	12/20/2001	PAD
EPA79	Potable Metals Digestion	Completed			11/26/2001	PAD
SM18	Tin	< MDL	ug/L	100	1/2/2002	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds	.	ug/L	12/7/2001	DLV
-------------------------	---	------	-----------	-----

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds.
 The recovery for the Surrogate Standard in this sample was 78%. Recoveries for 16 of the 44 compounds in the LCS Standard were outside of their acceptance ranges.

Approved Pam Dilsizian QA Officer

Date Approved: 1/2/2002

Environmental Laboratories

emailed Original 1/2/2002

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
Received BH
Bottle c549,464,922,124
Collected By: BOLTON
Comment:

Sample 36073

Sample 40250
ID of Source:
Collection 11/21/2001 AT 9:02:00
Submitted 11/22/2001 AT 9:40:00
Sample Type: POT_DW
PWS No.: **Source Code:**

Type Descriptor:
Sample chilled on arrival? : YES

Report

Sample No. AD28668

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
<u>ICPMS Metals for NEWQ</u>						
EPA94	Antimony	< MDL	ug/L	0.1	12/19/2001	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	12/19/2001	PAD
EPA94	Barium	15.0	ug/L	0.1	12/19/2001	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	12/19/2001	PAD
EPA94	Chromium	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Copper	4.6	ug/L	1.0	12/19/2001	PAD
EPA94	Lead	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Manganese	27.2	ug/L	1.0	12/19/2001	PAD
EPA94	Nickel	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Selenium	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Silver	< MDL	ug/L	2.0	12/19/2001	PAD
EPA94	Thallium	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Zinc	3.6	ug/L	2.0	12/19/2001	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	12/20/2001	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	12/20/2001	PAD
EPA94	Iron	34	ug/L	20.0	12/20/2001	PAD
EPA79	Mercury	< MDL	ug/L	0.40	11/27/2001	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 1/2/2002

Environmental Laboratories

emailed

Original 1/2/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
Distribution Laboratory
96-05 Horace Harding Ex.
Corona N.Y. 11368
Dr. Ian Hurley
Received BH
Bottle c549,464,922,124
Collected By: BOLTON
Comment:

Sample 36073

Sample 40250
ID of Source:
Collection 11/21/2001 AT 9:02:00
Submitted 11/22/2001 AT 9:40:00
Sample Type: POT_DW
PWS No.: **Source Code:**

Type Descriptor:
Sample chilled on arrival? : YES

Report

Sample No. AD28668

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	12/20/2001	PAD
EPA'79	Potable Metals Digestion	Completed			11/26/2001	PAD
SM18	Tin	< MDL	ug/L	100	1/2/2002	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds	.	ug/L	12/7/2001	DLV
-------------------------	---	------	-----------	-----

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds.

The recovery for the Surrogate Standard in this sample was 80%. Recoveries for 16 of the 44 compounds in the LCS Standard were outside of their acceptance ranges.

Approved Pam Dilsizian

QA Officer

Date Approved: 1/2/2002

Environmental Laboratories

emailed

Original 1/2/2002

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
Received BH
Bottle c555,p1093,1098,q936,934
Collected By: GERTSIK
Comment:

Sample 36083
Sample 51350
ID of Source:
Collection 11/21/2001 AT 1:39:00
Submitted 11/22/2001 AT 9:40:00
Sample Type: POT_DW
PWS No.: **Source Code:**

Type Descriptor:
Sample chilled on arrival? : YES

Report

Sample No. AD28669

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
ICPMS Metals for NEWQ						
EPA94	Antimony	< MDL	ug/L	0.1	12/19/2001	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	12/19/2001	PAD
EPA94	Barium	14.7	ug/L	0.1	12/19/2001	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	12/19/2001	PAD
EPA94	Chromium	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Copper	10.8	ug/L	1.0	12/19/2001	PAD
EPA94	Lead	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Manganese	19.1	ug/L	1.0	12/19/2001	PAD
EPA94	Nickel	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Selenium	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Silver	< MDL	ug/L	2.0	12/19/2001	PAD
EPA94	Thallium	< MDL	ug/L	1.0	12/19/2001	PAD
EPA94	Zinc	5.3	ug/L	2.0	12/19/2001	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	12/20/2001	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	12/20/2001	PAD
EPA94	Iron	24	ug/L	20.0	12/20/2001	PAD
EPA79	Mercury	< MDL	ug/L	0.40	11/27/2001	PAD

Approved Pam Dilisizian QA Officer

Date Approved: 1/2/2002

Environmental Laboratories

emailed Original 1/2/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
 Received BH
 Bottle c555,p1093,1098,q936,934
 Collected By: GERTSIK
 Comment:

Sample 36083

 Sample 51350
 ID of Source:
 Collection 11/21/2001 AT 1:39:00
 Submitted 11/22/2001 AT 9:40:00
 Sample Type: POT_DW
 PWS No.: Source Code:

 Type Descriptor:
 Sample chilled on arrival? : YES

Report

Sample No.		AD28669				
Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	12/20/2001	PAD
EPA'79	Potable Metals Digestion	Completed			11/26/2001	PAD
SM18	Tin	< MDL	ug/L	100	1/2/2002	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds	.	ug/L	12/7/2001	DLV
-------------------------	---	------	-----------	-----

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds, except for bis(2-Ethylhexyl)phthalate at an estimated level of 0.33 ug/L.
 The recovery for the Surrogate Standard in this sample was 102%. Recoveries for 9 of the 44 compounds in the LCS Standard were outside of their acceptance ranges.

Approved Pam Dilsizian	QA Officer	Date Approved: 1/2/2002
Environmental Laboratories		emailed Original 1/2/2002
NYS ELAP # 10108		Page 2 of 2