

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
 Received LK
 Bottle P1105, Q937
 Collected By: LAB
 Comment:

Sample FIELD BLANK

Sample FBD
 ID of Source:
 Collection 11/20/2001 AT
 Submitted 11/21/2001 AT
 Sample Type: POT_DW
 PWS No.:

Source Code:

Type Descriptor:
 Sample chilled on arrival? : YES

Report

Sample No. AD28600

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Organics						
<u>GCMS Scan For Unknowns</u>						
	Other unknown compounds	.	ug/L		12/7/2001	DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds.
 The recovery for the Surrogate Standard in this sample was 95%. Recoveries for 16 of the 44 compounds in the LCS Standard were outside of their acceptance ranges.

Approved Pam Dilsizian QA Officer

Environmental Laboratories

NYS ELAP # 10108

Date Approved: 1/10/2002

emailed Copy 1/10/2002

Page 1 of 1

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
 Received LK
 Bottle P1105, Q937
 Collected By: LAB
 Comment:

Sample FIELD BLANK

Sample FB 11/20/01
 ID of Source:
 Collection 11/20/2001 AT
 Submitted 11/21/2001 AT
 Sample Type: POT_DW
 PWS No.:

Source Code:

Type Descriptor:
 Sample chilled on arrival? : YES

Report

Sample No. AD28600

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Organics						
<u>GCMS Scan For Unknowns</u>						
	Other unknown compounds	.	ug/L		12/7/2001	DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds.
 The recovery for the Surrogate Standard in this sample was 95%. Recoveries for 16 of the 44 compounds in the LCS Standard were outside of their acceptance ranges.

Approved Donna Popp Chief of Env. Lab Services

Environmental Laboratories

NYS ELAP # 10108

Date Approved: 12/7/2001

emailed Original 12/7/2001

Page 1 of 1