

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Attention: Steve Schindler

Received EP

Bottle c276,p9166,9142,9169,q9145,14

Collected By: NEWMAN

Comment:

Sample NYC DEP/KENSICO LAB
19 WESTLAKE DRIVE
VALHALLA, N.Y.

Sample CATLEFF

ID of Source:

Collection 10/15/2001 AT

Submitted 10/15/2001 AT 2:10:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD25154

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
<u>Inorganics</u>						
EPA94	Beryllium	< MDL	ug/L	2	11/1/2001	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	11/1/2001	PAD
EPA94	Molybdenum	< MDL	ug/L	50.0	11/1/2001	PAD
EPA'79	Potable Metals Digestion	Completed			10/24/2001	PAD
SM18	Antimony	< MDL	ug/L	6	10/30/2001	PAD
SM18	Tin	< MDL	ug/L	100	10/19/2001	PAD
SM18	Titanium	< MDL	ug/L	100	10/25/2001	PAD
EPA94	Thallium	< MDL	ug/L	2	10/18/2001	PAD
EPA94	Vanadium	< MDL	ug/L	20.0	11/1/2001	PAD
<u>Organics</u>						
<u>Chlorine Pesticides - 508</u>						
	4,4-DDE	< MDL	ug/L	0.080	10/23/2001	DLV
	Aldrin	< MDL	ug/L	0.010	10/23/2001	DLV
	Chlordane	< MDL	ug/L	0.40	10/23/2001	DLV
	Dieldrin	< MDL	ug/L	0.016	10/23/2001	DLV
	Endrin	< MDL	ug/L	0.020	10/23/2001	DLV

Approved Pam Dilsizian

QA Officer

Date Approved: 12/24/2001

Environmental Laboratories

emailed

Original 12/24/2001

NYS ELAP # 10108

Page 1 of 6

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Sample NYC DEP/KENSICO LAB
19 WESTLAKE DRIVE
VALHALLA, N.Y.

Attention: Steve Schindler
Received EP

Sample CATLEFF
ID of Source:
Collection 10/15/2001 AT
Submitted 10/15/2001 AT 2:10:00
Sample Type: POT_DW
PWS No.:

Bottle c276,p9166,9142,9169,q9145,14
Collected By: NEWMAN
Comment:

Source Code: 000

Type Descriptor:
Sample chilled on arrival? : yes

Report

Sample No. AD25154

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
	Heptachlor	< MDL	ug/L	0.020	10/23/2001	DLV
	Heptachlor Epoxide	< MDL	ug/L	0.030	10/23/2001	DLV
	Lindane	< MDL	ug/L	0.030	10/23/2001	DLV
	Methoxychlor	< MDL	ug/L	0.039	10/23/2001	DLV
	PCB's	< MDL	ug/L	0.20	10/23/2001	DLV
	Propachlor	< MDL	ug/L	0.040	10/23/2001	DLV
	Toxaphene	< MDL	ug/L	0.40	10/23/2001	DLV

Recoveries for 3 of the 9 compounds in the LCS Standard were below their acceptance ranges.

Volatile Organic Compounds

524.2	*THM-Bromodichloromethane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	*THM-Bromoform	< MDL	ug/L	2.0	10/16/2001	SV
524.2	*THM-Chloroform	< MDL	ug/L	0.5	10/16/2001	SV
524.2	*THM-Dibromochloromethane	< MDL	ug/L	2.0	10/16/2001	SV
524.2	1,1,1,2-tetrachloroethane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,1,1- trichloroethane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,1,2,2-tetrachloroethane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,1,2-trichloroethane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,1-dichloroethane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,1-Dichloroethene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,1-dichloropropene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,2,3-trichlorobenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,2,3-trichloropropane	< MDL	ug/L	0.5	10/16/2001	SV

Approved Pam Dilsizian

QA Officer

Date Approved: 12/24/2001

Environmental Laboratories

emailed

Original 12/24/2001

NYS ELAP # 10108

Page 2 of 6

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Sample NYC DEP/KENSICO LAB
19 WESTLAKE DRIVE
VALHALLA, N.Y.

Attention: Steve Schindler

Received EP

Bottle c276,p9166,9142,9169,q9145,14

Collected By: NEWMAN

Comment:

Sample CATLEFF

ID of Source:

Collection 10/15/2001 AT

Submitted 10/15/2001 AT 2:10:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD25154

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
524.2	1,2,4-trichlorobenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,2,4-trimethylbenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,2-dichlorobenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,2-dichloroethane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,2-dichloropropane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,3,5-trimethylbenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,3-dichlorobenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,3-dichloropropane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,4-dichlorobenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	2,2-dichloropropane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	2-butanone (MEK)	< MDL	ug/L	2.0	10/16/2001	SV
524.2	2-chlorotoluene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	4-chlorotoluene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Benzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Bromobenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Bromochloromethane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Bromomethane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Carbon tetrachloride	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Chlorobenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Chloroethane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Chloromethane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	cis-1,2-dichloroethene	< MDL	ug/L	0.5	10/16/2001	SV

Approved Pam Dilsizian

QA Officer

Date Approved: 12/24/2001

Environmental Laboratories

emailed

Original 12/24/2001

NYS ELAP # 10108

Page 3 of 6

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Sample NYC DEP/KENSICO LAB
19 WESTLAKE DRIVE
VALHALLA, N.Y.

Attention: Steve Schindler

Received EP

Bottle c276,p9166,9142,9169,q9145,14

Collected By: NEWMAN

Comment:

Sample CATLEFF

ID of Source:

Collection 10/15/2001 AT

Submitted 10/15/2001 AT 2:10:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD25154

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
524.2	cis-1,3-dichloropropene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Dibromomethane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Dichlorodifluoromethane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Ethylbenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Hexachlorobutadiene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Isopropylbenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Methyl iso-butyl ketone (MIBK)	< MDL	ug/L	1.0	10/16/2001	SV
524.2	Methyl tert-butyl ether (MTBE)	< MDL	ug/L	1.0	10/16/2001	SV
524.2	Methylene Chloride	< MDL	ug/L	0.5	10/16/2001	SV
524.2	N-butylbenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	N-propylbenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Naphthalene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	O-xylene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	P & M-xylene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	P-isopropyltoluene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	SEC-butylbenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Styrene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	TERT-butylbenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Tetrachloroethene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Toluene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	trans-1,2-dichloroethene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	trans-1,3-dichloropropene	< MDL	ug/L	0.5	10/16/2001	SV

Approved Pam Dilsizian

QA Officer

Date Approved: 12/24/2001

Environmental Laboratories

emailed

Original 12/24/2001

NYS ELAP # 10108

Page 4 of 6

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Sample NYC DEP/KENSICO LAB
19 WESTLAKE DRIVE
VALHALLA, N.Y.

Attention: Steve Schindler
Received EP

Bottle c276,p9166,9142,9169,q9145,14

Collected By: NEWMAN

Comment:

Sample CATLEFF

ID of Source:

Collection 10/15/2001 AT

Submitted 10/15/2001 AT 2:10:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD25154

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
524.2	Trichloroethene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Trichlorofluoromethane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Vinyl chloride	< MDL	ug/L	0.5	10/16/2001	SV
<u>Organic Chemicals - 525.2</u>						
	2,4-Dinitrotoluene	< MDL	ug/L	2.0	10/25/2001	DLV
	2,6-Dinitrotoluene	< MDL	ug/L	2.0	10/25/2001	DLV
	4,4-DDE	< MDL	ug/L	0.80	10/25/2001	DLV
	Acetochlor	< MDL	ug/L	2.0	10/25/2001	DLV
	Alachlor	< MDL	ug/L	0.20	10/25/2001	DLV
	Atrazine	< MDL	ug/L	0.20	10/25/2001	DLV
	Benzo(a)pyrene	< MDL	ug/L	0.20	10/25/2001	DLV
	bis(2-Ethylhexyl)adipate	< MDL	ug/L	0.60	10/25/2001	DLV
	bis(2-Ethylhexyl)phthalate	0.75	ug/L	0.60	10/25/2001	DLV
	Butachlor	< MDL	ug/L	0.40	10/25/2001	DLV
	EPTC	< MDL	ug/L	1.0	10/25/2001	DLV
	Hexachlorobenzene	< MDL	ug/L	0.10	10/25/2001	DLV
	Hexachlorocyclopentadiene	< MDL	ug/L	0.10	10/25/2001	DLV
	Metolachlor	< MDL	ug/L	0.75	10/25/2001	DLV
	Metribuzin	< MDL	ug/L	0.15	10/25/2001	DLV
	Molinate	< MDL	ug/L	0.90	10/25/2001	DLV
	Simazine	< MDL	ug/L	0.40	10/25/2001	DLV
	Terbacil	< MDL	ug/L	2.0	10/25/2001	DLV

Approved Pam Dilsizian

QA Officer

Date Approved: 12/24/2001

Environmental Laboratories

emailed

Original 12/24/2001

NYS ELAP # 10108

Page 5 of 6

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Sample NYC DEP/KENSICO LAB
19 WESTLAKE DRIVE
VALHALLA, N.Y.

Attention: Steve Schindler

Received EP

Bottle c276,p9166,9142,9169,q9145,14

Collected By: NEWMAN

Comment:

Sample CATLEFF

ID of Source:

Collection 10/15/2001 AT

Submitted 10/15/2001 AT 2:10:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD25154

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Recoveries for 3 of the 18 compounds in the MDL and LCS Standards were outside of their acceptance ranges.						
Recoveries for 4 of the 18 compounds in the Spiked Sample were below their acceptance ranges.						
<u>GCMS Scan For Unknowns</u>						
	Other unknown compounds	.	ug/L		11/28/2001	DLV

A scan of the extract prepared for the 508 analysis, from 35-550 amu using a mass spectrometer, does not reveal the presence of non-target compounds.

Approved Pam Dilsizian

QA Officer

Date Approved: 12/24/2001

Environmental Laboratories

emailed

Original 12/24/2001

NYS ELAP # 10108

Page 6 of 6

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Sample NYC DEP/KENSICO LAB
19 WESTLAKE DRIVE
VALHALLA, N.Y.

Attention: Steve Schindler

Received EP

Bottle c284,p9153,9107,9112,q9147,91

Collected By: NEWMAN

Comment:

Sample DEL18

ID of Source:

Collection 10/15/2001 AT

Submitted 10/15/2001 AT 2:10:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD25155

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
<u>Inorganics</u>						
EPA94	Beryllium	< MDL	ug/L	2	11/1/2001	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	11/1/2001	PAD
EPA94	Molybdenum	< MDL	ug/L	50.0	11/1/2001	PAD
EPA'79	Potable Metals Digestion	Completed			10/24/2001	PAD
SM18	Antimony	< MDL	ug/L	6	10/30/2001	PAD
SM18	Tin	< MDL	ug/L	100	10/19/2001	PAD
SM18	Titanium	< MDL	ug/L	100	10/25/2001	PAD
EPA94	Thallium	< MDL	ug/L	2	10/18/2001	PAD
EPA94	Vanadium	< MDL	ug/L	20.0	11/1/2001	PAD
<u>Organics</u>						
<u>Chlorine Pesticides - 508</u>						
	4,4-DDE	< MDL	ug/L	0.080	10/23/2001	DLV
	Aldrin	< MDL	ug/L	0.010	10/23/2001	DLV
	Chlordane	< MDL	ug/L	0.40	10/23/2001	DLV
	Dieldrin	< MDL	ug/L	0.016	10/23/2001	DLV
	Endrin	< MDL	ug/L	0.020	10/23/2001	DLV

Approved Pam Dilsizian

QA Officer

Date Approved: 12/24/2001

Environmental Laboratories

emailed

Copy 12/24/2001

NYS ELAP # 10108

Page 1 of 6

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Sample NYC DEP/KENSICO LAB
19 WESTLAKE DRIVE
VALHALLA, N.Y.

Attention: Steve Schindler

Received EP

Bottle c284,p9153,9107,9112,q9147,91

Collected By: NEWMAN

Comment:

Sample DEL18

ID of Source:

Collection 10/15/2001 AT

Submitted 10/15/2001 AT 2:10:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD25155

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
	Heptachlor	< MDL	ug/L	0.020	10/23/2001	DLV
	Heptachlor Epoxide	< MDL	ug/L	0.030	10/23/2001	DLV
	Lindane	< MDL	ug/L	0.030	10/23/2001	DLV
	Methoxychlor	< MDL	ug/L	0.039	10/23/2001	DLV
	PCB's	< MDL	ug/L	0.20	10/23/2001	DLV
	Propachlor	< MDL	ug/L	0.040	10/23/2001	DLV
	Toxaphene	< MDL	ug/L	0.40	10/23/2001	DLV

Recoveries for 3 of the 9 compounds in the LCS Standard were below their acceptance ranges. Further analyses will attempt to evaluate the 4,4'-DDE reported.

Re-extraction and re-analyses of this sample does not reveal any presence of 4,4-DDE.

An interference or contaminant probably led to a false positive result for 4,4-DDE in the original analysis.

Volatile Organic Compounds

524.2	*THM-Bromodichloromethane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	*THM-Bromoform	< MDL	ug/L	2.0	10/16/2001	SV
524.2	*THM-Chloroform	< MDL	ug/L	0.5	10/16/2001	SV
524.2	*THM-Dibromochloromethane	< MDL	ug/L	2.0	10/16/2001	SV
524.2	1,1,1,2-tetrachloroethane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,1,1- trichloroethane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,1,2,2-tetrachloroethane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,1,2-trichloroethane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,1-dichloroethane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,1-Dichloroethene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,1-dichloropropene	< MDL	ug/L	0.5	10/16/2001	SV

Approved Pam Dilsizian

QA Officer

Date Approved: 12/24/2001

Environmental Laboratories

emailed

Copy 12/24/2001

NYS ELAP # 10108

Page 2 of 6

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Sample NYC DEP/KENSICO LAB
19 WESTLAKE DRIVE
VALHALLA, N.Y.

Attention: Steve Schindler

Received EP

Bottle c284,p9153,9107,9112,q9147,91

Collected By: NEWMAN

Comment:

Sample DEL18

ID of Source:

Collection 10/15/2001 AT

Submitted 10/15/2001 AT 2:10:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD25155

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
524.2	1,2,3-trichlorobenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,2,3-trichloropropane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,2,4-trichlorobenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,2,4-trimethylbenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,2-dichlorobenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,2-dichloroethane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,2-dichloropropane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,3,5-trimethylbenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,3-dichlorobenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,3-dichloropropane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	1,4-dichlorobenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	2,2-dichloropropane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	2-butanone (MEK)	< MDL	ug/L	2.0	10/16/2001	SV
524.2	2-chlorotoluene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	4-chlorotoluene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Benzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Bromobenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Bromochloromethane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Bromomethane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Carbon tetrachloride	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Chlorobenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Chloroethane	< MDL	ug/L	0.5	10/16/2001	SV

Approved Pam Dilsizian

QA Officer

Date Approved: 12/24/2001

Environmental Laboratories

emailed

Copy 12/24/2001

NYS ELAP # 10108

Page 3 of 6

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Sample NYC DEP/KENSICO LAB
19 WESTLAKE DRIVE
VALHALLA, N.Y.

Attention: Steve Schindler
Received EP

Bottle c284,p9153,9107,9112,q9147,91

Collected By: NEWMAN
Comment:

Sample DEL18

ID of Source:

Collection 10/15/2001 AT

Submitted 10/15/2001 AT 2:10:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD25155

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
524.2	Chloromethane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	cis-1,2-dichloroethene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	cis-1,3-dichloropropene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Dibromomethane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Dichlorodifluoromethane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Ethylbenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Hexachlorobutadiene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Isopropylbenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Methyl iso-butyl ketone (MIBK)	< MDL	ug/L	1.0	10/16/2001	SV
524.2	Methyl tert-butyl ether (MTBE)	< MDL	ug/L	1.0	10/16/2001	SV
524.2	Methylene Chloride	< MDL	ug/L	0.5	10/16/2001	SV
524.2	N-butylbenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	N-propylbenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Naphthalene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	O-xylene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	P & M-xylene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	P-isopropyltoluene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	SEC-butylbenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Styrene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	TERT-butylbenzene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Tetrachloroethene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Toluene	< MDL	ug/L	0.5	10/16/2001	SV

Approved Pam Dilsizian

QA Officer

Date Approved: 12/24/2001

Environmental Laboratories

emailed

Copy 12/24/2001

NYS ELAP # 10108

Page 4 of 6

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Sample NYC DEP/KENSICO LAB
19 WESTLAKE DRIVE
VALHALLA, N.Y.

Attention: Steve Schindler

Received EP

Bottle c284,p9153,9107,9112,q9147,91

Collected By: NEWMAN

Comment:

Sample DEL18

ID of Source:

Collection 10/15/2001 AT

Submitted 10/15/2001 AT 2:10:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD25155

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
524.2	trans-1,2-dichloroethene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	trans-1,3-dichloropropene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Trichloroethene	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Trichlorofluoromethane	< MDL	ug/L	0.5	10/16/2001	SV
524.2	Vinyl chloride	< MDL	ug/L	0.5	10/16/2001	SV
<u>Organic Chemicals - 525.2</u>						
	2,4-Dinitrotoluene	< MDL	ug/L	2.0	10/25/2001	DLV
	2,6-Dinitrotoluene	< MDL	ug/L	2.0	10/25/2001	DLV
	4,4-DDE	< MDL	ug/L	0.80	10/25/2001	DLV
	Acetochlor	< MDL	ug/L	2.0	10/25/2001	DLV
	Alachlor	< MDL	ug/L	0.20	10/25/2001	DLV
	Atrazine	< MDL	ug/L	0.20	10/25/2001	DLV
	Benzo(a)pyrene	< MDL	ug/L	0.20	10/25/2001	DLV
	bis(2-Ethylhexyl)adipate	< MDL	ug/L	0.60	10/25/2001	DLV
	bis(2-Ethylhexyl)phthalate	0.71	ug/L	0.60	10/25/2001	DLV
	Butachlor	< MDL	ug/L	0.40	10/25/2001	DLV
	EPTC	< MDL	ug/L	1.0	10/25/2001	DLV
	Hexachlorobenzene	< MDL	ug/L	0.10	10/25/2001	DLV
	Hexachlorocyclopentadiene	< MDL	ug/L	0.10	10/25/2001	DLV
	Metolachlor	< MDL	ug/L	0.75	10/25/2001	DLV
	Metribuzin	< MDL	ug/L	0.15	10/25/2001	DLV
	Molinate	< MDL	ug/L	0.90	10/25/2001	DLV

Approved Pam Dilsizian

QA Officer

Date Approved: 12/24/2001

Environmental Laboratories

emailed

Copy 12/24/2001

NYS ELAP # 10108

Page 5 of 6

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Sample NYC DEP/KENSICO LAB
19 WESTLAKE DRIVE
VALHALLA, N.Y.

Attention: Steve Schindler

Received EP

Bottle c284,p9153,9107,9112,q9147,91

Collected By: NEWMAN

Comment:

Sample DEL18

ID of Source:

Collection 10/15/2001 AT

Submitted 10/15/2001 AT 2:10:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD25155

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
	Simazine	< MDL	ug/L	0.40	10/25/2001	DLV
	Terbacil	< MDL	ug/L	2.0	10/25/2001	DLV

Recoveries for 3 of the 18 compounds in the MDL and LCS Standards were outside of their acceptance ranges.

Recoveries for 4 of the 18 compounds in the Spiked Sample were below their acceptance ranges.

GCMS Scan For Unknowns

Other unknown compounds	.	ug/L	11/28/2001	DLV
-------------------------	---	------	------------	-----

A scan of the extract prepared for the 508 analysis, from 35-550 amu using a mass spectrometer, does not reveal the presence of non-target compounds, except for the presence of the pesticide DDE.

Approved Pam Dilsizian

QA Officer

Date Approved: 12/24/2001

Environmental Laboratories

emailed

Copy 12/24/2001

NYS ELAP # 10108

Page 6 of 6

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Sample NYC DEP/KENSICO LAB
19 WESTLAKE DRIVE
VALHALLA, N.Y.

Attention: Steve Schindler

Received EP

Bottle c283,p9136,9165,9041,q7149,91

Collected By: NEWMAN

Comment:

Sample CROGH

ID of Source:

Collection 10/15/2001 AT

Submitted 10/15/2001 AT 2:10:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD25156

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
<u>Inorganics</u>						
EPA94	Beryllium	< MDL	ug/L	2	11/1/2001	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	11/1/2001	PAD
EPA94	Molybdenum	< MDL	ug/L	50.0	11/1/2001	PAD
EPA'79	Potable Metals Digestion	Completed			10/24/2001	PAD
SM18	Antimony	< MDL	ug/L	6	10/30/2001	PAD
SM18	Tin	< MDL	ug/L	100	10/19/2001	PAD
SM18	Titanium	< MDL	ug/L	100	10/25/2001	PAD
EPA94	Thallium	< MDL	ug/L	2	10/18/2001	PAD
EPA94	Vanadium	< MDL	ug/L	20.0	11/1/2001	PAD
<u>Organics</u>						
<u>Chlorine Pesticides - 508</u>						
	4,4-DDE	< MDL	ug/L	0.080	12/4/2001	DLV
	Aldrin	< MDL	ug/L	0.010	12/4/2001	DLV
	Chlordane	< MDL	ug/L	0.40	12/4/2001	DLV
	Dieldrin	< MDL	ug/L	0.016	12/4/2001	DLV
	Endrin	< MDL	ug/L	0.020	12/4/2001	DLV

Approved Pam Dilsizian

QA Officer

Date Approved: 12/6/2001

Environmental Laboratories

emailed

Original 12/6/2001

NYS ELAP # 10108

Page 1 of 6

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Attention: Steve Schindler

Received EP

Bottle c283,p9136,9165,9041,q7149,91

Collected By: NEWMAN

Comment:

Sample NYC DEP/KENSICO LAB
19 WESTLAKE DRIVE
VALHALLA, N.Y.

Sample CROGH

ID of Source:

Collection 10/15/2001 AT

Submitted 10/15/2001 AT 2:10:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD25156

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
	Heptachlor	< MDL	ug/L	0.020	12/4/2001	DLV
	Heptachlor Epoxide	< MDL	ug/L	0.030	12/4/2001	DLV
	Lindane	< MDL	ug/L	0.030	12/4/2001	DLV
	Methoxychlor	< MDL	ug/L	0.039	12/4/2001	DLV
	PCB's	< MDL	ug/L	0.20	12/4/2001	DLV
	Propachlor	< MDL	ug/L	0.040	12/4/2001	DLV
	Toxaphene	< MDL	ug/L	0.40	12/4/2001	DLV

The recovery for Toxaphene in the Spiked Sample was below its acceptance range.

Volatile Organic Compounds

524.2	*THM-Bromodichloromethane	< MDL	ug/L	.5	10/16/2020	SV
524.2	*THM-Bromoform	< MDL	ug/L	2.0	10/16/2020	SV
524.2	*THM-Chloroform	< MDL	ug/L	.5	10/16/2020	SV
524.2	*THM-Dibromochloromethane	< MDL	ug/L	2.0	10/16/2020	SV
524.2	1,1,1,2-tetrachloroethane	< MDL	ug/L	.5	10/16/2020	SV
524.2	1,1,1- trichloroethane	< MDL	ug/L	.5	10/16/2020	SV
524.2	1,1,2,2-tetrachloroethane	< MDL	ug/L	.5	10/16/2020	SV
524.2	1,1,2-trichloroethane	< MDL	ug/L	.5	10/16/2020	SV
524.2	1,1-dichloroethane	< MDL	ug/L	.5	10/16/2020	SV
524.2	1,1-Dichloroethene	< MDL	ug/L	.5	10/16/2020	SV
524.2	1,1-dichloropropene	< MDL	ug/L	.5	10/16/2020	SV
524.2	1,2,3-trichlorobenzene	< MDL	ug/L	.5	10/16/2020	SV
524.2	1,2,3-trichloropropane	< MDL	ug/L	.5	10/16/2020	SV

Approved Pam Dilsizian

QA Officer

Date Approved: 12/6/2001

Environmental Laboratories

emailed

Original 12/6/2001

NYS ELAP # 10108

Page 2 of 6

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Attention: Steve Schindler

Received EP

Bottle c283,p9136,9165,9041,q7149,91

Collected By: NEWMAN

Comment:

Sample NYC DEP/KENSICO LAB
19 WESTLAKE DRIVE
VALHALLA, N.Y.

Sample CROGH

ID of Source:

Collection 10/15/2001 AT

Submitted 10/15/2001 AT 2:10:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD25156

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
524.2	1,2,4-trichlorobenzene	< MDL	ug/L	.5	10/16/2020	SV
524.2	1,2,4-trimethylbenzene	< MDL	ug/L	.5	10/16/2020	SV
524.2	1,2-dichlorobenzene	< MDL	ug/L	.5	10/16/2020	SV
524.2	1,2-dichloroethane	< MDL	ug/L	.5	10/16/2020	SV
524.2	1,2-dichloropropane	< MDL	ug/L	.5	10/16/2020	SV
524.2	1,3,5-trimethylbenzene	< MDL	ug/L	.5	10/16/2020	SV
524.2	1,3-dichlorobenzene	< MDL	ug/L	.5	10/16/2020	SV
524.2	1,3-dichloropropane	< MDL	ug/L	.5	10/16/2020	SV
524.2	1,4-dichlorobenzene	< MDL	ug/L	.5	10/16/2020	SV
524.2	2,2-dichloropropane	< MDL	ug/L	.5	10/16/2020	SV
524.2	2-butanone (MEK)	< MDL	ug/L	2.0	10/16/2020	SV
524.2	2-chlorotoluene	< MDL	ug/L	.5	10/16/2020	SV
524.2	4-chlorotoluene	< MDL	ug/L	.5	10/16/2020	SV
524.2	Benzene	< MDL	ug/L	.5	10/16/2020	SV
524.2	Bromobenzene	< MDL	ug/L	.5	10/16/2020	SV
524.2	Bromochloromethane	< MDL	ug/L	.5	10/16/2020	SV
524.2	Bromomethane	< MDL	ug/L	.5	10/16/2020	SV
524.2	Carbon tetrachloride	< MDL	ug/L	.5	10/16/2020	SV
524.2	Chlorobenzene	< MDL	ug/L	.5	10/16/2020	SV
524.2	Chloroethane	< MDL	ug/L	.5	10/16/2020	SV
524.2	Chloromethane	< MDL	ug/L	.5	10/16/2020	SV
524.2	cis-1,2-dichloroethene	< MDL	ug/L	.5	10/16/2020	SV

Approved Pam Dilsizian

QA Officer

Date Approved: 12/6/2001

Environmental Laboratories

emailed

Original 12/6/2001

NYS ELAP # 10108

Page 3 of 6

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Attention: Steve Schindler

Received EP

Bottle c283,p9136,9165,9041,q7149,91

Collected By: NEWMAN

Comment:

Sample NYC DEP/KENSICO LAB
19 WESTLAKE DRIVE
VALHALLA, N.Y.

Sample CROGH

ID of Source:

Collection 10/15/2001 AT

Submitted 10/15/2001 AT 2:10:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD25156

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
524.2	cis-1,3-dichloropropene	< MDL	ug/L	.5	10/16/2020	SV
524.2	Dibromomethane	< MDL	ug/L	.5	10/16/2020	SV
524.2	Dichlorodifluoromethane	< MDL	ug/L	.5	10/16/2020	SV
524.2	Ethylbenzene	< MDL	ug/L	.5	10/16/2020	SV
524.2	Hexachlorobutadiene	< MDL	ug/L	.5	10/16/2020	SV
524.2	Isopropylbenzene	< MDL	ug/L	.5	10/16/2020	SV
524.2	Methyl iso-butyl ketone (MIBK)	< MDL	ug/L	1.0	10/16/2020	SV
524.2	Methyl tert-butyl ether (MTBE)	< MDL	ug/L	1.0	10/16/2020	SV
524.2	Methylene Chloride	< MDL	ug/L	.5	10/16/2020	SV
524.2	N-butylbenzene	< MDL	ug/L	.5	10/16/2020	SV
524.2	N-propylbenzene	< MDL	ug/L	.5	10/16/2020	SV
524.2	Naphthalene	< MDL	ug/L	.5	10/16/2020	SV
524.2	O-xylene	< MDL	ug/L	.5	10/16/2020	SV
524.2	P & M-xylene	< MDL	ug/L	.5	10/16/2020	SV
524.2	P-isopropyltoluene	< MDL	ug/L	.5	10/16/2020	SV
524.2	SEC-butylbenzene	< MDL	ug/L	.5	10/16/2020	SV
524.2	Styrene	< MDL	ug/L	.5	10/16/2020	SV
524.2	TERT-butylbenzene	< MDL	ug/L	.5	10/16/2020	SV
524.2	Tetrachloroethene	< MDL	ug/L	.5	10/16/2020	SV
524.2	Toluene	< MDL	ug/L	.5	10/16/2020	SV
524.2	trans-1,2-dichloroethene	< MDL	ug/L	.5	10/16/2020	SV
524.2	trans-1,3-dichloropropene	< MDL	ug/L	.5	10/16/2020	SV

Approved Pam Dilsizian

QA Officer

Date Approved: 12/6/2001

Environmental Laboratories

emailed

Original 12/6/2001

NYS ELAP # 10108

Page 4 of 6

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Sample NYC DEP/KENSICO LAB
19 WESTLAKE DRIVE
VALHALLA, N.Y.

Attention: Steve Schindler

Received EP

Bottle c283,p9136,9165,9041,q7149,91

Collected By: NEWMAN

Comment:

Sample CROGH

ID of Source:

Collection 10/15/2001 AT

Submitted 10/15/2001 AT 2:10:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD25156

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
524.2	Trichloroethene	< MDL	ug/L	.5	10/16/2001	SV
524.2	Trichlorofluoromethane	< MDL	ug/L	.5	10/16/2001	SV
524.2	Vinyl chloride	< MDL	ug/L	.5	10/16/2001	SV
<u>Organic Chemicals - 525.2</u>						
	2,4-Dinitrotoluene	< MDL	ug/L	2.0	10/26/2001	DLV
	2,6-Dinitrotoluene	< MDL	ug/L	2.0	10/26/2001	DLV
	4,4-DDE	< MDL	ug/L	0.80	10/26/2001	DLV
	Acetochlor	< MDL	ug/L	2.0	10/26/2001	DLV
	Alachlor	< MDL	ug/L	0.20	10/26/2001	DLV
	Atrazine	< MDL	ug/L	0.20	10/26/2001	DLV
	Benzo(a)pyrene	< MDL	ug/L	0.20	10/26/2001	DLV
	bis(2-Ethylhexyl)adipate	< MDL	ug/L	0.60	10/26/2001	DLV
	bis(2-Ethylhexyl)phthalate	0.62	ug/L	0.60	10/26/2001	DLV
	Butachlor	< MDL	ug/L	0.40	10/26/2001	DLV
	EPTC	< MDL	ug/L	1.0	10/26/2001	DLV
	Hexachlorobenzene	< MDL	ug/L	0.10	10/26/2001	DLV
	Hexachlorocyclopentadiene	< MDL	ug/L	0.10	10/26/2001	DLV
	Metolachlor	< MDL	ug/L	0.75	10/26/2001	DLV
	Metribuzin	< MDL	ug/L	0.15	10/26/2001	DLV
	Molinate	< MDL	ug/L	0.90	10/26/2001	DLV
	Simazine	< MDL	ug/L	0.40	10/26/2001	DLV
	Terbacil	< MDL	ug/L	2.0	10/26/2001	DLV

Approved Pam Dilsizian

QA Officer

Date Approved: 12/6/2001

Environmental Laboratories

emailed

Original 12/6/2001

NYS ELAP # 10108

Page 5 of 6

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Attention: Steve Schindler

Received EP

Bottle c283,p9136,9165,9041,q7149,91

Collected By: NEWMAN

Comment:

Sample NYC DEP/KENSICO LAB
19 WESTLAKE DRIVE
VALHALLA, N.Y.

Sample CROGH

ID of Source:

Collection 10/15/2001 AT

Submitted 10/15/2001 AT 2:10:00

Sample Type: POT_DW

PWS No.:

Source Code: 000

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AD25156

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
--------	------------------	---------	-------	-----	-------------	-----------

Recoveries for 4 of the 18 compounds in the MDL Standard were outside of their acceptance ranges. Recoveries for 3 of the 18 compounds in the LCS Standard were outside of their acceptance ranges. Recoveries for 3 of the 18 compounds in the Spiked Sample were outside of their acceptance ranges.

GCMS Scan For Unknowns

Other unknown compounds	.	ug/L	11/28/2001	DLV
-------------------------	---	------	------------	-----

A scan of the extract prepared for the 508 analysis, from 35-550 amu using a mass spectrometer, does not reveal the presence of non-target compounds.

Approved Pam Dilsizian

QA Officer

Date Approved: 12/6/2001

Environmental Laboratories

emailed

Original 12/6/2001

NYS ELAP # 10108

Page 6 of 6

No log in

Dell

10/15/01

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: NYCDEP Bureau Water
465 Columbus Ave.
Valhalla, NY 10595

Sample FIELD BLANK

Attention: Steve Schindler
Received KM
Bottle q9147
Collected By: LAB
Comment:

Sample FB 10/15/01
ID of Source:
Collection 10/15/2001 AT
Submitted 10/15/2001 AT 9:31:00
Sample Type: POT_DW
PWS No.: **Source Code:** 000

Type Descriptor:
Sample chilled on arrival? : YES

Report

Sample No. AD29752

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Organics						
<u>Chlorine Pesticides - 508</u>						
	4,4-DDE	< MDL	ug/L	0.080	12/21/2001	DLV
	Aldrin	< MDL	ug/L	0.010	12/21/2001	DLV
	Chlordane	< MDL	ug/L	0.40	12/21/2001	DLV
	Dieldrin	< MDL	ug/L	0.016	12/21/2001	DLV
	Endrin	< MDL	ug/L	0.020	12/21/2001	DLV
	Heptachlor	< MDL	ug/L	0.020	12/21/2001	DLV
	Heptachlor Epoxide	< MDL	ug/L	0.030	12/21/2001	DLV
	Lindane	< MDL	ug/L	0.030	12/21/2001	DLV
	Methoxychlor	< MDL	ug/L	0.039	12/21/2001	DLV
	PCB's	< MDL	ug/L	0.20	12/21/2001	DLV
	Propachlor	< MDL	ug/L	0.040	12/21/2001	DLV
	Toxaphene	< MDL	ug/L	0.40	12/21/2001	DLV

This is the Field Blank for sample AD25155, which was re-analyzed to confirm the presence or absence of 4,4'-DDE.
The recovery for the Surrogate Standard in this Field Blank was above its acceptance range.

Approved Pam Dilsizian

QA Officer

Date Approved: 12/24/2001

Environmental Laboratories

emailed

Original 12/24/2001

NYS ELAP # 10108

Page 1 of 1