

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received LK

Bottle C1479, Q2554,2568,P2717,2724

Collected By: TSUI

Comment:

Sample 3

Sample 3

ID of Source:

Collection 1/31/2002 AT 10:21:00

Submitted 1/31/2002 AT 11:28:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : YES

Report

Sample No. AE02453

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
<u>ICPMS Metals for NEWQ</u>						
EPA94	Antimony	< MDL	ug/L	0.1	2/4/2002	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Barium	13.2	ug/L	0.1	2/4/2002	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Chromium	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Copper	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Lead	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Manganese	19.1	ug/L	1.0	2/4/2002	PAD
EPA94	Nickel	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Selenium	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Silver	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Zinc	< MDL	ug/L	2.0	2/4/2002	PAD
Ag and Se did not recover. Results are suspect due to unsolved matrix effects						
EPA94200.	Aluminum	< MDL	ug/L	50.0	3/11/2002	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	3/11/2002	PAD
EPA94	Iron	23	ug/L	20.0	3/11/2002	PAD
EPA79	Mercury	< MDL	ug/L	0.40	2/5/2002	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 3/13/2002

Environmental Laboratories

emailed

Original 3/13/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received LK

Bottle C1479, Q2554,2568,P2717,2724

Collected By: TSUI

Comment:

Sample 3

Sample 3

ID of Source:

Collection 1/31/2002 AT 10:21:00

Submitted 1/31/2002 AT 11:28:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : YES

Report

Sample No. AE02453

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	3/11/2002	PAD
EPA'79	Potable Metals Digestion	Completed			2/4/2002	PAD
SM18	Tin	< MDL	ug/L	100	2/21/2002	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds

ug/L

3/7/2002

DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds. The recoveries for 2 of the 43 compounds in the LCS Standard were above their acceptance ranges, while 2 were below.

Approved Pam Dilsizian

QA Officer

Date Approved: 3/13/2002

Environmental Laboratories

emailed

Original 3/13/2002

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Sample 58

Distribution Laboratory
96-05 Horace Harding Ex.
Corona N.Y. 11368
Dr. Ian Hurley

Received LK

Bottle C1506, Q2522,2520,P2659,2663

Collected By: TSUI

Comment:

Sample 58

ID of Source:

Collection 1/31/2002 AT 9:43:00

Submitted 1/31/2002 AT 11:28:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : YES

Report

Sample No. AE02454

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
ICPMS Metals for NEWQ						
EPA94	Antimony	0.13	ug/L	0.1	2/4/2002	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Barium	12.9	ug/L	0.1	2/4/2002	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Chromium	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Copper	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Lead	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Manganese	16.7	ug/L	1.0	2/4/2002	PAD
EPA94	Nickel	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Selenium	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Silver	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Zinc	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	3/11/2002	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	3/11/2002	PAD
EPA94	Iron	< MDL	ug/L	20.0	3/11/2002	PAD
EPA79	Mercury	< MDL	ug/L	0.40	2/5/2002	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 3/13/2002

Environmental Laboratories

emailed

Original 3/13/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received LK

Bottle C1506, Q2522,2520,P2659,2663

Collected By: TSUI

Comment:

Sample 58

Sample 58

ID of Source:

Collection 1/31/2002 AT 9:43:00

Submitted 1/31/2002 AT 11:28:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : YES

Report

Sample No. AE02454

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	3/11/2002	PAD
EPA'79	Potable Metals Digestion	Completed			2/4/2002	PAD
SM18	Tin	< MDL	ug/L	100	2/21/2002	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds

ug/L

3/7/2002

DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds. The recoveries for 2 of the 43 compounds in the LCS Standard were above their acceptance ranges, while 2 were below.

Approved Pam Dilsizian

QA Officer

Date Approved: 3/13/2002

Environmental Laboratories

emailed

Original 3/13/2002

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
Received LK
Bottle C1402,Q211,P797,789
Collected By: AHMAD
Comment: Q587 BROKEN IN TRANSIT

Sample 03360

Sample 13550

ID of Source:

Collection 1/30/2002 AT 9:48:00

Submitted 1/31/2002 AT 11:28:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : YES

Report

Sample No. AE02455

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
<u>ICPMS Metals for NEWQ</u>						
EPA94	Antimony	< MDL	ug/L	0.1	2/4/2002	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Barium	12.7	ug/L	0.1	2/4/2002	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Chromium	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Copper	4.7	ug/L	1.0	2/4/2002	PAD
EPA94	Lead	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Manganese	18.0	ug/L	1.0	2/4/2002	PAD
EPA94	Nickel	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Selenium	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Silver	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Zinc	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	3/11/2002	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	3/11/2002	PAD
EPA94	Iron	43	ug/L	20.0	3/11/2002	PAD
EPA79	Mercury	< MDL	ug/L	0.40	2/5/2002	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 3/13/2002

Environmental Laboratories

emailed

Original 3/13/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
Received LK
Bottle C1402,Q211,P797,789
Collected By: AHMAD
Comment: Q587 BROKEN IN TRANSIT

Sample 03360

Sample 13550
ID of Source:
Collection 1/30/2002 AT 9:48:00
Submitted 1/31/2002 AT 11:28:00
Sample Type: POT_DW
PWS No.: **Source Code:**

Type Descriptor:
Sample chilled on arrival? : YES

Report

Sample No. AE02455

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	3/11/2002	PAD
EPA'79	Potable Metals Digestion	Completed			2/4/2002	PAD
SM18	Tin	< MDL	ug/L	100	2/21/2002	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds	ug/L	3/7/2002	DLV
-------------------------	------	----------	-----

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds. The recoveries for 2 of the 43 compounds in the LCS Standard were above their acceptance ranges, while 2 were below.

Approved Pam Dilsizian QA Officer

Date Approved: 3/13/2002

Environmental Laboratories

emailed

Original 3/13/2002

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
Received LK
Bottle C1284, Q2259,2253,P2488,2485
Collected By: KAGZI
Comment:

Sample 03371

Sample 24950
ID of Source:
Collection 1/30/2002 AT 8:44:00
Submitted 1/31/2002 AT 11:28:00
Sample Type: POT_DW
PWS No.: **Source Code:**

Type Descriptor:
Sample chilled on arrival? : YES

Report

Sample No. AE02456

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
<u>ICPMS Metals for NEWQ</u>						
EPA94	Antimony	< MDL	ug/L	0.1	2/4/2002	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Barium	14.2	ug/L	0.1	2/4/2002	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Chromium	1.1	ug/L	1.0	2/4/2002	PAD
EPA94	Copper	6.7	ug/L	1.0	2/4/2002	PAD
EPA94	Lead	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Manganese	20.0	ug/L	1.0	2/4/2002	PAD
EPA94	Nickel	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Selenium	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Silver	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Zinc	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	3/11/2002	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	3/11/2002	PAD
EPA94	Iron	27	ug/L	20.0	3/11/2002	PAD
EPA79	Mercury	< MDL	ug/L	0.40	2/5/2002	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 3/13/2002

Environmental Laboratories

emailed

Original 3/13/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory
96-05 Horace Harding Ex.
Corona N.Y. 11368
Dr. Ian Hurley

Received LK

Bottle C1284, Q2259,2253,P2488,2485

Collected By: KAGZI

Comment:

Sample 03371

Sample 24950

ID of Source:

Collection 1/30/2002 AT 8:44:00

Submitted 1/31/2002 AT 11:28:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : YES

Report

Sample No. AE02456

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	3/11/2002	PAD
EPA'79	Potable Metals Digestion	Completed			2/4/2002	PAD
SM18	Tin	< MDL	ug/L	100	2/21/2002	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds

ug/L

3/7/2002

DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds. The recoveries for 2 of the 43 compounds in the LCS Standard were above their acceptance ranges, while 2 were below.

Approved Pam Dilsizian

QA Officer

Date Approved: 3/13/2002

Environmental Laboratories

emailed

Original 3/13/2002

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received LK

Bottle C1269,Q2258,233,P2334,2541

Collected By: GERTSIK

Comment:

Sample 03378

Sample 34850

ID of Source:

Collection 1/30/2002 AT 9:15:00

Submitted 1/31/2002 AT 11:28:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : YES

Report

Sample No. AE02457

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
<u>ICPMS Metals for NEWQ</u>						
EPA94	Antimony	< MDL	ug/L	0.1	2/4/2002	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Barium	16.2	ug/L	0.1	2/4/2002	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Chromium	1.1	ug/L	1.0	2/4/2002	PAD
EPA94	Copper	8.0	ug/L	1.0	2/4/2002	PAD
EPA94	Lead	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Manganese	28.9	ug/L	1.0	2/4/2002	PAD
EPA94	Nickel	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Selenium	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Silver	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Zinc	2.6	ug/L	2.0	2/4/2002	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	3/11/2002	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	3/11/2002	PAD
EPA94	Iron	36	ug/L	20.0	3/11/2002	PAD
EPA79	Mercury	< MDL	ug/L	0.40	2/5/2002	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 3/13/2002

Environmental Laboratories

emailed

Original 3/13/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received LK

Bottle C1269,Q2258,233,P2334,2541

Collected By: GERTSIK

Comment:

Sample 03378

Sample 34850

ID of Source:

Collection 1/30/2002 AT 9:15:00

Submitted 1/31/2002 AT 11:28:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : YES

Report

Sample No. AE02457

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	3/11/2002	PAD
EPA'79	Potable Metals Digestion	Completed			2/4/2002	PAD
SM18	Tin	< MDL	ug/L	100	2/21/2002	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds

ug/L

3/7/2002

DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds. The recoveries for 9 of the 43 compounds in the LCS Standard were below their acceptance ranges.

Approved Pam Dilsizian

QA Officer

Date Approved: 3/13/2002

Environmental Laboratories

emailed

Original 3/13/2002

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory
96-05 Horace Harding Ex.
Corona N.Y. 11368
Dr. Ian Hurley

Received LK

Bottle C1403,Q228,184,P2621,796

Collected By: SPERANZA

Comment:

Sample 03391

Sample 44550

ID of Source:

Collection 1/30/2002 AT 8:26:00

Submitted 1/31/2002 AT 11:28:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : YES

Report

Sample No. AE02458

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
<u>ICPMS Metals for NEWQ</u>						
EPA94	Antimony	< MDL	ug/L	0.1	2/4/2002	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Barium	13.4	ug/L	0.1	2/4/2002	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Chromium	1.1	ug/L	1.0	2/4/2002	PAD
EPA94	Copper	7.1	ug/L	1.0	2/4/2002	PAD
EPA94	Lead	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Manganese	13.9	ug/L	1.0	2/4/2002	PAD
EPA94	Nickel	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Selenium	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Silver	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Zinc	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	3/11/2002	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	3/11/2002	PAD
EPA94	Iron	32	ug/L	20.0	3/11/2002	PAD
EPA79	Mercury	< MDL	ug/L	0.40	2/5/2002	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 4/4/2002

Environmental Laboratories

emailed

Original 4/4/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received LK

Bottle C1403,Q228,184,P2621,796

Collected By: SPERANZA

Comment:

Sample 03391

Sample 44550

ID of Source:

Collection 1/30/2002 AT 8:26:00

Submitted 1/31/2002 AT 11:28:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : YES

Report

Sample No. AE02458

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	3/11/2002	PAD
EPA'79	Potable Metals Digestion	Completed			2/4/2002	PAD
SM18	Tin	< MDL	ug/L	100	2/21/2002	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds

ug/L

4/3/2002

DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds. The recoveries for 9 of the 43 compounds in the LCS Standard were low. This sample was not reextracted. A discount will be applied to its

Approved Pam Dilsizian

QA Officer

Date Approved: 4/4/2002

Environmental Laboratories

emailed

Original 4/4/2002

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received LK

Bottle C1405,Q219,594,P2617,2620

Collected By: BROWN

Comment:

Sample 03399

Sample 52150

ID of Source:

Collection 1/30/2002 AT 10:25:00

Submitted 1/31/2002 AT 11:28:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : YES

Report

Sample No. AE02459

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
<u>ICPMS Metals for NEWQ</u>						
EPA94	Antimony	< MDL	ug/L	0.1	2/4/2002	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Barium	13.5	ug/L	0.1	2/4/2002	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Chromium	1.5	ug/L	1.0	2/4/2002	PAD
EPA94	Copper	6.7	ug/L	1.0	2/4/2002	PAD
EPA94	Lead	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Manganese	15.6	ug/L	1.0	2/4/2002	PAD
EPA94	Nickel	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Selenium	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Silver	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Zinc	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	3/11/2002	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	3/11/2002	PAD
EPA94	Iron	< MDL	ug/L	20.0	3/11/2002	PAD
EPA79	Mercury	< MDL	ug/L	0.40	2/5/2002	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 3/13/2002

Environmental Laboratories

emailed

Original 3/13/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received LK

Bottle C1405,Q219,594,P2617,2620

Collected By: BROWN

Comment:

Sample 03399

Sample 52150

ID of Source:

Collection 1/30/2002 AT 10:25:00

Submitted 1/31/2002 AT 11:28:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : YES

Report

Sample No. AE02459

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	3/11/2002	PAD
EPA'79	Potable Metals Digestion	Completed			2/4/2002	PAD
SM18	Tin	< MDL	ug/L	100	2/21/2002	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds

ug/L

3/7/2002

DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds. The recoveries for 9 of the 43 compounds in the LCS Standard were below their acceptance ranges.

Approved Pam Dilsizian

QA Officer

Date Approved: 3/13/2002

Environmental Laboratories

emailed

Original 3/13/2002

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
 Received LK
 Bottle Q834, P1000
 Collected By: LAB
 Comment:

Sample FBD

Sample FBD

ID of Source:
 Collection 1/31/2002 AT 12:00:00
 Submitted 1/31/2002 AT 11:31:00
 Sample Type: POT_DW
 PWS No.:

Source Code:

Type Descriptor:
 Sample chilled on arrival? : YES

Report

Sample No. AE02460

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Organics						
<u>GCMS Scan For Unknowns</u>						
	Other unknown compounds	.	ug/L		3/7/2002	DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds. The recoveries for 9 of the 43 compounds in the LCS Standard were below their acceptance ranges. This sample is the Field Blank. It is being reextracted and reanalyzed because of its low Surrogate Standard recovery.

Approved Pam Dilsizian

QA Officer

Date Approved: 3/8/2002

Environmental Laboratories

emailed

Original 3/8/2002

NYS ELAP # 10108

Page 1 of 1