

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory
96-05 Horace Harding Ex.
Corona N.Y. 11368
Dr. Ian Hurley

Received KM

Bottle C1469,P2751,2754,Q2582,2579

Collected By: TSUI

Comment:

Sample 3

Sample 3

ID of Source:

Collection 2/1/2002 AT 10:28:00

Submitted 2/1/2002 AT 11:25:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival?: YES

Report

Sample No. AE02561

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
<u>ICPMS Metals for NEWQ</u>						
EPA94	Antimony	< MDL	ug/L	0.1	2/7/2002	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	2/7/2002	PAD
EPA94	Barium	13.4	ug/L	0.1	2/7/2002	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	2/7/2002	PAD
EPA94	Chromium	< MDL	ug/L	1.0	2/7/2002	PAD
EPA94	Copper	< MDL	ug/L	1.0	2/7/2002	PAD
EPA94	Lead	< MDL	ug/L	1.0	2/7/2002	PAD
EPA94	Manganese	22.2	ug/L	1.0	2/7/2002	PAD
EPA94	Nickel	< MDL	ug/L	1.0	2/7/2002	PAD
EPA94	Selenium	< MDL	ug/L	1.0	2/7/2002	PAD
EPA94	Silver	< MDL	ug/L	2.0	2/7/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	2/7/2002	PAD
EPA94	Zinc	< MDL	ug/L	2.0	2/7/2002	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	3/12/2002	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	3/12/2002	PAD
EPA94	Iron	33	ug/L	20.0	3/12/2002	PAD
EPA79	Mercury	< MDL	ug/L	0.40	2/5/2002	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 3/15/2002

Environmental Laboratories

emailed

Original 3/15/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory
96-05 Horace Harding Ex.
Corona N.Y. 11368
Dr. Ian Hurley

Received KM

Bottle C1469,P2751,2754,Q2582,2579

Collected By: TSUI

Comment:

Sample 3

Sample 3.

ID of Source:

Collection 2/1/2002 AT 10:28:00

Submitted 2/1/2002 AT 11:25:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : YES

Report

Sample No. AE02561

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	3/12/2002	PAD
EPA'79	Potable Metals Digestion	Completed			2/4/2002	PAD
SM18	Tin	< MDL	ug/L	100	2/21/2002	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds	.	ug/L	3/7/2002	DLV
-------------------------	---	------	----------	-----

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds. The recoveries for 9 of the 43 compounds in the LCS Standard were below their acceptance ranges. Because of the low Surrogate Standard recovery this sample will be reextracted and reanalyzed.

Approved Pam Dilsizian

QA Officer

Date Approved: 3/15/2002

Environmental Laboratories

emailed

Original 3/15/2002

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received KM

Bottle C1456,P2768,2746,Q2593,2592

Collected By: VASSILEV/TSUI

Comment:

Sample 58

Sample 58

ID of Source:

Collection 2/1/2002 AT 9:56:00

Submitted 2/1/2002 AT 11:25:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : YES

Report

Sample No. AE02562

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
<u>ICPMS Metals for NEWQ</u>						
EPA94	Antimony	0.16	ug/L	0.1	2/7/2002	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	2/7/2002	PAD
EPA94	Barium	12.6	ug/L	0.1	2/7/2002	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	2/7/2002	PAD
EPA94	Chromium	< MDL	ug/L	1.0	2/7/2002	PAD
EPA94	Copper	< MDL	ug/L	1.0	2/7/2002	PAD
EPA94	Lead	< MDL	ug/L	1.0	2/7/2002	PAD
EPA94	Manganese	19.6	ug/L	1.0	2/7/2002	PAD
EPA94	Nickel	< MDL	ug/L	1.0	2/7/2002	PAD
EPA94	Selenium	< MDL	ug/L	1.0	2/7/2002	PAD
EPA94	Silver	< MDL	ug/L	2.0	2/7/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	2/7/2002	PAD
EPA94	Zinc	< MDL	ug/L	2.0	2/7/2002	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	3/12/2002	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	3/12/2002	PAD
EPA94	Iron	35	ug/L	20.0	3/12/2002	PAD
EPA79	Mercury	< MDL	ug/L	0.40	2/5/2002	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 3/13/2002

Environmental Laboratories

emailed

Original 3/13/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received KM

Bottle C1456,P2768,2746,Q2593,2592

Collected By: VASSILEV/TSUI

Comment:

Sample 58

Sample 58

ID of Source:

Collection 2/1/2002 AT 9:56:00

Submitted 2/1/2002 AT 11:25:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : YES

Report

Sample No. AE02562

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	3/12/2002	PAD
EPA79	Potable Metals Digestion	Completed			2/4/2002	PAD
SM18	Tin	< MDL	ug/L	100	2/21/2002	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds

ug/L

3/7/2002

DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds. The recoveries for 9 of the 43 compounds in the LCS Standard were below their acceptance ranges.

Approved Pam Dilsizian

QA Officer

Date Approved: 3/13/2002

Environmental Laboratories

emailed

Original 3/13/2002

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
Received KM
Bottle C1478,P2705,2710,Q2539,2552
Collected By: PATEL
Comment:

Sample 03469

Sample 12550
ID of Source:
Collection 1/31/2002 AT 10:06:00
Submitted 2/1/2002 AT 11:25:00
Sample Type: POT_DW
PWS No.: **Source Code:**

Type Descriptor:
Sample chilled on arrival? : YES

Report

Sample No. AE02563

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
<u>ICPMS Metals for NEWQ</u>						
EPA94	Antimony	< MDL	ug/L	0.1	2/4/2002	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Barium	13.2	ug/L	0.1	2/4/2002	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Chromium	1.1	ug/L	1.0	2/4/2002	PAD
EPA94	Copper	6.8	ug/L	1.0	2/4/2002	PAD
EPA94	Lead	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Manganese	11.6	ug/L	1.0	2/4/2002	PAD
EPA94	Nickel	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Selenium	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Silver	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Zinc	3.2	ug/L	2.0	2/4/2002	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	3/12/2002	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	3/12/2002	PAD
EPA94	Iron	32	ug/L	20.0	3/12/2002	PAD
EPA79	Mercury	< MDL	ug/L	0.40	2/5/2002	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 3/13/2002

Environmental Laboratories

emailed

Original 3/13/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
 Received KM
 Bottle C1478,P2705,2710,Q2539,2552
 Collected By: PATEL
 Comment:

Sample 03469

 Sample 12550
 ID of Source:
 Collection 1/31/2002 AT 10:06:00
 Submitted 2/1/2002 AT 11:25:00
 Sample Type: POT_DW
 PWS No.: Source Code:

 Type Descriptor:
 Sample chilled on arrival? : YES

Report

Sample No.		AE02563				
Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	3/12/2002	PAD
EPA'79	Potable Metals Digestion	Completed			2/4/2002	PAD
SM18	Tin	< MDL	ug/L	100	2/21/2002	PAD
Organics						
<u>GCMS Scan For Unknowns</u>						
	Other unknown compounds	.	ug/L		3/7/2002	DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds. The recoveries for 9 of the 43 compounds in the LCS Standard were below their acceptance ranges.

Approved Pam Dilsizian QA Officer

Date Approved: 3/13/2002

Environmental Laboratories

emailed Original 3/13/2002

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received KM

Bottle C1489,P2722,2716,Q2550,2551

Collected By: KAGZI

Comment:

Sample 03508

Sample 24450

ID of Source:

Collection 1/31/2002 AT 8:49:00

Submitted 2/1/2002 AT 11:25:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : YES

Report

Sample No. AE02564

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
<u>ICPMS Metals for NEWQ</u>						
EPA94	Antimony	< MDL	ug/L	0.1	2/4/2002	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Barium	14.9	ug/L	0.1	2/4/2002	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Chromium	1.0	ug/L	1.0	2/4/2002	PAD
EPA94	Copper	38.7	ug/L	1.0	2/4/2002	PAD
EPA94	Lead	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Manganese	18.2	ug/L	1.0	2/4/2002	PAD
EPA94	Nickel	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Selenium	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Silver	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Zinc	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	3/12/2002	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	3/12/2002	PAD
EPA94	Iron	43	ug/L	20.0	3/12/2002	PAD
EPA79	Mercury	< MDL	ug/L	0.40	2/5/2002	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 3/13/2002

Environmental Laboratories

emailed

Original 3/13/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received KM

Bottle C1489,P2722,2716,Q2550,2551

Collected By: KAGZI

Comment:

Sample 03508

Sample 24450

ID of Source:

Collection 1/31/2002 AT 8:49:00

Submitted 2/1/2002 AT 11:25:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : YES

Report

Sample No. AE02564

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	3/12/2002	PAD
EPA'79	Potable Metals Digestion	Completed			2/4/2002	PAD
SM18	Tin	< MDL	ug/L	100	2/21/2002	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds

ug/L

3/7/2002

DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds. The recoveries for 9 of the 43 compounds in the LCS Standard were below their acceptance ranges.

Approved Pam Dilsizian

QA Officer

Date Approved: 3/13/2002

Environmental Laboratories

emailed

Original 3/13/2002

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Sample 03513

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received KM

Bottle C1471,P2739,2740,Q2569,2567

Collected By: BOLTON

Comment:

Sample 31650

ID of Source:

Collection 1/31/2002 AT 9:25:00

Submitted 2/1/2002 AT 11:25:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : YES

Report

Sample No. AE02565

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
<u>ICPMS Metals for NEWQ</u>						
EPA94	Antimony	< MDL	ug/L	0.1	2/4/2002	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Barium	13.1	ug/L	0.1	2/4/2002	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Chromium	1.2	ug/L	1.0	2/4/2002	PAD
EPA94	Copper	4.9	ug/L	1.0	2/4/2002	PAD
EPA94	Lead	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Manganese	21.6	ug/L	1.0	2/4/2002	PAD
EPA94	Nickel	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Selenium	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Silver	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Zinc	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	3/12/2002	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	3/12/2002	PAD
EPA94	Iron	43	ug/L	20.0	3/12/2002	PAD
EPA79	Mercury	< MDL	ug/L	0.40	2/5/2002	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 3/13/2002

Environmental Laboratories

emailed

Original 3/13/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received KM

Bottle C1471,P2739,2740,Q2569,2567

Collected By: BOLTON

Comment:

Sample 03513

Sample 31650

ID of Source:

Collection 1/31/2002 AT 9:25:00

Submitted 2/1/2002 AT 11:25:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : YES

Report

Sample No. AE02565

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	3/12/2002	PAD
EPA'79	Potable Metals Digestion	Completed			2/4/2002	PAD
SM18	Tin	< MDL	ug/L	100	2/21/2002	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds

ug/L

3/7/2002

DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds. The recoveries for 9 of the 43 compounds in the LCS Standard were below their acceptance ranges. Because of the low Surrogate Standard recovery this sample will be reextracted and reanalyzed.

Approved Pam Dilsizian

QA Officer

Date Approved: 3/13/2002

Environmental Laboratories

emailed

Original 3/13/2002

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received KM

Bottle C1477,P2736,2738,Q2559,2560

Collected By: DESSOURCES

Comment:

Sample 03526

Sample 42350

ID of Source:

Collection 1/31/2002 AT 10:45:00

Submitted 2/1/2002 AT 11:25:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : YES

Report

Sample No. AE02566

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
<u>ICPMS Metals for NEWQ</u>						
EPA94	Antimony	13.5	ug/L	0.1	2/4/2002	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Barium	< MDL	ug/L	0.1	2/4/2002	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Chromium	1.5	ug/L	1.0	2/4/2002	PAD
EPA94	Copper	5.9	ug/L	1.0	2/4/2002	PAD
EPA94	Lead	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Manganese	13.7	ug/L	1.0	2/4/2002	PAD
EPA94	Nickel	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Selenium	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Silver	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Zinc	3.5	ug/L	2.0	2/4/2002	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	3/12/2002	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	3/12/2002	PAD
EPA94	Iron	37	ug/L	20.0	3/12/2002	PAD
EPA79	Mercury	< MDL	ug/L	0.40	2/5/2002	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 3/13/2002

Environmental Laboratories

emailed

Original 3/13/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received KM

Bottle C1477,P2736,2738,Q2559,2560

Collected By: DESSOURCES

Comment:

Sample 03526

Sample 42350

ID of Source:

Collection 1/31/2002 AT 10:45:00

Submitted 2/1/2002 AT 11:25:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : YES

Report

Sample No. AE02566

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	3/12/2002	PAD
EPA'79	Potable Metals Digestion	Completed			2/4/2002	PAD
SM18	Tin	< MDL	ug/L	100	2/21/2002	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds

ug/L

3/7/2002

DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds. The recoveries for 9 of the 43 compounds in the LCS Standard were below their acceptance ranges.

Approved Pam Dilsizian

QA Officer

Date Approved: 3/13/2002

Environmental Laboratories

emailed

Original 3/13/2002

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Sample 03536

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received KM

Bottle C1475,P2723,2734,Q2546,2553

Collected By: BROWN

Comment:

Sample 57750

ID of Source:

Collection 1/31/2002 AT 10:10:00

Submitted 2/1/2002 AT 11:25:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : YES

Report

Sample No. AE02567

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
<u>ICPMS Metals for NEWQ</u>						
EPA94	Antimony	< MDL	ug/L	0.1	2/4/2002	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Barium	13.3	ug/L	0.1	2/4/2002	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Chromium	1.7	ug/L	1.0	2/4/2002	PAD
EPA94	Copper	4.7	ug/L	1.0	2/4/2002	PAD
EPA94	Lead	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Manganese	12.0	ug/L	1.0	2/4/2002	PAD
EPA94	Nickel	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Selenium	1.0	ug/L	1.0	2/4/2002	PAD
EPA94	Silver	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	2/4/2002	PAD
EPA94	Zinc	< MDL	ug/L	2.0	2/4/2002	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	3/12/2002	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	3/12/2002	PAD
EPA94	Iron	28	ug/L	20.0	3/12/2002	PAD
EPA79	Mercury	< MDL	ug/L	0.40	2/5/2002	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 3/26/2002

Environmental Laboratories

emailed

Original 3/26/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
Received KM
Bottle C1475,P2723,2734,Q2546,2553
Collected By: BROWN
Comment:

Sample 03536

Sample 57750
ID of Source:
Collection 1/31/2002 AT 10:10:00
Submitted 2/1/2002 AT 11:25:00
Sample Type: POT_DW
PWS No.: **Source Code:**

Type Descriptor:
Sample chilled on arrival? : YES

Report

Sample No. AE02567

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	3/12/2002	PAD
EPA'79	Potable Metals Digestion	Completed			2/4/2002	PAD
SM18	Tin	< MDL	ug/L	100	2/21/2002	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds	ug/L	3/25/2002	DLV
-------------------------	------	-----------	-----

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds. The recoveries for 8 of the compounds in the LCS Standard are high while 1 was low. New control limits will be calculated.

Approved Pam Dilsizian

QA Officer

Date Approved: 3/26/2002

Environmental Laboratories

emailed Original 3/26/2002

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
Received KM
Bottle P1003,Q163
Collected By: LAB
Comment:

Sample FBD

Sample FBD

ID of Source:
Collection 2/1/2002 AT 12:00:00
Submitted 2/1/2002 AT 11:25:00
Sample Type: POT_DW
PWS No.:

Source Code:

Type Descriptor:
Sample chilled on arrival? : YES

Report

Sample No. AE02568

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Organics						
<u>GCMS Scan For Unknowns</u>						
	Other unknown compounds		ug/L		3/25/2002	DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds. The recoveries for 8 of the compounds in the LCS Standard are high while 1 was low. New control limits will be calculated.

Approved Pam Dilsizian QA Officer

Date Approved: 3/26/2002

Environmental Laboratories

emailed Original 3/26/2002

NYS ELAP # 10108

Page 1 of 1