

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory
96-05 Horace Harding Ex.
Corona N.Y. 11368
Dr. Ian Hurley

Received LK

Bottle C1626,Q2457,2641,2726,852,26

Collected By: PATEL

Comment:

Sample 3

Sample 3

ID of Source:

Collection 2/11/2002 AT 6:05:00

Submitted 2/11/2002 AT 11:03:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : YES

Report

Sample No. AE03281

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
<u>ICPMS Metals for NEWQ</u>						
EPA94	Antimony	< MDL	ug/L	0.1	3/7/2002	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	3/7/2002	PAD
EPA94	Barium	13.9	ug/L	0.1	3/7/2002	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	3/7/2002	PAD
EPA94	Chromium	< MDL	ug/L	1.0	3/7/2002	PAD
EPA94	Copper	< MDL	ug/L	1.0	3/7/2002	PAD
EPA94	Lead	< MDL	ug/L	1.0	3/7/2002	PAD
EPA94	Manganese	21.9	ug/L	1.0	3/7/2002	PAD
EPA94	Nickel	< MDL	ug/L	1.0	3/7/2002	PAD
EPA94	Selenium	< MDL	ug/L	1.0	3/7/2002	PAD
EPA94	Silver	< MDL	ug/L	2.0	3/7/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	3/7/2002	PAD
EPA94	Zinc	< MDL	ug/L	2.0	3/7/2002	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	3/26/2002	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	3/26/2002	PAD
EPA94	Iron	36	ug/L	20.0	3/26/2002	PAD
EPA79	Mercury	< MDL	ug/L	0.40	2/13/2002	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 3/29/2002

Environmental Laboratories

emailed

Original 3/29/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory
96-05 Horace Harding Ex.
Corona N.Y. 11368
Dr. Ian Hurley

Received LK

Bottle C1626,Q2457,2641,2726,852,26

Collected By: PATEL

Comment:

Sample 3

Sample 3

ID of Source:

Collection 2/11/2002 AT 6:05:00

Submitted 2/11/2002 AT 11:03:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : YES

Report

Sample No. AE03281

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	3/26/2002	PAD
EPA'79	Potable Metals Digestion	Completed			2/14/2002	PAD
SM18	Tin	< MDL	ug/L	100	2/25/2002	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds

ug/L

3/29/2002

DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds. The recoveries for the compounds in the LCS Standard were high. New limits will be calculated.

Approved Pam Dilsizian

QA Officer

Date Approved: 3/29/2002

Environmental Laboratories

emailed

Original 3/29/2002

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory
96-05 Horace Harding Ex.
Corona N.Y. 11368
Dr. Ian Hurley

Received LK

Bottle C1439,Q785,2667,P1175,1161

Collected By: PATEL

Comment:

Sample 58

Sample 58

ID of Source:

Collection 2/11/2002 AT 9:25:00

Submitted 2/11/2002 AT 11:03:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : YES

Report

Sample No. AE03282

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
ICPMS Metals for NEWQ						
EPA94	Antimony	< MDL	ug/L	0.1	3/7/2002	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	3/7/2002	PAD
EPA94	Barium	14.0	ug/L	0.1	3/7/2002	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	3/7/2002	PAD
EPA94	Chromium	< MDL	ug/L	1.0	3/7/2002	PAD
EPA94	Copper	< MDL	ug/L	1.0	3/7/2002	PAD
EPA94	Lead	< MDL	ug/L	1.0	3/7/2002	PAD
EPA94	Manganese	19.5	ug/L	1.0	3/7/2002	PAD
EPA94	Nickel	< MDL	ug/L	1.0	3/7/2002	PAD
EPA94	Selenium	< MDL	ug/L	1.0	3/7/2002	PAD
EPA94	Silver	< MDL	ug/L	2.0	3/7/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	3/7/2002	PAD
EPA94	Zinc	< MDL	ug/L	2.0	3/7/2002	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	3/26/2002	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	3/26/2002	PAD
EPA94	Iron	30	ug/L	20.0	3/26/2002	PAD
EPA79	Mercury	< MDL	ug/L	0.40	2/13/2002	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 3/29/2002

Environmental Laboratories

emailed

Original 3/29/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory
96-05 Horace Harding Ex.
Corona N.Y. 11368
Dr. Ian Hurley

Received LK

Bottle C1439,Q785,2667,P1175,1161

Collected By: PATEL

Comment:

Sample 58

Sample 58

ID of Source:

Collection 2/11/2002 AT 9:25:00

Submitted 2/11/2002 AT 11:03:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : YES

Report

Sample No. AE03282

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	3/26/2002	PAD
EPA'79	Potable Metals Digestion	Completed			2/14/2002	PAD
SM18	Tin	< MDL	ug/L	100	2/25/2002	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds

ug/L

3/29/2002

DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds. The recoveries for the compounds in the LCS Standard were high. New limits will be calculated.

Approved Pam Dilsizian

QA Officer

Date Approved: 3/29/2002

Environmental Laboratories

emailed

Original 3/29/2002

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
 Received LK
 Bottle Q862, P1113
 Collected By: LAB
 Comment:

Sample FBD
 Sample FBD
 ID of Source:
 Collection 2/11/2002 AT 12:00:00
 Submitted 2/11/2002 AT 11:04:00
 Sample Type: POT_DW
 PWS No.: Source Code:
 Type Descriptor:
 Sample chilled on arrival? : YES

Report

Sample No. AE03283

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Organics						
<u>GCMS Scan For Unknowns</u>						
	Other unknown compounds	.	ug/L		3/29/2002	DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds. The recoveries for the compounds in the LCS Standard were high. New limits will be calculated.

Approved Pam Dilsizian QA Officer

Date Approved: 3/29/2002

Environmental Laboratories emailed Original 3/29/2002