

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle c1545,p1460,q2808,2809

Collected By: SACKETT

Comment: p1488 broken in transit

Sample 3

Sample 3

ID of Source:

Collection 2/20/2002 AT 10:12:00

Submitted 2/20/2002 AT 11:32:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AE04125

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
<u>ICPMS Metals for NEWQ</u>						
EPA94	Antimony	< MDL	ug/L	0.1	3/14/2002	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	3/14/2002	PAD
EPA94	Barium	14.0	ug/L	0.1	3/14/2002	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	3/14/2002	PAD
EPA94	Chromium	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Copper	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Lead	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Manganese	20.9	ug/L	1.0	3/14/2002	PAD
EPA94	Nickel	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Selenium	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Silver	< MDL	ug/L	2.0	3/14/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Zinc	< MDL	ug/L	2.0	3/14/2002	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	3/29/2002	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	3/29/2002	PAD
EPA94	Iron	24	ug/L	20.0	3/29/2002	PAD
EPA79	Mercury	< MDL	ug/L	0.40	2/21/2002	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 4/2/2002

Environmental Laboratories

emailed

Original 4/2/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle c1545,p1460,q2808,2809

Collected By: SACKETT

Comment: p1488 broken in transit

Sample 3

Sample 3

ID of Source:

Collection 2/20/2002 AT 10:12:00

Submitted 2/20/2002 AT 11:32:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AE04125

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	3/29/2002	PAD
EPA'79	Potable Metals Digestion	Completed			2/21/2002	PAD
SM18	Tin	< MDL	ug/L	100	2/26/2002	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds

ug/L

4/1/2002

DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds.

The recoveries for compounds in the LCS Standard were high. New recovery limits will be calculated.

Approved Pam Dilsizian

QA Officer

Date Approved: 4/2/2002

Environmental Laboratories

emailed

Original 4/2/2002

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory
96-05 Horace Harding Ex.
Corona N.Y. 11368
Dr. Ian Hurley

Received EP

Bottle c1525,p1661,p1667,q1391,1283

Collected By: SACKETT

Comment:

Sample 58

Sample 58

ID of Source:

Collection 2/20/2002 AT 9:40:00

Submitted 2/20/2002 AT 11:32:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AE04126

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
<u>ICPMS Metals for NEWQ</u>						
EPA94	Antimony	< MDL	ug/L	0.1	3/14/2002	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	3/14/2002	PAD
EPA94	Barium	13.9	ug/L	0.1	3/14/2002	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	3/14/2002	PAD
EPA94	Chromium	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Copper	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Lead	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Manganese	21.0	ug/L	1.0	3/14/2002	PAD
EPA94	Nickel	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Selenium	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Silver	< MDL	ug/L	2.0	3/14/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Zinc	< MDL	ug/L	2.0	3/14/2002	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	3/29/2002	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	3/29/2002	PAD
EPA94	Iron	24	ug/L	20.0	3/29/2002	PAD
EPA79	Mercury	< MDL	ug/L	0.40	2/21/2002	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 4/2/2002

Environmental Laboratories

emailed

Original 4/2/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle c1525,p1661,p1667,q1391,1283

Collected By: SACKETT

Comment:

Sample 58

Sample 58

ID of Source:

Collection 2/20/2002 AT 9:40:00

Submitted 2/20/2002 AT 11:32:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AE04126

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	3/29/2002	PAD
EPA'79	Potable Metals Digestion	Completed			2/21/2002	PAD
SM18	Tin	< MDL	ug/L	100	2/26/2002	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds

ug/L

4/1/2002

DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds.

The recoveries for compounds in the LCS Standard were high. New recovery limits will be calculated.

Approved Pam Dilsizian

QA Officer

Date Approved: 4/2/2002

Environmental Laboratories

emailed

Original 4/2/2002

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle c1431,p2794,1162,q313,2368

Collected By: VASSILEV

Comment:

Sample 14350

Sample 13450

ID of Source:

Collection 02/19/02 AT 9:22:00

Submitted 02/20/02 AT 11:32:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Report

Sample No. AE04127

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
ICPMS Metals for NEWQ						
EPA94	Antimony	< MDL	ug/L	0.1	03/14/02	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	03/14/02	PAD
EPA94	Barium	27.4	ug/L	0.1	03/14/02	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	03/14/02	PAD
EPA94	Chromium	1.5	ug/L	1.0	03/14/02	PAD
EPA94	Copper	< MDL	ug/L	1.0	03/14/02	PAD
EPA94	Lead	< MDL	ug/L	1.0	03/14/02	PAD
EPA94	Manganese	61.3	ug/L	1.0	03/14/02	PAD
EPA94	Nickel	10.8	ug/L	1.0	03/14/02	PAD
EPA94	Selenium	1.2	ug/L	1.0	03/14/02	PAD
EPA94	Silver	< MDL	ug/L	2.0	03/14/02	PAD
EPA94	Thallium	< MDL	ug/L	1.0	03/14/02	PAD
EPA94	Zinc	< MDL	ug/L	2.0	03/14/02	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	04/02/02	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	04/02/02	PAD
EPA94	Iron	60	ug/L	20.0	04/02/02	PAD
EPA79	Mercury	< MDL	ug/L	0.40	02/21/02	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 04/10/02

Environmental Laboratories

emailed

Copy 04/16/02

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle c1431,p2794,1162,q313,2368

Collected By: VASSILEV

Comment:

Sample 14350

Sample 13450

ID of Source:

Collection 02/19/02 AT 9:22:00

Submitted 02/20/02 AT 11:32:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Report

Sample No. AE04127

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	04/02/02	PAD
EPA'79	Potable Metals Digestion	Completed			02/21/02	PAD
SM18	Tin	< MDL	ug/L	100	02/26/02	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds

ug/L

04/01/02

DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds.

The recoveries for compounds in the LCS Standard were high. New recovery limits will be calculated.

Approved Pam Dilsizian

QA Officer

Date Approved: 04/10/02

Environmental Laboratories

emailed

Copy 04/16/02

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle c9476,p2840,2842,q2675,2673

Collected By: GERTSIK

Comment:

Sample 25950

Sample 25950

ID of Source:

Collection 2/19/2002 AT 11:34:00

Submitted 2/20/2002 AT 11:32:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AE04128

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
<u>ICPMS Metals for NEWQ</u>						
EPA94	Antimony	< MDL	ug/L	0.1	3/14/2002	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	3/14/2002	PAD
EPA94	Barium	15.6	ug/L	0.1	3/14/2002	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	3/14/2002	PAD
EPA94	Chromium	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Copper	14.2	ug/L	1.0	3/14/2002	PAD
EPA94	Lead	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Manganese	26.2	ug/L	1.0	3/14/2002	PAD
EPA94	Nickel	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Selenium	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Silver	< MDL	ug/L	2.0	3/14/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Zinc	< MDL	ug/L	2.0	3/14/2002	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	4/2/2002	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	4/2/2002	PAD
EPA94	Iron	53	ug/L	20.0	4/2/2002	PAD
EPA79	Mercury	< MDL	ug/L	0.40	2/21/2002	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 4/2/2002

Environmental Laboratories

emailed

Original 4/2/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle c9476,p2840,2842,q2675,2673

Collected By: GERTSIK

Comment:

Sample 25950

Sample 25950

ID of Source:

Collection 2/19/2002 AT 11:34:00

Submitted 2/20/2002 AT 11:32:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AE04128

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	4/2/2002	PAD
EPA'79	Potable Metals Digestion	Completed			2/21/2002	PAD
SM18	Tin	< MDL	ug/L	100	2/26/2002	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds

ug/L

4/1/2002

DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds.

The recoveries for compounds in the LCS Standard were high. New recovery limits will be calculated.

Approved Pam Dilsizian

QA Officer

Date Approved: 4/2/2002

Environmental Laboratories

emailed

Original 4/2/2002

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
Distribution Laboratory
96-05 Horace Harding Ex.
Corona N.Y. 11368
Dr. Ian Hurley

Received EP

Bottle c1598,p13,471,q1309,2802

Collected By: BOLTON

Comment:

Sample 34350

Sample 34350

ID of Source:

Collection 2/19/2002 AT 10:33:00

Submitted 2/20/2002 AT 11:32:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AE04129

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
<u>ICPMS Metals for NEWQ</u>						
EPA94	Antimony	< MDL	ug/L	0.1	3/14/2002	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	3/14/2002	PAD
EPA94	Barium	17.2	ug/L	0.1	3/14/2002	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	3/14/2002	PAD
EPA94	Chromium	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Copper	8.3	ug/L	1.0	3/14/2002	PAD
EPA94	Lead	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Manganese	31.6	ug/L	1.0	3/14/2002	PAD
EPA94	Nickel	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Selenium	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Silver	< MDL	ug/L	2.0	3/14/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Zinc	< MDL	ug/L	2.0	3/14/2002	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	4/2/2002	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	4/2/2002	PAD
EPA94	Iron	44	ug/L	20.0	4/2/2002	PAD
EPA79	Mercury	< MDL	ug/L	0.40	2/21/2002	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 4/2/2002

Environmental Laboratories

emailed

Original 4/2/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
Distribution Laboratory
96-05 Horace Harding Ex.
Corona N.Y. 11368
Dr. Ian Hurley

Received EP

Bottle c1598,p13,471,q1309,2802

Collected By: BOLTON

Comment:

Sample 34350

Sample 34350

ID of Source:

Collection 2/19/2002 AT 10:33:00

Submitted 2/20/2002 AT 11:32:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AE04129

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	4/2/2002	PAD
EPA'79	Potable Metals Digestion	Completed			2/21/2002	PAD
SM18	Tin	< MDL	ug/L	100	2/27/2002	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds

ug/L

4/1/2002

DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds.

The recoveries for compounds in the LCS Standard were high. New recovery limits will be calculated.

Approved Pam Dilsizian

QA Officer

Date Approved: 4/2/2002

Environmental Laboratories

emailed

Original 4/2/2002

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle c1551,p948,955,q2466,2467,263

Collected By: ASPERANZA

Comment:

Sample 40550

Sample 40550

ID of Source:

Collection 2/19/2002 AT 1:02:00

Submitted 2/20/2002 AT 11:32:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AE04130

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
<u>ICPMS Metals for NEWQ</u>						
EPA94	Antimony	< MDL	ug/L	0.1	3/14/2002	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	3/14/2002	PAD
EPA94	Barium	15.3	ug/L	0.1	3/14/2002	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	3/14/2002	PAD
EPA94	Chromium	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Copper	8.1	ug/L	1.0	3/14/2002	PAD
EPA94	Lead	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Manganese	44.5	ug/L	1.0	3/14/2002	PAD
EPA94	Nickel	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Selenium	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Silver	< MDL	ug/L	2.0	3/14/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Zinc	2.4	ug/L	2.0	3/14/2002	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	4/2/2002	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	4/2/2002	PAD
EPA94	Iron	66	ug/L	20.0	4/2/2002	PAD
EPA79	Mercury	< MDL	ug/L	0.40	2/21/2002	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 4/2/2002

Environmental Laboratories

emailed

Original 4/2/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle c1551,p948,955,q2466,2467,263

Collected By: ASPERANZA

Comment:

Sample 40550

Sample 40550

ID of Source:

Collection 2/19/2002 AT 1:02:00

Submitted 2/20/2002 AT 11:32:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AE04130

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	4/2/2002	PAD
EPA'79	Potable Metals Digestion	Completed			2/21/2002	PAD
SM18	Tin	< MDL	ug/L	100	2/27/2002	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds

ug/L

4/1/2002

DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds.

The recoveries for compounds in the LCS Standard were high. New recovery limits will be calculated.

Approved Pam Dilsizian

QA Officer

Date Approved: 4/2/2002

Environmental Laboratories

emailed

Original 4/2/2002

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle c1554,p1054,1865,q1803,1408

Collected By: PATEL

Comment:

Sample 53550

Sample 53550

ID of Source:

Collection 2/19/2002 AT 11:54:00

Submitted 2/20/2002 AT 11:32:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AE04131

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
ICPMS Metals for NEWQ						
EPA94	Antimony	< MDL	ug/L	0.1	3/14/2002	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	3/14/2002	PAD
EPA94	Barium	13.6	ug/L	0.1	3/14/2002	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	3/14/2002	PAD
EPA94	Chromium	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Copper	6.1	ug/L	1.0	3/14/2002	PAD
EPA94	Lead	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Manganese	18.2	ug/L	1.0	3/14/2002	PAD
EPA94	Nickel	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Selenium	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Silver	< MDL	ug/L	2.0	3/14/2002	PAD
EPA94	Thallium	< MDL	ug/L	1.0	3/14/2002	PAD
EPA94	Zinc	3.0	ug/L	2.0	3/14/2002	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	4/2/2002	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	4/2/2002	PAD
EPA94	Iron	43	ug/L	20.0	4/2/2002	PAD
EPA79	Mercury	< MDL	ug/L	0.40	2/21/2002	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 4/2/2002

Environmental Laboratories

emailed

Original 4/2/2002

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received EP

Bottle c1554,p1054,1865,q1803,1408

Collected By: PATEL

Comment:

Sample 53550

Sample 53550

ID of Source:

Collection 2/19/2002 AT 11:54:00

Submitted 2/20/2002 AT 11:32:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample chilled on arrival? : yes

Report

Sample No. AE04131

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	4/2/2002	PAD
EPA'79	Potable Metals Digestion	Completed			2/21/2002	PAD
SM18	Tin	< MDL	ug/L	100	2/27/2002	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds

ug/L

4/1/2002

DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds.

The recoveries for compounds in the LCS Standard were high. New recovery limits will be calculated.

Approved Pam Dilsizian

QA Officer

Date Approved: 4/2/2002

Environmental Laboratories

emailed

Original 4/2/2002

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
Received EP
Bottle p1691,q1385
Collected By: LAB
Comment:

Sample FBD

Sample FBD
ID of Source:
Collection 2/20/2002 AT 12:00:00
Submitted 2/20/2002 AT 11:41:00
Sample Type: DI WATER
PWS No.: **Source Code:**
Type Descriptor:
Sample chilled on arrival? : yes

Report

Sample No. AE04132

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Organics						
<u>GCMS Scan For Unknowns</u>						
	Other unknown compounds	.	ug/L		4/1/2002	DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds.
 The recoveries for compounds in the LCS Standard were high. New recovery limits will be calculated.

Approved Pam Dilsizian QA Officer

Environmental Laboratories

NYS ELAP # 10108

Date Approved: 4/1/2002

emailed Original 4/2/2002