

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received PH

Bottle C0055,P3286,3273,Q3107,3109

Collected By: YACOUN

Comment:

Sample 3

Sample 3

ID of Source:

Collection 03/09/02 AT 9:04:00

Submitted 03/09/02 AT 10:06:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Report

Sample No. AE05908

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
ICPMS Metals for NEWQ						
EPA94	Antimony	< MDL	ug/L	0.1	04/05/02	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	04/05/02	PAD
EPA94	Barium	13.5	ug/L	0.1	04/05/02	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	04/05/02	PAD
EPA94	Chromium	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Copper	1.7	ug/L	1.0	04/05/02	PAD
EPA94	Lead	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Manganese	18.7	ug/L	1.0	04/05/02	PAD
EPA94	Nickel	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Selenium	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Silver	< MDL	ug/L	2.0	04/05/02	PAD
EPA94	Thallium	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Zinc	< MDL	ug/L	2.0	04/05/02	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	04/12/02	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	04/12/02	PAD
EPA94	Iron	33	ug/L	20.0	04/12/02	PAD
EPA79	Mercury	< MDL	ug/L	0.40	03/13/02	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 04/12/02

Environmental Laboratories

emailed

Copy 04/16/02

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received PH

Bottle C0055,P3286,3273,Q3107,3109

Collected By: YACoub

Comment:

Report

Sample 3

Sample 3

ID of Source:

Collection 03/09/02 AT 9:04:00

Submitted 03/09/02 AT 10:06:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample No. AE05908

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	04/12/02	PAD
EPA'79	Potable Metals Digestion	Completed			03/18/02	PAD
SM18	Tin	< MDL	ug/L	100	03/18/02	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds

ug/L

04/12/02

DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds. The recoveries for 8 of the 43 compounds in the LCS Standard were high.

Approved Pam Dilsizian

QA Officer

Date Approved: 04/12/02

Environmental Laboratories

emailed

Copy 04/16/02

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Sample 58

Distribution Laboratory
96-05 Horace Harding Ex.
Corona N.Y. 11368
Dr. Ian Hurley

Received PH

Bottle C9692,P2836,1527,Q19,3146

Collected By: RIBANDO

Comment:

Sample 58

ID of Source:

Collection 03/09/02 AT 8:30:00

Submitted 03/09/02 AT 10:06:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Report

Sample No. AE05909

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
ICPMS Metals for NEWQ						
EPA94	Antimony	< MDL	ug/L	0.1	04/05/02	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	04/05/02	PAD
EPA94	Barium	13.6	ug/L	0.1	04/05/02	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	04/05/02	PAD
EPA94	Chromium	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Copper	1.0	ug/L	1.0	04/05/02	PAD
EPA94	Lead	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Manganese	16.2	ug/L	1.0	04/05/02	PAD
EPA94	Nickel	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Selenium	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Silver	< MDL	ug/L	2.0	04/05/02	PAD
EPA94	Thallium	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Zinc	< MDL	ug/L	2.0	04/05/02	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	04/12/02	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	04/12/02	PAD
EPA94	Iron	29	ug/L	20.0	04/12/02	PAD
EPA79	Mercury	< MDL	ug/L	0.40	03/13/02	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 04/12/02

Environmental Laboratories

emailed

Copy 04/16/02

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
Received PH
Bottle C9692,P2836,1527,Q19,3146
Collected By: RIBANDO
Comment:

Sample 58

Sample 58
ID of Source:
Collection 03/09/02 AT 8:30:00
Submitted 03/09/02 AT 10:06:00
Sample Type: POT_DW
PWS No.:
Source Code:
Type Descriptor:

Report

Sample No. AE05909

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	04/12/02	PAD
EPA'79	Potable Metals Digestion	Completed			03/18/02	PAD
SM18	Tin	< MDL	ug/L	100	03/18/02	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds	ug/L	03/21/02	DLV
-------------------------	------	----------	-----

The GCMS scan, from 35 to 550 amu, reveals the presence of a poorly identified hydrocarbon at 30.61 minutes, with an estimated concentration of 2.7 ug/L. Recoveries for 27 of 43 compounds in the LCS Standard were above their acceptance ranges, while one was below. Recalculation of recovery values will be performed.

Approved Pam Dilsizian

QA Officer

Date Approved: 04/12/02

Environmental Laboratories

emailed

Copy 04/16/02

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
Received PH
Bottle C0091,P420,2876,Q3080,3081
Collected By: BOLTON
Comment:

Sample 07493

Sample 17050
ID of Source:
Collection 03/08/02 AT 12:00:00
Submitted 03/09/02 AT 10:06:00
Sample Type: POT_DW
PWS No.:
Source Code:
Type Descriptor:

Report

Sample No. AE05910

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
ICPMS Metals for NEWQ						
EPA94	Antimony	< MDL	ug/L	0.1	04/05/02	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	04/05/02	PAD
EPA94	Barium	13.5	ug/L	0.1	04/05/02	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	04/05/02	PAD
EPA94	Chromium	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Copper	7.1	ug/L	1.0	04/05/02	PAD
EPA94	Lead	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Manganese	16.4	ug/L	1.0	04/05/02	PAD
EPA94	Nickel	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Selenium	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Silver	< MDL	ug/L	2.0	04/05/02	PAD
EPA94	Thallium	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Zinc	< MDL	ug/L	2.0	04/05/02	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	04/12/02	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	04/12/02	PAD
EPA94	Iron	57	ug/L	20.0	04/12/02	PAD
EPA79	Mercury	< MDL	ug/L	0.40	03/13/02	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 04/12/02

Environmental Laboratories

emailed

Copy 04/16/02

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received PH

Bottle C0091,P420,2876,Q3080,3081

Collected By: BOLTON

Comment:

Report

Sample 07493

Sample 17050

ID of Source:

Collection 03/08/02 AT 12:00:00

Submitted 03/09/02 AT 10:06:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Sample No. AE05910

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	04/12/02	PAD
EPA'79	Potable Metals Digestion	Completed			03/18/02	PAD
SM18	Tin	< MDL	ug/L	100	03/18/02	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds

ug/L

03/21/02

DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds. Recoveries for 27 of 43 compounds in the LCS Standard were above their acceptance ranges, while one was below. Recalculation of recovery values will be performed.

Approved Pam Dilsizian

QA Officer

Date Approved: 04/12/02

Environmental Laboratories

emailed

Copy 04/16/02

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received PH

Bottle C0985,P1181,340,Q3076,3078

Collected By: NAKHLA

Comment:

Sample 07495

Sample 20650

ID of Source:

Collection 03/08/02 AT 11:19:00

Submitted 03/09/02 AT 10:06:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Report

Sample No. AE05911

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
ICPMS Metals for NEWQ						
EPA94	Antimony	< MDL	ug/L	0.1	04/05/02	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	04/05/02	PAD
EPA94	Barium	16.9	ug/L	0.1	04/05/02	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	04/05/02	PAD
EPA94	Chromium	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Copper	13.1	ug/L	1.0	04/05/02	PAD
EPA94	Lead	1.1	ug/L	1.0	04/05/02	PAD
EPA94	Manganese	36.6	ug/L	1.0	04/05/02	PAD
EPA94	Nickel	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Selenium	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Silver	< MDL	ug/L	2.0	04/05/02	PAD
EPA94	Thallium	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Zinc	< MDL	ug/L	2.0	04/05/02	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	04/12/02	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	04/12/02	PAD
EPA94	Iron	91	ug/L	20.0	04/12/02	PAD
EPA79	Mercury	< MDL	ug/L	0.40	03/13/02	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 04/12/02

Environmental Laboratories

emailed

Copy 04/16/02

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Sample 07495

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received PH

Bottle C0985,P1181,340,Q3076,3078

Collected By: NAKHLA

Comment:

Sample 20650

ID of Source:

Collection 03/08/02 AT 11:19:00

Submitted 03/09/02 AT 10:06:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Report

Sample No. AE05911

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	04/12/02	PAD
EPA'79	Potable Metals Digestion	Completed			03/18/02	PAD
SM18	Tin	< MDL	ug/L	100	03/18/02	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds

ug/L

03/21/02

DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds. Recoveries for 27 of 43 compounds in the LCS Standard were above their acceptance ranges, while one was below. Recalculation of recovery values will be performed.

Approved Pam Disizian

QA Officer

Date Approved: 04/12/02

Environmental Laboratories

emailed

Copy 04/16/02

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
Received PH
Bottle C9835,P416,Q3069,3066
Collected By: BROWN
Comment: P3253 BOTTLE BROKEN

Sample 07506

Sample 35150

ID of Source:
Collection 03/08/02 AT 9:44:00
Submitted 03/09/02 AT 10:06:00
Sample Type: POT_DW
PWS No.:
Source Code:
Type Descriptor:

Report

Sample No. AE05912

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
ICPMS Metals for NEWQ						
EPA94	Antimony	< MDL	ug/L	0.1	04/08/02	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	04/08/02	PAD
EPA94	Barium	14.1	ug/L	0.1	04/08/02	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	04/08/02	PAD
EPA94	Chromium	< MDL	ug/L	1.0	04/08/02	PAD
EPA94	Copper	5.0	ug/L	1.0	04/08/02	PAD
EPA94	Lead	< MDL	ug/L	1.0	04/08/02	PAD
EPA94	Manganese	18.8	ug/L	1.0	04/08/02	PAD
EPA94	Nickel	< MDL	ug/L	1.0	04/08/02	PAD
EPA94	Selenium	< MDL	ug/L	1.0	04/08/02	PAD
EPA94	Silver	< MDL	ug/L	2.0	04/08/02	PAD
EPA94	Thallium	< MDL	ug/L	1.0	04/08/02	PAD
EPA94	Zinc	< MDL	ug/L	2.0	04/08/02	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	04/12/02	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	04/12/02	PAD
EPA94	Iron	36	ug/L	20.0	04/12/02	PAD
EPA79	Mercury	< MDL	ug/L	0.40	03/13/02	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 04/12/02

Environmental Laboratories

emailed

Copy 04/16/02

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received PH

Bottle C9835,P416,Q3069,3066

Collected By: BROWN

Comment: P3253 BOTTLE BROKEN

Sample 07506

Sample 35150

ID of Source:

Collection 03/08/02 AT 9:44:00

Submitted 03/09/02 AT 10:06:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Report

Sample No. AE05912

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	04/12/02	PAD
EPA79	Potable Metals Digestion	Completed			03/18/02	PAD
SM18	Tin	< MDL	ug/L	100	03/18/02	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds

ug/L

03/21/02

DLV

The GCMS scan, from 35 to 550 amu, reveals the presence of a poorly identified hydrocarbon at 30.61 minutes, with an estimated concentration of 0.77 ug/L. Recoveries for 27 of 43 compounds in the LCS Standard were above their acceptance ranges, while one was below. Recalculation of recovery values will be performed.

Approved Pam Dilsizian

QA Officer

Date Approved: 04/12/02

Environmental Laboratories

emailed

Copy 04/16/02

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received PH

Bottle C9985,P3265,3266,Q3070,3071

Collected By: AHMAD

Comment:

Sample 07519

Sample 45050

ID of Source:

Collection 03/08/02 AT 10:55:00

Submitted 03/09/02 AT 10:06:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Report

Sample No. AE05913

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
ICPMS Metals for NEWQ						
EPA94	Antimony	< MDL	ug/L	0.1	04/05/02	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	04/05/02	PAD
EPA94	Barium	13.6	ug/L	0.1	04/05/02	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	04/05/02	PAD
EPA94	Chromium	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Copper	11.5	ug/L	1.0	04/05/02	PAD
EPA94	Lead	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Manganese	19.7	ug/L	1.0	04/05/02	PAD
EPA94	Nickel	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Selenium	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Silver	< MDL	ug/L	2.0	04/05/02	PAD
EPA94	Thallium	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Zinc	3.1	ug/L	2.0	04/05/02	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	04/12/02	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	04/12/02	PAD
EPA94	Iron	36	ug/L	20.0	04/12/02	PAD
EPA79	Mercury	< MDL	ug/L	0.40	03/13/02	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 04/12/02

Environmental Laboratories

emailed

Copy 04/16/02

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
Received PH
Bottle C9985,P3265,3266,Q3070,3071
Collected By: AHMAD
Comment:

Sample 07519

Sample 45050
ID of Source:
Collection 03/08/02 AT 10:55:00
Submitted 03/09/02 AT 10:06:00
Sample Type: POT_DW
PWS No.:
Source Code:
Type Descriptor:

Report

Sample No. AE05913

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	04/12/02	PAD
EPA'79	Potable Metals Digestion	Completed			03/18/02	PAD
SM18	Tin	< MDL	ug/L	100	03/18/02	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds	.	ug/L	03/21/02	DLV
-------------------------	---	------	----------	-----

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds. Recoveries for 27 of 43 compounds in the LCS Standard were above their acceptance ranges, while one was below. Recalculation of recovery values will be performed.

Approved Pam Dilisizian

QA Officer

Date Approved: 04/12/02

Environmental Laboratories

emailed

Copy 04/16/02

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC

Sample 07522

Distribution Laboratory

96-05 Horace Harding Ex.

Corona N.Y. 11368

Dr. Ian Hurley

Received PH

Bottle C9992,P3263,3289,Q3075,3084

Collected By: PATEL

Comment:

Sample 52050

ID of Source:

Collection 03/08/02 AT 9:38:00

Submitted 03/09/02 AT 10:06:00

Sample Type: POT_DW

PWS No.:

Source Code:

Type Descriptor:

Report

Sample No. AE05914

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Inorganics						
ICPMS Metals for NEWQ						
EPA94	Antimony	< MDL	ug/L	0.1	04/05/02	PAD
EPA94	Arsenic	< MDL	ug/L	2.0	04/05/02	PAD
EPA94	Barium	13.1	ug/L	0.1	04/05/02	PAD
EPA94	Beryllium	< MDL	ug/L	2.0	04/05/02	PAD
EPA94	Chromium	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Copper	8.8	ug/L	1.0	04/05/02	PAD
EPA94	Lead	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Manganese	18.5	ug/L	1.0	04/05/02	PAD
EPA94	Nickel	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Selenium	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Silver	< MDL	ug/L	2.0	04/05/02	PAD
EPA94	Thallium	< MDL	ug/L	1.0	04/05/02	PAD
EPA94	Zinc	5.4	ug/L	2.0	04/05/02	PAD
EPA94200.	Aluminum	< MDL	ug/L	50.0	04/12/02	PAD
EPA94	Cobalt	< MDL	ug/L	20.0	04/12/02	PAD
EPA94	Iron	50	ug/L	20.0	04/12/02	PAD
EPA79	Mercury	< MDL	ug/L	0.40	03/13/02	PAD

Approved Pam Dilsizian

QA Officer

Date Approved: 04/12/02

Environmental Laboratories

emailed

Copy 04/16/02

NYS ELAP # 10108

Page 1 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
Received PH
Bottle C9992,P3263,3289,Q3075,3084
Collected By: PATEL
Comment:

Sample 07522

Sample 52050

ID of Source:
Collection 03/08/02 AT 9:38:00
Submitted 03/09/02 AT 10:06:00
Sample Type: POT_DW
PWS No.:
Source Code:
Type Descriptor:

Report

Sample No. AE05914

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
EPA94	Molybdenum	< MDL	ug/L	50.0	04/12/02	PAD
EPA'79	Potable Metals Digestion	Completed			03/18/02	PAD
SM18	Tin	< MDL	ug/L	100	03/18/02	PAD

Organics

GCMS Scan For Unknowns

Other unknown compounds	ug/L	04/05/02	DLV
-------------------------	------	----------	-----

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds.

Approved Pam Dilsizian

QA Officer

Date Approved: 04/12/02

Environmental Laboratories

emailed

Copy 04/16/02

NYS ELAP # 10108

Page 2 of 2

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
Received PH
Bottle P3303,Q3141
Collected By: LAB
Comment:

Sample FBD

Sample FBD
ID of Source:
Collection 3/9/2002 AT 12:00:00
Submitted 3/9/2002 AT 10:06:00
Sample Type: POT_DW
PWS No.: **Source Code:**

Type Descriptor:
Sample chilled on arrival? : YES

Report

Sample No. AE05915

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Organics						
GCMS Scan For Unknowns						
	Other unknown compounds	.	ug/L		4/5/2002	DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds.

Approved Pam Dilsizian **QA Officer**

Environmental Laboratories

NYS ELAP # 10108

Date Approved: 4/8/2002

emailed Original 4/8/2002

REPORT OF ANALYSIS

Westchester County Department of Labs and Research 2 Dana Road Valhalla, New York

Agency: New York City DEP DWQC
 Distribution Laboratory
 96-05 Horace Harding Ex.
 Corona N.Y. 11368
 Dr. Ian Hurley
Received PH
Bottle P3303,Q3141
Collected By: LAB
Comment:

Sample FBD

Sample FBD
ID of Source:
Collection 3/9/2002 AT 12:00:00
Submitted 3/9/2002 AT 10:06:00
Sample Type: POT_DW
PWS No.: **Source Code:**

Type Descriptor:
Sample chilled on arrival? : YES

Report

Sample No. AE05915

Method	Test Description	Results	Units	MDL	Analyzed on	Validator
Organics						
<u>GCMS Scan For Unknowns</u>						
	Other unknown compounds	.	ug/L		4/5/2002	DLV

The GCMS scan, from 35 to 550 amu, does not reveal the presence of unusual compounds.

Approved Pam Dilsizian **QA Officer**

Environmental Laboratories

NYS ELAP # 10108

Date Approved: 4/8/2002

emailed Original 4/8/2002