

Page 83

Lab #	Description	Sample vol/wt	Drip rate ml/min	Surr add	Spk add	IS add	Final vol.	T.V. pos.	pH
L6	L6 rat Pest MDL Pest PRL 1242 22542 22543 22544 22545 22546 22547	50W	—	✓	—	2nd	—	—	7
									4 5 6 7 8 9
QC Reviewed by: <i>[Signature]</i> 11/19/01									
Remarks									
22545 contaminated by pipet bulb?									
✓ 20E68									
drip rate must be between 5-15 ml/min for liq-liq extractors drip rate -									
Extraction type - Sup		Solv. lot# V16469		NaCl lot# T12417		Na2SO4 lot#			
STANDARD	INT STD LOG#								
CALIBRATION									
REF/SPIKE	2062w pest								
MDL/MDF	1870c pest								
Q. CHECK	—								
INT. STD.	—								
SURROGATE	1834w								
LPC	—								
ENDRIN/DDT									

Analyst hmy/SD/M.C
revision 10/15/99