

Method 508

Date Extracted 9/25/01

Page 1

Lab #	Description	Sample vol/wt	Drip rate ml/min	Surr add	Spk add	IS add	Final vol.	T.V. pos.	pH =7	Count
	UP	50ml	—	✓		—	2ml		✓	
	Net toluene									
	MDL TOL									
	PRC 1232									
23318	spike pest				✓					
23318	spike dup pest 4th				✓					
23324	spike TOL 500				✓					
23324	spike dup TOL				✓					
	23270									4
	23271									5
	23272									6
	23318									7
	23319									8
	23320									9
	23321									10
	23322									11
	23323									12
	23324									13
Remarks										
drip rate must be between 5-15 ml/min for liq-liq extractors drip rate - Extraction type <u>Sep Tex</u> Solv. lot# <u>T 12617</u> NaCl lot# <u>118617</u> Na2SO4 lot# <u>118617</u> STANDARD INT STD LOG# CALIBRATION REF/SPIKE <u>pot 20644 to 16784</u> MDL/MDRF Q. CHECK INT. STD. SURROGATE <u>20644</u> LPC ENDRIN/DDT										

QC Reviewed by
4/12/13/01

Analyst my/AD/CM
revision 4/12/01

Verified by _____

Reviewed by DN
Date 10/17/11