




12/5/01

westchester county labs

Folder: RUNS1
Protocol: hg-cv
POST-RUN REPORT

Page 160

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Standard: 1 Rep: 1								
				Seq: 1081	14:56:13	05	Dec 01	HG
Hg	.000	ppb	-955					=
*** Standard: 2 Rep: 1								
				Seq: 1082	14:58:02	05	Dec 01	HG
Hg	.400	ppb	29750					=
*** Standard: 3 Rep: 1								
				Seq: 1083	14:59:54	05	Dec 01	HG
Hg	1.00	ppb	77524					=
*** Standard: 4 Rep: 1								
				Seq: 1084	15:01:31	05	Dec 01	HG
Hg	3.00	ppb	237665					=
*** Standard: 5 Rep: 1								
				Seq: 1085	15:03:08	05	Dec 01	HG
Hg	5.00	ppb	391777					=
*** Standard: 6 Rep: 1								
				Seq: 1086	15:04:46	05	Dec 01	HG
Hg	10.0	ppb	771705					=
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		96.9	.969	1.00	ppb	.000		=
*** Sample ID: 28751 pot								
				Seq: 1091	15:14:19	05	Dec 01	HG
Hg	-.034	ppb	.000	-.034				=
*** Sample ID: 28751d								
				Seq: 1092	15:16:28	05	Dec 01	HG
Hg	-.021	ppb	.000	-.021				=
*** Sample ID: 28751 spk								
				Seq: 1093	15:18:15	05	Dec 01	HG
Hg	5.04	ppb	.000	5.04				=
*** Sample ID: 28752								
				Seq: 1094	15:19:51	05	Dec 01	HG
Hg	-.048	ppb	.000	-.048				=
*** Sample ID: 28753								
				Seq: 1095	15:23:24	05	Dec 01	HG
Hg	-.029	ppb	.000	-.029				=
*** Sample ID: 28754								
				Seq: 1096	15:25:27	05	Dec 01	HG
Hg	-.026	ppb	.000	-.026				=
*** Sample ID: 28755								
				Seq: 1097	15:27:07	05	Dec 01	HG
Hg	-.023	ppb	.000	-.023				=

westchester county labs

Folder: RUNS1

Page 161

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
=====								
*** Check Standard: 2 Ck21.0 qc Seq: 1098 15:28:48 05 Dec 01 HG								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		97.8 ✓	.978	1.00	ppb	.000		
=====								
*** Sample ID: 28756 Seq: 1099 15:30:27 05 Dec 01 HG								
Hg	-.021	ppb	.000	-.021				
=====								
*** Sample ID: 28757 Seq: 1100 15:32:03 05 Dec 01 HG								
Hg	.022	ppb	.000	.022				
=====								
*** Sample ID: 28759 Seq: 1101 15:33:59 05 Dec 01 HG								
Hg	-.023	ppb	.000	-.023				
=====								
*** Sample ID: 28760 Seq: 1102 15:35:46 05 Dec 01 HG								
Hg	-.020	ppb	.000	-.020				
=====								
*** Sample ID: 28801 Seq: 1103 15:37:43 05 Dec 01 HG								
Hg	-.023	ppb	.000	-.023				
=====								
*** Sample ID: 28802 Seq: 1104 15:39:21 05 Dec 01 HG								
Hg	-.020	ppb	.000	-.020				
=====								
*** Sample ID: 28897 Seq: 1108 15:47:25 05 Dec 01 HG								
Hg	-.027	ppb	.000	-.027				
=====								
*** Check Standard: 2 Ck21.0 qc Seq: 1109 15:49:33 05 Dec 01 HG								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		97.7 ✓	.977	1.00	ppb	.000		
=====								
*** Sample ID: 28898 Seq: 1110 15:51:31 05 Dec 01 HG								
Hg	-.020	ppb	.000	-.020				
=====								
*** Sample ID: 28899 Seq: 1111 15:53:41 05 Dec 01 HG								
Hg	-.024	ppb	.000	-.024				
=====								
*** Sample ID: 28900 Seq: 1112 15:55:19 05 Dec 01 HG								
Hg	-.021	ppb	.000	-.021				
=====								
*** Sample ID: 28901 Seq: 1113 15:56:56 05 Dec 01 HG								
Hg	-.024	ppb	.000	-.024				
=====								

westchester county labs

Folder: RUNS1

Page 162

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 28902 Seq: 1114 15:58:33 05 Dec 01 HG								
Hg	-.018	ppb	.000	-.018				
=====								
*** Sample ID: 28903 Seq: 1115 16:00:10 05 Dec 01 HG								
Hg	-.017	ppb	.000	-.017				
=====								
*** Sample ID: 28903d Seq: 1116 16:01:58 05 Dec 01 HG								
Hg	-.039	ppb	.000	-.039				
=====								
*** Sample ID: 28903 spk Seq: 1117 16:03:36 05 Dec 01 HG								
Hg	4.67	ppb	.000	4.67				
=====								
*** Check Standard: 2 Ck21.0 qc Seq: 1120 16:08:42 05 Dec 01 HG								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		99.7	.997	1.00	ppb	.000		
=====								
*** Sample ID: 29022 pot Seq: 1125 16:19:17 05 Dec 01 HG								
Hg	-.009	ppb	.000	-.009				
=====								
*** Sample ID: 29023 Seq: 1126 16:21:14 05 Dec 01 HG								
Hg	-.016	ppb	.000	-.016				
=====								
*** Sample ID: 29024 Seq: 1127 16:22:53 05 Dec 01 HG								
Hg	-.025	ppb	.000	-.025				
=====								
*** Sample ID: 29025 Seq: 1128 16:24:50 05 Dec 01 HG								
Hg	-.009	ppb	.000	-.009				
=====								
*** Sample ID: 29026 Seq: 1129 16:26:27 05 Dec 01 HG								
Hg	-.027	ppb	.000	-.027				
=====								
*** Sample ID: 29027 Seq: 1130 16:28:08 05 Dec 01 HG								
Hg	-.028	ppb	.000	-.028				
=====								
*** Check Standard: 2 Ck21.0 qc Seq: 1131 16:30:47 05 Dec 01 HG								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		98.6	.986	1.00	ppb	.000		
=====								
*** Sample ID: 29028 Seq: 1132 16:32:55 05 Dec 01 HG								
Hg	-.021	ppb	.000	-.021				
=====								
*** Check Standard: 2 Ck21.0 qc Seq: 1142 16:51:04 05 Dec 01 HG								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		