




westchester county labs

Folder: RUNS1

Page 176

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Standard: 1 Rep: 1								
				Seq: 1217	15:48:23	11 Dec 01	HG	
Hg	.000	ppb	411					=
*** Standard: 2 Rep: 1								
				Seq: 1218	15:50:00	11 Dec 01	HG	
Hg	.400	ppb	28390					=
*** Standard: 3 Rep: 1								
				Seq: 1219	15:51:40	11 Dec 01	HG	
Hg	1.00	ppb	72560					=
*** Standard: 4 Rep: 1								
				Seq: 1220	15:53:32	11 Dec 01	HG	
Hg	3.00	ppb	219849					=
*** Standard: 5 Rep: 1								
				Seq: 1221	15:55:23	11 Dec 01	HG	
Hg	5.00	ppb	363098					=
*** Standard: 6 Rep: 1								
				Seq: 1222	15:57:01	11 Dec 01	HG	
Hg	10.0	ppb	710510					=
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		99.3	.993	1.00	ppb	.000		=
*** Sample ID: 29454 pot								
				Seq: 1224	16:01:56	11 Dec 01	HG	
Hg	.006	ppb	.000	.006				=
*** Sample ID: 29517 pot								
				Seq: 1225	16:04:07	11 Dec 01	HG	
Hg	-.027	ppb	.000	-.027				=
*** Sample ID: 29517d								
				Seq: 1226	16:06:02	11 Dec 01	HG	
Hg	-.042	ppb	.000	-.042				=
*** Sample ID: 29517 spk								
				Seq: 1227	16:08:28	11 Dec 01	HG	
Hg	5.28	ppb	.000	5.28				=
*** Sample ID: 29518								
				Seq: 1228	16:10:06	11 Dec 01	HG	
Hg	-.049	ppb	.000	-.049				=
*** Sample ID: 29519								
				Seq: 1229	16:11:42	11 Dec 01	HG	
Hg	-.036	ppb	.000	-.036				=
*** Sample ID: 29520								
				Seq: 1230	16:13:29	11 Dec 01	HG	
Hg	-.020	ppb	.000	-.020				=

westchester county labs

Folder: RUNS1

Page 177

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 29521								
				Seq: 1231	16:15:07	11 Dec 01	HG	
Hg	-.017	ppb	.000	-.017				
=====								
*** Sample ID: 29522								
				Seq: 1232	16:17:45	11 Dec 01	HG	
Hg	-.057	ppb	.000	-.057				
=====								
*** Sample ID: 29523								
				Seq: 1233	16:19:21	11 Dec 01	HG	
Hg	-.065	ppb	.000	-.065				
=====								
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		96.1	.961	1.00	ppb	.000		
=====								
*** Sample ID: 29584								
				Seq: 1235	16:22:56	11 Dec 01	HG	
Hg	-.024	ppb	.000	-.024				
=====								
*** Sample ID: 29664								
				Seq: 1236	16:24:33	11 Dec 01	HG	
Hg	-.029	ppb	.000	-.029				
=====								
*** Sample ID: 29665								
				Seq: 1237	16:26:09	11 Dec 01	HG	
Hg	-.023	ppb	.000	-.023				
=====								
*** Sample ID: 29666								
				Seq: 1238	16:28:45	11 Dec 01	HG	
Hg	-.042	ppb	.000	-.042				
=====								
*** Sample ID: 29667								
				Seq: 1239	16:30:31	11 Dec 01	HG	
Hg	-.065	ppb	.000	-.065				
=====								
*** Sample ID: 29668								
				Seq: 1240	16:32:20	11 Dec 01	HG	
Hg	-.067	ppb	.000	-.067				
=====								
*** Sample ID: 29669								
				Seq: 1241	16:34:09	11 Dec 01	HG	
Hg	-.036	ppb	.000	-.036				
=====								
*** Sample ID: 29670								
				Seq: 1242	16:35:49	11 Dec 01	HG	
Hg	-.027	ppb	.000	-.027				
=====								
*** Sample ID: 29704								
				Seq: 1243	16:37:58	11 Dec 01	HG	
Hg	-.007	ppb	.000	-.007				
=====								

westchester county labs

Folder: RUNS1

Page 178

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 29706								
				Seq: 1244	16:39:41	11 Dec 01	HG	
Hg	-.041	ppb	.000	-.041				
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		96.6	.966	1.00	ppb	.000		
*** Sample ID: 29708								
				Seq: 1246	16:42:58	11 Dec 01	HG	
Hg	-.081	ppb	.000	-.081				
*** Sample ID: 29781								
				Seq: 1247	16:44:56	11 Dec 01	HG	
Hg	-.074	ppb	.000	-.074				
*** Sample ID: 29782								
				Seq: 1248	16:46:44	11 Dec 01	HG	
Hg	-.034	ppb	.000	-.034				
*** Sample ID: 29783								
				Seq: 1249	16:48:32	11 Dec 01	HG	
Hg	-.017	ppb	.000	-.017				
*** Sample ID: 29783d								
				Seq: 1250	16:50:18	11 Dec 01	HG	
Hg	-.022	ppb	.000	-.022				
*** Sample ID: 29783 spk								
				Seq: 1251	16:52:05	11 Dec 01	HG	
Hg	5.42	ppb	.000	5.42				
*** Sample ID: 29784								
				Seq: 1252	16:53:45	11 Dec 01	HG	
Hg	-.068	ppb	.000	-.068				
*** Sample ID: 29785								
				Seq: 1253	16:55:38	11 Dec 01	HG	
Hg	-.064	ppb	.000	-.064				
*** Sample ID: 29786								
				Seq: 1254	16:57:37	11 Dec 01	HG	
Hg	-.033	ppb	.000	-.033				
*** Sample ID: 29787								
				Seq: 1255	16:59:17	11 Dec 01	HG	
Hg	-.025	ppb	.000	-.025				
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		103.	1.03	1.00	ppb	.000		