




12/14/01

westchester county labs

Folder: RUNS1
Protocol: hg-cv
POST-RUN REPORT

Page 186

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Standard: 1 Rep: 1								
				Seq:	1273	14:51:59	14 Dec 01	HG
Hg	.000	ppb	-1281					=
*** Standard: 2 Rep: 1								
				Seq:	1274	14:53:37	14 Dec 01	HG
Hg	.400	ppb	25114					=
*** Standard: 3 Rep: 1								
				Seq:	1275	14:55:15	14 Dec 01	HG
Hg	1.00	ppb	70193					=
*** Standard: 4 Rep: 1								
				Seq:	1276	14:56:53	14 Dec 01	HG
Hg	3.00	ppb	228056					=
*** Standard: 5 Rep: 1								
				Seq:	1277	14:58:31	14 Dec 01	HG
Hg	5.00	ppb	384771					=
*** Standard: 6 Rep: 1								
				Seq:	1278	15:00:14	14 Dec 01	HG
Hg	10.0	ppb	774976					=
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		103. ✓	1.03	1.00	ppb	.000		=
*** Sample ID: 29925 pot								
				Seq:	1280	15:05:27	14 Dec 01	HG
Hg	.071	ppb	.000	.071				=
*** Sample ID: 29925d								
				Seq:	1281	15:07:22	14 Dec 01	HG
Hg	.020	ppb	.000	.020				=
*** Sample ID: 29925 spk								
				Seq:	1282	15:09:21	14 Dec 01	HG
Hg	5.44	ppb	.000	5.44				=
*** Sample ID: 29926								
				Seq:	1283	15:11:18	14 Dec 01	HG
Hg	.144	ppb	.000	.144				=
*** Sample ID: 29927								
				Seq:	1284	15:13:28	14 Dec 01	HG
Hg	.108	ppb	.000	.108				=
*** Sample ID: 29928								
				Seq:	1285	15:15:05	14 Dec 01	HG
Hg	.053	ppb	.000	.053				=
*** Sample ID: 29929								
				Seq:	1286	15:17:01	14 Dec 01	HG
Hg	.027	ppb	.000	.027				=

westchester county labs


Folder: RUNS1

Page 187

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 29930								
				Seq: 1287	15:18:38	14 Dec 01	HG	
Hg	.026	ppb	.000	.026				
=====								
*** Sample ID: 29931								
				Seq: 1288	15:20:16	14 Dec 01	HG	
Hg	.022	ppb	.000	.022				
=====								
*** Sample ID: 30009								
				Seq: 1289	15:22:15	14 Dec 01	HG	
Hg	.040	ppb	.000	.040				
=====								
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		111. ✓	1.11	1.00	ppb	.000		
=====								
*** Sample ID: 30010								
				Seq: 1291	15:26:08	14 Dec 01	HG	
Hg	.090	ppb	.000	.090				
=====								
*** Sample ID: 30011								
				Seq: 1292	15:27:46	14 Dec 01	HG	
Hg	.094	ppb	.000	.094				
=====								
*** Sample ID: 30012								
				Seq: 1293	15:29:22	14 Dec 01	HG	
Hg	.072	ppb	.000	.072				
=====								
*** Sample ID: 30013								
				Seq: 1294	15:30:58	14 Dec 01	HG	
Hg	.028	ppb	.000	.028				
=====								
*** Sample ID: 30014								
				Seq: 1295	15:32:35	14 Dec 01	HG	
Hg	.035	ppb	.000	.035				
=====								
*** Sample ID: 30015								
				Seq: 1296	15:34:53	14 Dec 01	HG	
Hg	.099	ppb	.000	.099				
=====								
*** Sample ID: 30021								
				Seq: 1297	15:36:29	14 Dec 01	HG	
Hg	.039	ppb	.000	.039				
=====								
*** Sample ID: 30022								
				Seq: 1298	15:38:25	14 Dec 01	HG	
Hg	.089	ppb	.000	.089				
=====								
*** Sample ID: 30085								
				Seq: 1299	15:40:54	14 Dec 01	HG	
Hg	.073	ppb	.000	.073				
=====								



westchester county labs

Folder: RUNS1

Page 188

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 30086								
				Seq: 1300	15:42:31	14 Dec 01	HG	
Hg	.040	ppb	.000	.040				=
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		107. ✓	1.07	1.00	ppb	.000		=
*** Standard: 1 Rep: 1								
				Seq: 1313	16:07:34	14 Dec 01	HG	
Hg	.000	ppb	-1235					=
*** Standard: 2 Rep: 1								
				Seq: 1314	16:09:13	14 Dec 01	HG	
Hg	.400	ppb	30228					=
*** Standard: 3 Rep: 1								
				Seq: 1315	16:11:51	14 Dec 01	HG	
Hg	1.00	ppb	81864					=
*** Standard: 4 Rep: 1								
				Seq: 1316	16:13:29	14 Dec 01	HG	
Hg	3.00	ppb	256084					=
*** Standard: 5 Rep: 1								
				Seq: 1317	16:15:06	14 Dec 01	HG	
Hg	5.00	ppb	426621					=
*** Standard: 6 Rep: 1								
				Seq: 1318	16:17:16	14 Dec 01	HG	
Hg	10.0	ppb	833718					=
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		96.0 ✓	.960	1.00	ppb	.000		=
*** Sample ID: 30251								
				Seq: 1320	16:23:02	14 Dec 01	HG	
Hg	-.037	ppb	.000	-.037				=
*** Sample ID: 30252								
				Seq: 1321	16:24:41	14 Dec 01	HG	
Hg	-.019	ppb	.000	-.019				=
*** Sample ID: 30252d								
				Seq: 1322	16:26:39	14 Dec 01	HG	
Hg	.005	ppb	.000	.005				=
*** Sample ID: 30252 spk								
				Seq: 1323	16:28:35	14 Dec 01	HG	
Hg	4.47	ppb	.000	4.47				=
*** Sample ID: 30253								
				Seq: 1324	16:30:17	14 Dec 01	HG	
Hg	.038	ppb	.000	.038				=

westchester county labs

Folder: RUNS1

Page 189

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 30254								
				Seq: 1325	16:31:57	14 Dec 01	HG	
Hg	-.011	ppb	.000	-.011				
=====								
*** Sample ID: 30255								
				Seq: 1326	16:33:37	14 Dec 01	HG	
Hg	-.043	ppb	.000	-.043				
=====								
*** Sample ID: 30256								
				Seq: 1327	16:35:27	14 Dec 01	HG	
Hg	-.041	ppb	.000	-.041				
=====								
*** Sample ID: 30257								
				Seq: 1328	16:37:04	14 Dec 01	HG	
Hg	-.037	ppb	.000	-.037				
=====								
*** Sample ID: 30350								
				Seq: 1329	16:38:41	14 Dec 01	HG	
Hg	-.010	ppb	.000	-.010				
=====								
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		100. ✓	1.00	1.00	ppb	.000		
=====								
*** Sample ID: 30351								
				Seq: 1331	16:42:11	14 Dec 01	HG	
Hg	.027	ppb	.000	.027				
=====								
*** Sample ID: 30352								
				Seq: 1332	16:44:05	14 Dec 01	HG	
Hg	.031	ppb	.000	.031				
=====								
*** Sample ID: 30353								
				Seq: 1333	16:45:48	14 Dec 01	HG	
Hg	-.010	ppb	.000	-.010				
=====								
*** Sample ID: 30354								
				Seq: 1334	16:47:45	14 Dec 01	HG	
Hg	-.042	ppb	.000	-.042				
=====								
*** Sample ID: 30355								
				Seq: 1335	16:49:26	14 Dec 01	HG	
Hg	-.010	ppb	.000	-.010				
=====								
*** Sample ID: 30356								
				Seq: 1336	16:51:16	14 Dec 01	HG	
Hg	-.008	ppb	.000	-.008				
=====								
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		94.5 ✓	.945	1.00	ppb	.000		