




1/3/02

westchester county labs

Folder: RUNS1

Page 211

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Standard: 1 Rep: 1								
				Seq:	1543	15:57:31	03 Jan 02	HG
Hg	.000	ppb	2695					=
*** Standard: 2 Rep: 1								
				Seq:	1544	15:59:18	03 Jan 02	HG
Hg	.400	ppb	26696					=
*** Standard: 3 Rep: 1								
				Seq:	1545	16:01:27	03 Jan 02	HG
Hg	1.00	ppb	72070					=
*** Standard: 4 Rep: 1								
				Seq:	1546	16:03:16	03 Jan 02	HG
Hg	3.00	ppb	225046					=
*** Standard: 5 Rep: 1								
				Seq:	1547	16:04:55	03 Jan 02	HG
Hg	5.00	ppb	385651					=
*** Standard: 6 Rep: 1								
				Seq:	1548	16:06:45	03 Jan 02	HG
Hg	10.0	ppb	767161					=
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg	H	118.	1.18	1.00	ppb	.000		=
(qc - re-powder, sample tip adjusted)								
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg	✓	103. ✓	1.03	1.00	ppb	.000		=
*** Sample ID: 31339 pot								
				Seq:	1551	16:12:38	03 Jan 02	HG
Hg	.158	ppb	.000	.158				=
*** Sample ID: 31340								
				Seq:	1552	16:14:44	03 Jan 02	HG
Hg	.138	ppb	.000	.138				=
*** Sample ID: 31341								
				Seq:	1553	16:16:34	03 Jan 02	HG
Hg	.117	ppb	.000	.117				=
*** Sample ID: 31342								
				Seq:	1554	16:18:23	03 Jan 02	HG
Hg	.111	ppb	.000	.111				=
*** Sample ID: 31343								
				Seq:	1555	16:20:39	03 Jan 02	HG
Hg	.087	ppb	.000	.087				=
*** Sample ID: 31344								
				Seq:	1556	16:22:47	03 Jan 02	HG
Hg	.082	ppb	.000	.082				=

westchester county labs

Folder: RUNS1

Page 212

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
=====								
*** Sample ID: 31345				Seq: 1557	16:24:47	03 Jan 02	HG	
Hg	.085	ppb	.000	.085				
=====								
*** Sample ID: 31348				Seq: 1558	16:26:27	03 Jan 02	HG	
Hg	✓ .090	ppb	.000	.090				
=====								
*** Sample ID: 31348d				Seq: 1559	16:28:05	03 Jan 02	HG	
Hg	✓ .085	ppb	.000	.085				
=====								
*** Sample ID: 31348 spk				Seq: 1560	16:29:45	03 Jan 02	HG	
Hg	4.96	ppb	.000	4.96				
=====								
*** Check Standard: 2 Ck21.0 qc				Seq: 1561	16:31:43	03 Jan 02	HG	
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		102.✓	1.02	1.00	ppb	.000		
=====								