




1/10/02

westchester county labs

Folder: RUNS1

Page 232

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Standard: 1 Rep: 1								
				Seq: 1730	15:13:20	10 Jan 02	HG	
Hg	.000	ppb	4166					
*** Standard: 2 Rep: 1								
				Seq: 1731	15:16:28	10 Jan 02	HG	
Hg	.400	ppb	36601					
*** Standard: 3 Rep: 1								
				Seq: 1732	15:18:55	10 Jan 02	HG	
Hg	1.00	ppb	76276					
*** Standard: 4 Rep: 1								
				Seq: 1733	15:21:05	10 Jan 02	HG	
Hg	3.00	ppb	222853					
*** Standard: 5 Rep: 1								
				Seq: 1734	15:23:04	10 Jan 02	HG	
Hg	5.00	ppb	359935					
*** Standard: 6 Rep: 1								
				Seq: 1735	15:24:41	10 Jan 02	HG	
Hg	10.0	ppb	728281					
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Seq: 1736	15:27:17	10 Jan 02	HG
Hg		104. ✓	1.04	1.00	Units	SD/RSD		
					ppb	.000		
*** Sample ID: 31637 pot								
				Seq: 1737	15:29:03	10 Jan 02	HG	
Hg	.026	ppb	.000	.026				
*** Sample ID: 31638								
				Seq: 1738	15:30:45	10 Jan 02	HG	
Hg	.012	ppb	.000	.012				
*** Sample ID: 31639								
				Seq: 1739	15:32:22	10 Jan 02	HG	
Hg	.007	ppb	.000	.007				
*** Sample ID: 31640								
				Seq: 1740	15:34:03	10 Jan 02	HG	
Hg	-.062	ppb	.000	-.062				
*** Sample ID: 31641								
				Seq: 1741	15:36:20	10 Jan 02	HG	
Hg	-.080	ppb	.000	-.080				
*** Sample ID: 31642								
				Seq: 1742	15:37:56	10 Jan 02	HG	
Hg	-.033	ppb	.000	-.033				
*** Sample ID: 31643								
				Seq: 1743	15:39:55	10 Jan 02	HG	
Hg	.007	ppb	.000	.007				

westchester county labs

Folder: RUNS1

Page 233

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 31645								
				Seq: 1744	15:41:32	10 Jan 02	HG	
Hg	.036	ppb	.000	.036				
=====								
*** Sample ID: 31646								
				Seq: 1745	15:43:19	10 Jan 02	HG	
Hg	-.012	ppb	.000	-.012				
=====								
*** Sample ID: 31662								
				Seq: 1746	15:45:16	10 Jan 02	HG	
Hg	-.036	ppb	.000	-.036				
=====								
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		94.3	.943	1.00	ppb	.000		
=====								
*** Sample ID: 31663								
				Seq: 1748	15:49:04	10 Jan 02	HG	
Hg	-.043	ppb	.000	-.043				
=====								
*** Sample ID: 31663d								
				Seq: 1749	15:51:00	10 Jan 02	HG	
Hg	.016	ppb	.000	.016				
=====								
*** Sample ID: 31663 spk								
				Seq: 1750	15:52:39	10 Jan 02	HG	
Hg	5.19	ppb	.000	5.19				
=====								
*** Sample ID: 003								
				Seq: 1751	15:55:04	10 Jan 02	HG	
Hg	.020	ppb	.000	.020				
=====								
*** Sample ID: 004								
				Seq: 1752	15:57:10	10 Jan 02	HG	
Hg	-.053	ppb	.000	-.053				
=====								
*** Sample ID: 005								
				Seq: 1753	15:58:50	10 Jan 02	HG	
Hg	-.045	ppb	.000	-.045				
=====								
*** Sample ID: 006								
				Seq: 1754	16:01:32	10 Jan 02	HG	
Hg	-.016	ppb	.000	-.016				
=====								
*** Sample ID: 007								
				Seq: 1755	16:03:22	10 Jan 02	HG	
Hg	.026	ppb	.000	.026				
=====								
*** Sample ID: 008								
				Seq: 1756	16:05:54	10 Jan 02	HG	
Hg	.022	ppb	.000	.022				
=====								

westchester county labs

Folder: RUNS1

Page 234

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 009								
				Seq: 1757	16:08:11	10 Jan 02	HG	
Hg	-.066	ppb	.000	-.066				
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		93.1	.931	1.00	ppb	.000		
*** Sample ID: 027								
				Seq: 1759	16:12:10	10 Jan 02	HG	
Hg	-.008	ppb	.000	-.008				
*** Sample ID: 028								
				Seq: 1760	16:14:22	10 Jan 02	HG	
Hg	.013	ppb	.000	.013				
*** Sample ID: 163								
				Seq: 1763	16:20:10	10 Jan 02	HG	
Hg	-.041	ppb	.000	-.041				
*** Sample ID: 164								
				Seq: 1764	16:22:06	10 Jan 02	HG	
Hg	-.058	ppb	.000	-.058				
*** Sample ID: 165								
				Seq: 1765	16:23:57	10 Jan 02	HG	
Hg	.005	ppb	.000	.005				
*** Sample ID: 166								
				Seq: 1766	16:25:38	10 Jan 02	HG	
Hg	.024	ppb	.000	.024				
*** Sample ID: 167								
				Seq: 1767	16:27:16	10 Jan 02	HG	
Hg	.020	ppb	.000	.020				
*** Sample ID: 168								
				Seq: 1768	16:28:53	10 Jan 02	HG	
Hg	-.041	ppb	.000	-.041				
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		94.0	.940	1.00	ppb	.000		
*** Sample ID: 169								
				Seq: 1770	16:32:43	10 Jan 02	HG	
Hg	-.068	ppb	.000	-.068				
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		91.9	.919	1.00	ppb	.000		