




1/11/02

westchester county labs

Folder: RUNS1

Page 241

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
=====								
*** Standard: 1 Rep: 1				Seq: 1790	14:59:20	11 Jan 02	HG	
Hg	.000	ppb	3810					=
*** Standard: 2 Rep: 1				Seq: 1791	15:01:02	11 Jan 02	HG	
Hg	.400	ppb	26114					=
*** Standard: 3 Rep: 1				Seq: 1792	15:02:41	11 Jan 02	HG	
Hg	1.00	ppb	67186					=
*** Standard: 4 Rep: 1				Seq: 1793	15:04:38	11 Jan 02	HG	
Hg	3.00	ppb	206315					=
*** Standard: 5 Rep: 1				Seq: 1794	15:06:15	11 Jan 02	HG	
Hg	5.00	ppb	337499					=
*** Standard: 6 Rep: 1				Seq: 1795	15:08:04	11 Jan 02	HG	
Hg	10.0	ppb	657467					=
*** Check Standard: 2	Ck21.0 qc			Seq: 1796	15:09:43	11 Jan 02	HG	
Line Flag	%Rcv.	Found	True	Units	SD/RSD			
Hg	87.5	.875	1.00	ppb	.000			=
*** Sample ID: 275 pot				Seq: 1797	15:11:29	11 Jan 02	HG	
Hg	.070	ppb	.000	.070				=
=====								
*** Sample ID: 275d				Seq: 1798	15:13:48	11 Jan 02	HG	
Hg	.101	ppb	.000	.101				=
=====								
*** Sample ID: 275 spk				Seq: 1799	15:15:25	11 Jan 02	HG	
Hg	5.05	ppb	.000	5.05				=
=====								
*** Sample ID: 276				Seq: 1800	15:17:06	11 Jan 02	HG	
Hg	.130	ppb	.000	.130				=
=====								
*** Sample ID: 277				Seq: 1801	15:18:44	11 Jan 02	HG	
Hg	.126	ppb	.000	.126				=
=====								
*** Sample ID: 278				Seq: 1802	15:20:26	11 Jan 02	HG	
Hg	.028	ppb	.000	.028				=
=====								
*** Sample ID: 279				Seq: 1803	15:23:16	11 Jan 02	HG	
Hg	.080	ppb	.000	.080				=

westchester county labs

Folder: RUNS1

Page 242

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 280								
				Seq: 1804	15:25:23	11 Jan 02	HG	
Hg	.121	ppb	.000	.121				=
*** Sample ID: 281								
				Seq: 1805	15:26:59	11 Jan 02	HG	
Hg	.160	ppb	.000	.160				=
*** Sample ID: 314								
				Seq: 1806	15:28:40	11 Jan 02	HG	
Hg	.180	ppb	.000	.180				=
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		98.1	.981	1.00	ppb	.000		=
*** Sample ID: 315								
				Seq: 1808	15:32:00	11 Jan 02	HG	
Hg	.109	ppb	.000	.109				=
*** Sample ID: 316								
				Seq: 1809	15:33:39	11 Jan 02	HG	
Hg	.093	ppb	.000	.093				=
*** Sample ID: 317								
				Seq: 1810	15:35:36	11 Jan 02	HG	
Hg	.089	ppb	.000	.089				=
*** Sample ID: 318								
				Seq: 1811	15:37:15	11 Jan 02	HG	
Hg	.081	ppb	.000	.081				=
*** Sample ID: 319								
				Seq: 1812	15:38:51	11 Jan 02	HG	
Hg	.081	ppb	.000	.081				=
*** Sample ID: 320								
				Seq: 1813	15:40:29	11 Jan 02	HG	
Hg	.153	ppb	.000	.153				=
*** Sample ID: 327								
				Seq: 1814	15:42:48	11 Jan 02	HG	
Hg	.177	ppb	.000	.177				=
*** Sample ID: 328								
				Seq: 1815	15:46:08	11 Jan 02	HG	
Hg	.071	ppb	.000	.071				=
*** Sample ID: 352								
				Seq: 1816	15:47:56	11 Jan 02	HG	
Hg	.066	ppb	.000	.066				=

westchester county labs

Folder: RUNS1
Protocol: hg-cv
POST-RUN REPORT

Page 243

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 364								
				Seq: 1817	15:49:45	11 Jan 02	HG	
Hg	.068	ppb	.000	.068				=
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		95.3	.953	1.00	ppb	.000		=
*** Sample ID: 365								
				Seq: 1819	15:53:34	11 Jan 02	HG	
Hg	.071	ppb	.000	.071				=
*** Sample ID: 530								
				Seq: 1820	15:55:13	11 Jan 02	HG	
Hg	.042	ppb	.000	.042				=
*** Sample ID: 531								
				Seq: 1821	15:57:01	11 Jan 02	HG	
Hg	.028	ppb	.000	.028				=
*** Sample ID: 555								
				Seq: 1822	15:59:19	11 Jan 02	HG	
Hg	-.044	ppb	.000	-.044				=
*** Sample ID: 555d								
				Seq: 1823	16:00:55	11 Jan 02	HG	
Hg	-.061	ppb	.000	-.061				=
*** Sample ID: 555 spk								
				Seq: 1824	16:02:34	11 Jan 02	HG	
Hg	4.33	ppb	.000	4.33				=
*** Sample ID: 556								
				Seq: 1825	16:04:11	11 Jan 02	HG	
Hg	-.052	ppb	.000	-.052				=
*** Sample ID: 557								
				Seq: 1826	16:06:40	11 Jan 02	HG	
Hg	.046	ppb	.000	.046				=
*** Sample ID: 558								
				Seq: 1827	16:08:22	11 Jan 02	HG	
Hg	.051	ppb	.000	.051				=
*** Sample ID: 559								
				Seq: 1828	16:10:02	11 Jan 02	HG	
Hg	.013	ppb	.000	.013				=
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		93.4	.934	1.00	ppb	.000		=

westchester county labs

Folder: RUNS1

Page 244

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 560								
				Seq: 1830	16:13:51	11 Jan 02	HG	
Hg	-.076	ppb	.000	-.076				
*** Sample ID: 561								
				Seq: 1831	16:15:29	11 Jan 02	HG	
Hg	-.017	ppb	.000	-.017				
*** Sample ID: 640								
				Seq: 1832	16:17:28	11 Jan 02	HG	
Hg	-.010	ppb	.000	-.010				
*** Sample ID: 642								
				Seq: 1833	16:19:16	11 Jan 02	HG	
Hg	.009	ppb	.000	.009				
*** Sample ID: 644								
				Seq: 1834	16:22:07	11 Jan 02	HG	
Hg	-.016	ppb	.000	-.016				
*** Sample ID: 646								
				Seq: 1835	16:23:53	11 Jan 02	HG	
Hg	-.052	ppb	.000	-.052				
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		90.7	.907	1.00	ppb	.000		