

WinHg Database 1.3

File Utility Help

RN↓ RN↑ ?

Protocol hg-cv Dataset/Proto RUNS1 /hg-cv

Protocol Line info Cal Curve Report Ctrl Chart Viewer

Reset

Calib Coeffs

New Cal

Update Coeffs

Spike Coeffs

A

B 1.37985e-5

C -2.19088e-2

Rho .999861

Type Linear

Accept

Rel Abs. 721654

Accepted

New


Conc. 10.0

Include S1 Rep 1 ☒ 2 ☐ 3 ☐ 4 ☐ 5 ☐

S	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3
01	.00000	-.035	-.035	-.932	0	-.932		
02	.40000	.351	-.049	27038	0%	27037		
03	1.0000	1.01	.006	74525	0%	74525		
04	3.0000	3.04	.036	221613	0%	221613		
05	5.0000	5.11	.105	371584	0%	371583		
06	10.000	9.94	-.064	721655	0%	721654		

Ready

NUM



1/23/02

westchester county labs

Folder: RUNS1

Page 279

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Standard: 1 Rep: 1				Seq: 1980	14:59:04	23 Jan 02	HG	
Hg	.000	ppb	-932					
*** Standard: 2 Rep: 1				Seq: 1981	15:00:55	23 Jan 02	HG	
Hg	.400	ppb	27037					
*** Standard: 3 Rep: 1				Seq: 1982	15:02:31	23 Jan 02	HG	
Hg	1.00	ppb	74525					
*** Standard: 4 Rep: 1				Seq: 1983	15:04:22	23 Jan 02	HG	
Hg	3.00	ppb	221613					
*** Standard: 5 Rep: 1				Seq: 1984	15:06:00	23 Jan 02	HG	
Hg	5.00	ppb	371583					
*** Standard: 6 Rep: 1				Seq: 1985	15:08:13	23 Jan 02	HG	
Hg	10.0	ppb	721654					
*** Check Standard: 2 Ck21.0 qc				Seq: 1986	15:12:35	23 Jan 02	HG	
Line Flag %Rcv. Found True				Units	SD/RSD			
Hg L 37.8 378 1.00				ppb	.000			
(sample tip error)								
*** Check Standard: 2 Ck21.0 qc				Seq: 1987	15:14:53	23 Jan 02	HG	
Line Flag %Rcv. Found True				Units	SD/RSD			
Hg ✓ 109.1 1.09 1.00				ppb	.000			
*** Sample ID: 1573 pot				Seq: 1988	15:16:31	23 Jan 02	HG	
Hg	.067	ppb	.000	.067				
*** Sample ID: 1220				Seq: 1989	15:18:12	23 Jan 02	HG	
Hg	-.002	ppb	.000	-.002				
*** Sample ID: 1220d				Seq: 1990	15:20:09	23 Jan 02	HG	
Hg L -.035	ppb	.000	-.035					
*** Sample ID: 1220 spk				Seq: 1991	15:22:40	23 Jan 02	HG	
Hg 5.58	ppb	.000	5.58					
*** Sample ID: 1221				Seq: 1992	15:24:26	23 Jan 02	HG	
Hg	-.015	ppb	.000	-.015				
*** Sample ID: 1222				Seq: 1993	15:27:03	23 Jan 02	HG	
Hg	.055	ppb	.000	.055				

westchester county labs

Folder: RUNS1

Page 280

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 1223								
				Seq: 1994	15:28:49	23	Jan 02	HG
Hg	.067	ppb	.000	.067				
*** Sample ID: 1224								
				Seq: 1995	15:30:29	23	Jan 02	HG
Hg	-.024	ppb	.000	-.024				
*** Sample ID: 1225								
				Seq: 1996	15:32:07	23	Jan 02	HG
Hg	-.027	ppb	.000	-.027				
*** Sample ID: 1226								
				Seq: 1997	15:33:47	23	Jan 02	HG
Hg	-.040	ppb	.000	-.040				
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		104. ✓	1.04	1.00	ppb	.000		
*** Sample ID: 1330								
				Seq: 1999	15:37:29	23	Jan 02	HG
Hg	-.004	ppb	.000	-.004				
*** Sample ID: 1331								
				Seq: 2000	15:39:20	23	Jan 02	HG
Hg	.029	ppb	.000	.029				
*** Sample ID: 1332								
				Seq: 2001	15:40:57	23	Jan 02	HG
Hg	.008	ppb	.000	.008				
*** Sample ID: 1333								
				Seq: 2002	15:42:38	23	Jan 02	HG
Hg	-.006	ppb	.000	-.006				
*** Sample ID: 1334								
				Seq: 2003	15:44:50	23	Jan 02	HG
Hg	-.046	ppb	.000	-.046				
*** Sample ID: 1335								
				Seq: 2004	15:47:18	23	Jan 02	HG
Hg	-.051	ppb	.000	-.051				
*** Sample ID: 1436								
				Seq: 2005	15:48:55	23	Jan 02	HG
Hg	-.024	ppb	.000	-.024				
*** Sample ID: 1437								
				Seq: 2006	15:50:34	23	Jan 02	HG
Hg	.013	ppb	.000	.013				

westchester county labs

Folder: RUNS1

Page 281

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 1438								
				Seq: 2007	15:52:30	23 Jan 02	HG	
Hg	.029	ppb	.000	.029				
*** Sample ID: 1439								
				Seq: 2008	15:54:27	23 Jan 02	HG	
Hg	.003	ppb	.000	.003				
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		111.	1.11	1.00	ppb	.000		
*** Sample ID: 1440								
				Seq: 2010	15:57:51	23 Jan 02	HG	
Hg	-.059	ppb	.000	-.059				
*** Sample ID: 1441								
				Seq: 2011	15:59:43	23 Jan 02	HG	
Hg	-.056	ppb	.000	-.056				
*** Sample ID: 1442								
				Seq: 2012	16:01:24	23 Jan 02	HG	
Hg	-.023	ppb	.000	-.023				
*** Sample ID: 1442d								
				Seq: 2013	16:03:10	23 Jan 02	HG	
Hg	.017	ppb	.000	.017				
*** Sample ID: 1442 spk								
				Seq: 2014	16:05:28	23 Jan 02	HG	
Hg	5.62	ppb	.000	5.62				
*** Sample ID: 1498								
				Seq: 2015	16:07:06	23 Jan 02	HG	
Hg	-.035	ppb	.000	-.035				
*** Sample ID: 1499								
				Seq: 2016	16:09:03	23 Jan 02	HG	
Hg	-.063	ppb	.000	-.063				
*** Sample ID: 1500								
				Seq: 2017	16:10:43	23 Jan 02	HG	
Hg	-.049	ppb	.000	-.049				
*** Sample ID: 1501								
				Seq: 2018	16:12:59	23 Jan 02	HG	
Hg	-.027	ppb	.000	-.027				
*** Sample ID: 1502								
				Seq: 2019	16:14:36	23 Jan 02	HG	
Hg	.006	ppb	.000	.006				

westchester county labs

Folder: RUNS1

Page 282

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Check Standard: 2 Ck21.0 qc Seq: 2020 16:16:15 23 Jan 02 HG								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		113.	1.13	1.00	ppb	.000		
*** Sample ID: 1503 Seq: 2021 16:18:02 23 Jan 02 HG								
Hg	.029	ppb	.000	.029				
*** Sample ID: 1504 Seq: 2022 16:19:45 23 Jan 02 HG								
Hg	-.011	ppb	.000	-.011				
*** Sample ID: 1506 Seq: 2023 16:21:37 23 Jan 02 HG								
Hg	-.041	ppb	.000	-.041				
*** Sample ID: 1507 Seq: 2024 16:23:14 23 Jan 02 HG								
Hg	-.052	ppb	.000	-.052				
*** Sample ID: 1509 Seq: 2025 16:24:54 23 Jan 02 HG								
Hg	-.048	ppb	.000	-.048				
*** Sample ID: 1510 Seq: 2026 16:26:37 23 Jan 02 HG								
Hg	-.028	ppb	.000	-.028				
*** Sample ID: 1547 Seq: 2027 16:28:19 23 Jan 02 HG								
Hg	.036	ppb	.000	.036				
*** Sample ID: 1548 Seq: 2028 16:30:26 23 Jan 02 HG								
Hg	.016	ppb	.000	.016				
*** Check Standard: 2 Ck21.0 qc Seq: 2031 16:35:27 23 Jan 02 HG								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		105.	1.05	1.00	ppb	.000		