




1/25/02

westchester county labs

Folder: RUNS1

Page 272

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Standard: 1 Rep: 1				Seq: 2053	15:25:12	25 Jan 02	HG	
Hg	.000	ppb	-678					
*** Standard: 2 Rep: 1				Seq: 2054	15:27:23	25 Jan 02	HG	
Hg	.400	ppb	24219					
*** Standard: 3 Rep: 1				Seq: 2055	15:29:04	25 Jan 02	HG	
Hg	1.00	ppb	66710					
*** Standard: 4 Rep: 1				Seq: 2056	15:30:42	25 Jan 02	HG	
Hg	3.00	ppb	209769					
*** Standard: 5 Rep: 1				Seq: 2057	15:32:30	25 Jan 02	HG	
Hg	5.00	ppb	351511					
*** Standard: 6 Rep: 1				Seq: 2058	15:34:43	25 Jan 02	HG	
Hg	10.0	ppb	638271					
*** Check Standard: 2 Ck21.0 qc				Seq: 2059	15:44:56	25 Jan 02	HG	
Line Flag %Rcv. Found True Units SD/RSD								
Hg 105. ✓ 1.05 1.00 ppb .000								
*** Sample ID: 1610 pot				Seq: 2060	15:47:22	25 Jan 02	HG	
Hg	-.064	ppb	.000	-.064				
*** Sample ID: 1612				Seq: 2061	15:49:01	25 Jan 02	HG	
Hg	-.083	ppb	.000	-.083				
*** Sample ID: 1630				Seq: 2062	15:50:49	25 Jan 02	HG	
Hg	-.065	ppb	.000	-.065				
*** Sample ID: 1632				Seq: 2063	15:52:26	25 Jan 02	HG	
Hg	-.114	ppb	.000	-.114				
*** Sample ID: 1687				Seq: 2064	15:54:34	25 Jan 02	HG	
Hg	-.111	ppb	.000	-.111				
*** Sample ID: 1687 spk				Seq: 2065	15:56:23	25 Jan 02	HG	
Hg 5.12 ¹⁰² ppb .000 5.12								
*** Sample ID: 1687d				Seq: 2066	15:58:03	25 Jan 02	HG	
Hg	-.064	ppb	.000	-.064				

westchester county labs

Folder: RUNS1

Page 274

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5

*** Sample ID: 1801				Seq: 2080	16:24:05	25 Jan 02	HG	
Hg	-.057	ppb	.000	-.057				
=								
*** Check Standard: 2 Ck21.0 qc				Seq: 2081	16:26:32	25 Jan 02	HG	
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		97.6	.976	1.00	ppb	.000		
=								
*** Sample ID: 1802				Seq: 2082	16:28:11	25 Jan 02	HG	
Hg	-.123	ppb	.000	-.123				
=								
*** Check Standard: 2 Ck21.0 qc				Seq: 2092	16:48:22	25 Jan 02	HG	
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		106.	1.06	1.00	ppb	.000		
=								

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 1687d spk				Seq: 2067	15:59:53	25 Jan 02	HG	
Hg	4.78	ppb	.000	4.78				
*** Sample ID: 1688				Seq: 2068	16:01:30	25 Jan 02	HG	
Hg	-.089	ppb	.000	-.089				
*** Sample ID: 1689				Seq: 2069	16:03:12	25 Jan 02	HG	
Hg	-.087	ppb	.000	-.087				
*** Check Standard: 2 Ck21.0 qc				Seq: 2070	16:04:53	25 Jan 02	HG	
Line Flag %Rcv. Found True Units SD/RSD								
Hg	94.6		.946	1.00	ppb	.000		
*** Sample ID: 1690				Seq: 2071	16:07:30	25 Jan 02	HG	
Hg	-.088	ppb	.000	-.088				
*** Sample ID: 1691				Seq: 2072	16:09:08	25 Jan 02	HG	
Hg	-.047	ppb	.000	-.047				
*** Sample ID: 1692				Seq: 2073	16:11:06	25 Jan 02	HG	
Hg	-.056	ppb	.000	-.056				
*** Sample ID: 1693				Seq: 2074	16:13:37	25 Jan 02	HG	
Hg	-.090	ppb	.000	-.090				
*** Sample ID: 1796				Seq: 2075	16:15:27	25 Jan 02	HG	
Hg	-.126	ppb	.000	-.126				
*** Sample ID: 1797				Seq: 2076	16:17:35	25 Jan 02	HG	
Hg	-.131	ppb	.000	-.131				
*** Sample ID: 1798				Seq: 2077	16:19:12	25 Jan 02	HG	
Hg	-.131	ppb	.000	-.131				
*** Sample ID: 1799				Seq: 2078	16:20:49	25 Jan 02	HG	
Hg	-.070	ppb	.000	-.070				
*** Sample ID: 1800				Seq: 2079	16:22:28	25 Jan 02	HG	
Hg	-.074	ppb	.000	-.074				