

westchester county labs

Folder: RUNS1

Page 283

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Standard: 1 Rep: 1				Seq: 2099	15:00:01	30 Jan 02	HG	
Hg	.000	ppb	2535					
*** Standard: 2 Rep: 1				Seq: 2100	15:01:47	30 Jan 02	HG	
Hg	.400	ppb	27228					
*** Standard: 3 Rep: 1				Seq: 2101	15:03:36	30 Jan 02	HG	
Hg	1.00	ppb	58941					
*** Standard: 4 Rep: 1				Seq: 2102	15:05:45	30 Jan 02	HG	
Hg	3.00	ppb	195461					
*** Standard: 5 Rep: 1				Seq: 2103	15:07:23	30 Jan 02	HG	
Hg	5.00	ppb	344157					
*** Standard: 6 Rep: 1				Seq: 2104	15:09:36	30 Jan 02	HG	
Hg	10.0	ppb	686742					
*** Check Standard: 2 Ck21.0 qc				Seq: 2105	15:12:12	30 Jan 02	HG	
Line Flag %Rcv. Found True Units SD/RSD								
Hg	106. ✓	1.06	1.00	ppb	.000			
*** Sample ID: 1910 pot				Seq: 2106	15:14:01	30 Jan 02	HG	
Hg	✓ .143	ppb	.000	.143				
*** Sample ID: 1910 spk				Seq: 2107	15:15:38	30 Jan 02	HG	
Hg	5.32	ppb	.000	5.32				
*** Sample ID: 1910d				Seq: 2108	15:17:14	30 Jan 02	HG	
Hg	✓ .116	ppb	.000	.116				
*** Sample ID: 1910d spk				Seq: 2109	15:18:49	30 Jan 02	HG	
Hg	4.84	ppb	.000	4.84				
*** Sample ID: 1911				Seq: 2110	15:20:29	30 Jan 02	HG	
Hg	.145	ppb	.000	.145				
*** Sample ID: 1912				Seq: 2111	15:22:18	30 Jan 02	HG	
Hg	.180	ppb	.000	.180				
*** Sample ID: 1913				Seq: 2112	15:23:56	30 Jan 02	HG	
Hg	.141	ppb	.000	.141				

1/30/02

westchester county labs

Folder: RUNS1

Page 284

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 1914								
				Seq: 2113	15:25:46	30	Jan 02	HG
Hg	.107	ppb	.000	.107				
*** Sample ID: 1915								
				Seq: 2114	15:27:46	30	Jan 02	HG
Hg	.137	ppb	.000	.137				
*** Sample ID: 1916								
				Seq: 2115	15:29:26	30	Jan 02	HG
Hg	.080	ppb	.000	.080				
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		108.	1.08	1.00	ppb	.000		
*** Sample ID: 1960								
				Seq: 2117	15:33:44	30	Jan 02	HG
Hg	.169	ppb	.000	.169				
*** Sample ID: 1961								
				Seq: 2118	15:35:24	30	Jan 02	HG
Hg	.152	ppb	.000	.152				
*** Sample ID: 1962								
				Seq: 2119	15:37:41	30	Jan 02	HG
Hg	.131	ppb	.000	.131				
*** Sample ID: 1963								
				Seq: 2120	15:39:19	30	Jan 02	HG
Hg	.121	ppb	.000	.121				
*** Sample ID: 1964								
				Seq: 2121	15:41:05	30	Jan 02	HG
Hg	.068	ppb	.000	.068				
*** Sample ID: 1965								
				Seq: 2122	15:43:03	30	Jan 02	HG
Hg	.059	ppb	.000	.059				
*** Sample ID: 1966								
				Seq: 2123	15:44:53	30	Jan 02	HG
Hg	.083	ppb	.000	.083				
*** Sample ID: 1968								
				Seq: 2124	15:46:34	30	Jan 02	HG
Hg	.126	ppb	.000	.126				
*** Sample ID: 1969								
				Seq: 2125	15:48:23	30	Jan 02	HG
Hg	.142	ppb	.000	.142				

westchester county labs

Folder: RUNS1

Page 285

Protocol: hg-cv

POST-RUN REPORT

Line	Conc.	Units	SD/RSD	1	2	3	4	5
*** Sample ID: 2015								
				Seq: 2126	15:50:00	30	Jan 02	HG
Hg	.128	ppb	.000	.128				
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg		113.1	1.13	1.00	ppb	.000		
*** Sample ID: 2016								
				Seq: 2128	15:53:22	30	Jan 02	HG
Hg	.182	ppb	.000	.182				
*** Sample ID: 2145								
				Seq: 2129	15:56:56	30	Jan 02	HG
Hg	.058	ppb	.000	.058				
*** Sample ID: 2145 spk								
				Seq: 2130	15:58:35	30	Jan 02	HG
Hg	5.17	ppb	.000	5.17				
*** Sample ID: 2145d								
				Seq: 2131	16:00:11	30	Jan 02	HG
Hg	.136	ppb	.000	.136				
*** Sample ID: 2145d spk								
				Seq: 2132	16:01:47	30	Jan 02	HG
Hg	5.56	ppb	.000	5.56				
*** Sample ID: 2146								
				Seq: 2133	16:03:55	30	Jan 02	HG
Hg	.133	ppb	.000	.133				
*** Sample ID: 2147								
				Seq: 2134	16:05:37	30	Jan 02	HG
Hg	.121	ppb	.000	.121				
*** Sample ID: 2148								
				Seq: 2135	16:08:43	30	Jan 02	HG
Hg	.051	ppb	.000	.051				
*** Sample ID: 2149								
				Seq: 2136	16:10:30	30	Jan 02	HG
Hg	.089	ppb	.000	.089				
*** Sample ID: 2150								
				Seq: 2137	16:12:19	30	Jan 02	HG
Hg	.101	ppb	.000	.101				
*** Check Standard: 2 Ck21.0 qc								
Line	Flag	%Rcv.	Found	True	Units	SD/RSD		
Hg	H	116.5	1.16	1.00	ppb	.000		

(O.K. to report LML results)

westchester county labs

Folder: RUNS1
Protocol: hg-cv
POST-RUN REPORT

Page 286

Line	Conc.	Units	SD/RSD	1	2	3	4	5	
*** Standard: 1 Rep: 1									
				Seq:	2140	16:21:02	30 Jan 02	HG	
Hg	.000	ppb	-494						=
*** Standard: 2 Rep: 1									
				Seq:	2141	16:22:40	30 Jan 02	HG	
Hg	.400	ppb	30973						=
*** Standard: 3 Rep: 1									
				Seq:	2142	16:24:40	30 Jan 02	HG	
Hg	1.00	ppb	68663						=
*** Standard: 4 Rep: 1									
				Seq:	2143	16:26:20	30 Jan 02	HG	
Hg	3.00	ppb	217043						=
*** Standard: 5 Rep: 1									
				Seq:	2144	16:27:59	30 Jan 02	HG	
Hg	5.00	ppb	371161						=
*** Standard: 6 Rep: 1									
				Seq:	2145	16:30:41	30 Jan 02	HG	
Hg	10.0	ppb	714663						=
*** Check Standard: 2 Ck21.0 qc									
Line	Flag	%Rcv.	Found	True	Units	SD/RSD			
Hg		94.6 ✓	.946	1.00	ppb	.000			=
*** Sample ID: 2151									
				Seq:	2147	16:36:16	30 Jan 02	HG	
Hg	-.004	ppb	.000						=
*** Check Standard: 2 Ck21.0 qc									
Line	Flag	%Rcv.	Found	True	Units	SD/RSD			
Hg		102. ✓	1.02	1.00	ppb	.000			=