

Daily Performance Report

Sample ID: DPC

Sample Date/Time: Wednesday, December 19, 2001 07:58:39

Sample Description: DAILY PERFORMANCE CHECK

Method File: d:\Elandata\Method\DAIly PERF TO SAVE REPORT.mth

Dataset File: d:\Elandata\Dataset\daily performance\DPC.1090

Tuning File: d:\Elandata\Tuning\default.tun

Optimization File: d:\Elandata\Optimize\default.dac

Dual Detector Mode: Pulse

Acq. Dead Time(ns): 35

Current Dead Time (ns): 35

Summary

Analyte	Mass	Meas. Intens. Mean	Net Intens. Mean	Net Intens. SD	Net Intens. RSD
Mg	24.0	52142.6	52142.589 ✓	333.784	0.6
Rh	102.9	339883.5	339883.468 ✓	3371.696	1.0
In	114.9	486151.5	486151.537	3798.239	0.8
Pb	208.0	301052.0	301052.043 ✓	1403.797	0.5
[> Ba	137.9	448849.5	448849.545	2928.793	0.7
[Ba++	69.0	6223.1	0.014 ✓	0.000	0.8
[> Ce	139.9	547004.7	547004.742	4670.142	0.9
[CeO	155.9	15731.1	0.029 ✓	0.000	1.1
Bkgd	220.0	17.7 ✓	17.734	1.205	6.8

Current Optimization File Data

Current Value	Description
0.87	Nebulizer Gas Flow
5.75	Lens Voltage
1000.00	ICP RF Power
-1950.00	Analog Stage Voltage
1300.00	Pulse Stage Voltage
70.00	Discriminator Threshold
-6.00	AC Rod Offset
60.00	Service DAC 1
0.00	Quadrupole Rod Offset

Current Autolens Data

Analyte	Mass	Num of Pts	DAC Value	Maximum Intensity
Be	9	21	4.8	8680.2
Al	27	21	5.0	264330.4
Sc	45	21	5.0	1390659.5
V	51	21	5.3	314490.6
Cr	52	20	5.3	305433.6
Mn	55	21	5.3	473883.0
Ni	58	20	5.3	243380.1
Co	59	21	5.3	360386.7
Cu	63	21	5.3	206675.8
Zn	64	21	5.3	73297.5
Ge	74	21	5.3	320871.7
As	75	21	5.3	37199.7
Se	82	21	5.5	4007.6
Mo	98	21	5.5	139614.3
Ag	107	21	5.5	251831.5

Sample ID: DPC

Report Date/Time: Wednesday, December 19, 2001 08:02:05

Page 1

Cd	114	21	5.5	116126.5
In	115	13	7.0	1840544.5
Sb	121	21	5.5	146066.9
Ba	138	21	5.8	628568.4
Ho	165	8	4.8	1775794.9
Hg	202	21	7.0	48066.7
Tl	205	21	7.0	645387.9
Pb	208	21	6.8	454952.8
Bi	209	10	8.3	1831789.2