

Method 508Date Extracted 11/6/01Page 78

Lab #	Description	Sample vol/wt	Drip rate ml/min	Surr add	Spk add	IS add	Final vol.	T.V. pos.	pH = 7	Count
	LAB BLANK	500 mL		1 mL			2 mL			
	Ref CHL + Post									
	Ref TOX									
	PAH 1232									
26530	S CHL				1 mL					
26532	S D CHL									
26533	S TOX									
26533	S D TOX									
26536	S Post									
26536	S D Post									
	26475									1
	26476									2
	26477									3
	26478									4
	21532									5
	26533									6
	26534									7
	26535									8
	26536									9
	26537									10
Remarks										
Reviewed by <u>AS 3/13/02</u>										
drip rate must be between 5-15 ml/min for liq-liq extractors										
drip rate -										
Extraction type - <u>Solvent</u> Solv. lot# <u>V42 E17</u> NaCl lot# <u>V11599</u> Na2SO4 lot# <u>V32603</u>										
STANDARD	INT STD LOG#									
CALIBRATION										
REF/SPIKE	<u>TOX: 1992</u>	<u>CHL: 1993</u>								<u>POST: 2026</u>
MDL/MDRF										
Q. CHECK										
INT. STD.										
SURROGATE	<u>1839b</u>									
LPC										
ENDRIN/DDT										

Analyst ML/AD
revision 4/12/01

Verified by _____

Reviewed by DN
Date 12/26/11