

User: Nefliu, Marcela
 Westchester Dept. of Labs & Research
 Date: 12/12/2001 Time: 16:34:59

Sample: AD26538
 Analysis: \$LP508 Pesticides Lab Performance Check
 Units: ug
 Started: 11/7/01 16:20 Ended: 11/30/01 16:20 Analyst: MN
 Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	PGF		0.99		0.80-1.1
2	Resolution		1.6		>0.50
3	S/N		130		>3

Quantitation Report

(Not Reviewed)

Signal #1 : C:\HPCHEM\1\DATA\508104\011130D\1130LPC.D\ECD1A.CH Vial: 6
Signal #2 : C:\HPCHEM\1\DATA\508104\011130D\1130LPC.D\ECD2B.CH
Acq On : 30 Nov 2001 5:11 pm Operator: Nefliu
Sample : LPC Inst : 104
Misc : Multiplr: 1.00
IntFile Signal #1: events.e IntFile Signal #2: events2.e
Quant Time: Nov 30 18:04 2001 Quant Results File: 508F2.RES

Quant Method : C:\HPCHEM\1\METHODS\508F2.M (Chemstation Integrator)
Title : 508 CHLORINATED PESTICIDES 9/6/00
Last Update : Tue Oct 31 13:22:28 2000
Response via : Initial Calibration
DataAcq Meth : 508F2.M

Volume Inj. :
Signal #1 Phase : Signal #2 Phase:
Signal #1 Info : Signal #2 Info :

Compound	RT#1	RT#2	Resp#1	Resp#2	ug/L	ug/L

System Monitoring Compounds						
9) S 2,4-DDD SURROGAT	38.56	34.37	2003	6304	0.001	0.001 #
Spiked Amount	0.200		Recovery	=	0.50%	0.50%
Target Compounds						
3) HEPTACHLOR	26.36	22.28	1237826	6565	0.513	0.001 #
4) ALDRIN	28.45f	24.75	331028	37367	0.056	0.003 #
5) HEPTACHLOR EPOXI	33.01	0.00	1119	0	0.000	N.D. #
6) DIELDRIN	0.00	33.72f	0	1081	N.D.	0.000 #
7) ENDRIN	38.56	34.37	2003	6304	0.001	0.001 #
8) METHOXYCHLOR	0.00	41.51	0	4478	N.D.	0.002 #

$\frac{S}{N} = 130$
 $R = 0.99$
 $PGF = 1.6$
 0.99

Quantitation Report

Signal #1 : C:\HPCHEM\1\DATA\508104\011130D\1130LPC.D\ECD1A.CH Vial: 6
Signal #2 : C:\HPCHEM\1\DATA\508104\011130D\1130LPC.D\ECD2B.CH
Acq On : 30 Nov 2001 5:11 pm Operator: Nefliu
Sample : LPC Inst : 104
Misc : Multiplr: 1.00
IntFile Signal #1: events.e IntFile Signal #2: events2.e
Quant Time: Nov 30 18:04 2001 Quant Results File: 508F2.RES

Quant Method : C:\HPCHEM\1\METHODS\508F2.M (Chemstation Integrator)
Title : 508 CHLORINATED PESTICIDES 9/6/00
Last Update : Tue Oct 31 13:22:28 2000
Response via : Multiple Level Calibration
DataAcq Meth : 508F2.M

Volume Inj. :
Signal #1 Phase : Signal #2 Phase:
Signal #1 Info : Signal #2 Info :

