

User: Nefliu, Marcela
Westchester Dept. of Labs & Research
Date: 12/12/2001 Time: 16:41:43

Sample: AD26679
Analysis: \$508 Chlorine Pesticides - 508
Units: ug
Started: 11/7/01 16:38 Ended: 11/30/01 16:38 Analyst: MN
Result source: manual entry
Page: 1

	Component Name	Viol	Result	Qualifier	MDL
1	Propachlor		< MDL		0.040
2	Lindane		< MDL		0.030
3	Heptachlor		< MDL		0.020
4	Aldrin		< MDL		0.010
5	Heptachlor Epoxide		< MDL		0.030
6	4,4-DDE		< MDL		0.080
7	Dieldrin		< MDL		0.016
8	Endrin		< MDL		0.020
9	Methoxychlor		< MDL		0.039
10	Chlordane		< MDL		0.40
11	Toxaphene		< MDL		0.40
12	PCB's		< MDL		0.20
13	Surr_2,4-DDD		0.21		

Quantitation Report

(QT Reviewed)

Signal #1 : C:\HPCHEM\1\DATA\508\113001\26679.D\ECD1A.CH Vial: 62
 Signal #2 : C:\HPCHEM\1\DATA\508\113001\26679.D\ECD2B.CH
 Acq On : 2 Dec 2001 5:12 am Operator: Nefliu
 Sample : sample Inst : 105
 Misc : Multiplr: 1.00
 IntFile Signal #1: EVENTS.E IntFile Signal #2: EVENTS.E
 Quant Time: Dec 10 13:09 2001 Quant Results File: 121001P2.RES

Quant Method : C:\HPCHEM\1\METHODS\508\121001P2.M (Chemstation Integrator)
 Title : 508 Chlorinated Pesticides
 Last Update : Mon Dec 10 13:04:49 2001
 Response via : Initial Calibration
 DataAcq Meth : PESTPCB.M

Volume Inj. :
 Signal #1 Phase : Signal #2 Phase:
 Signal #1 Info : Signal #2 Info :

Compound	RT#1	RT#2	Resp#1	Resp#2	ug/l	ug/l

System Monitoring Compounds						
10) S 2,4 DDD Surrogat	21.61	21.90	3382808.	255715	0.181m	0.207
Spiked Amount	0.200		Recovery	=	90.50%	103.50%
Target Compounds						
1) Propachlor	14.30	13.81	103504	440	0.034	0.002 #
3) Heptachlor	0.00	17.19	0	3046	N.D. d	0.001
4) Aldrin	17.93	0.00	7064	0	0.000	N.D. d#
5) Heptachlor Epoxi	19.74	19.91f	62981	148	0.001	0.000 #
6) 4,4-DDE	20.87	21.14	28834	6017	0.001	0.002 #
7) Dieldrin	21.37	21.48	274	9893	0.000	0.004 #
8) Endrin	22.05f	22.38f	72351	992	0.003	0.001 #

 (f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

26679.D 121001P2.M Mon Dec 10 13:09:24 2001


Page 1

Quantitation Report

Signal #1 : C:\HPCHEM\1\DATA\508\113001\26679.D\ECD1A.CH Vial: 62
 Signal #2 : C:\HPCHEM\1\DATA\508\113001\26679.D\ECD2B.CH
 Acq On : 2 Dec 2001 5:12 am Operator: Nefliu
 Sample : sample Inst : 105
 Misc : Multiplr: 1.00
 IntFile Signal #1: EVENTS.E IntFile Signal #2: EVENTS.E
 Quant Time: Dec 10 13:09 2001 Quant Results File: 121001P2.RES

Quant Method : C:\HPCHEM\1\METHODS\508\121001P2.M (Chemstation Integrator)
 Title : 508 Chlorinated Pesticides
 Last Update : Mon Dec 10 13:04:49 2001
 Response via : Multiple Level Calibration
 DataAcq Meth : PESTPCB.M

Volume Inj. :
 Signal #1 Phase : Signal #2 Phase:
 Signal #1 Info : Signal #2 Info :



Quantitation Report

(QT Reviewed)

Signal #1 : C:\HPCHEM\1\DATA\508\113001\26679.D\ECD1A.CH Vial: 62
Signal #2 : C:\HPCHEM\1\DATA\508\113001\26679.D\ECD2B.CH
Acq On : 2 Dec 2001 5:12 am Operator: Nefliu
Sample : sample Inst : 105
Misc : Multiplr: 1.00
IntFile Signal #1: events.e IntFile Signal #2: events2.e
Quant Time: Dec 10 15:02 2001 Quant Results File: 121001C.RES

Quant Method : C:\HPCHEM\1\METHODS\508\121001C.M (Chemstation Integrator)
Title : 508 CHLORDANE
Last Update : Wed Nov 28 13:19:08 2001
Response via : Initial Calibration
DataAcq Meth : PESTPCB.M

Volume Inj. :
Signal #1 Phase : Signal #2 Phase:
Signal #1 Info : Signal #2 Info :

Compound	RT#1	RT#2	Resp#1	Resp#2	ug/L	ug/L

System Monitoring Compounds						
4) S 2,4-DDD Surrogat	21.61	21.90	3541107	246789	0.146	0.166
Spiked Amount 0.200			Recovery	=	73.00%	83.00%
Target Compounds						
1) Chlordane 1	20.00	16.68	3552	4238	0.004	0.033 #
2) Chlordane 2	20.51	0.00	340	0	0.000	N.D. d#
3) Chlordane 3	20.67	0.00	159	0	0.000	N.D. d#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

26679.D 121001C.M Mon Dec 10 15:02:57 2001


Page 1

Quantitation Report

Signal #1 : C:\HPCHEM\1\DATA\508\113001\26679.D\ECD1A.CH Vial: 62
Signal #2 : C:\HPCHEM\1\DATA\508\113001\26679.D\ECD2B.CH
Acq On : 2 Dec 2001 5:12 am Operator: Nefliu
Sample : sample Inst : 105
Misc : Multiplr: 1.00
IntFile Signal #1: events.e IntFile Signal #2: events2.e
Quant Time: Dec 10 15:02 2001 Quant Results File: 121001C.RES

Quant Method : C:\HPCHEM\1\METHODS\508\121001C.M (Chemstation Integrator)
Title : 508 CHLORDANE
Last Update : Wed Nov 28 13:19:08 2001
Response via : Multiple Level Calibration
DataAcq Meth : PESTPCB.M

Volume Inj. :
Signal #1 Phase : Signal #2 Phase:
Signal #1 Info : Signal #2 Info :



Quantitation Report

(QT Reviewed)

Signal #1 : C:\HPCHEM\1\DATA\508\113001\26679.D\ECD1A.CH Vial: 62
Signal #2 : C:\HPCHEM\1\DATA\508\113001\26679.D\ECD2B.CH
Acq On : 2 Dec 2001 5:12 am Operator: Nefliu
Sample : sample Inst : 105
Misc : Multiplr: 1.00
IntFile Signal #1: events.e IntFile Signal #2: events2.e
Quant Time: Dec 10 12:14 2001 Quant Results File: 121001T.RES

Quant Method : C:\HPCHEM\1\METHODS\508\121001T.M (Chemstation Integrator)
Title : 508 Toxaphene
Last Update : Fri Dec 07 12:12:37 2001
Response via : Initial Calibration
DataAcq Meth : PESTPCB.M

Volume Inj. :
Signal #1 Phase : Signal #2 Phase:
Signal #1 Info : Signal #2 Info :

Compound	RT#1	RT#2	Resp#1	Resp#2	Conc#1	Conc#2

System Monitoring Compounds						
6) S 2,4-DDD Surrogat	21.61	21.90	3750228	254569	0.139	0.178 #
Spiked Amount	0.200		Recovery	=	69.50%	89.00%

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

26679.D 121001T.M Mon Dec 10 12:14:35 2001



Page 1

Quantitation Report

Signal #1 : C:\HPCHEM\1\DATA\508\113001\26679.D\ECD1A.CH Vial: 62
Signal #2 : C:\HPCHEM\1\DATA\508\113001\26679.D\ECD2B.CH
Acq On : 2 Dec 2001 5:12 am Operator: Nefliu
Sample : sample Inst : 105
Misc : Multiplr: 1.00
IntFile Signal #1: events.e IntFile Signal #2: events2.e
Quant Time: Dec 10 12:14 2001 Quant Results File: 121001T.RES

Quant Method : C:\HPCHEM\1\METHODS\508\121001T.M (Chemstation Integrator)
Title : 508 Toxaphene
Last Update : Fri Dec 07 12:12:37 2001
Response via : Multiple Level Calibration
DataAcq Meth : PESTPCB.M

Volume Inj. :
Signal #1 Phase : Signal #2 Phase:
Signal #1 Info : Signal #2 Info :



Quantitation Report

(QT Reviewed)

Signal #1 : C:\HPCHEM\1\DATA\508105\113001\26679.D\ECD1A.CH Vial: 62
Signal #2 : C:\HPCHEM\1\DATA\508105\113001\26679.D\ECD2B.CH
Acq On : 2 Dec 2001 5:12 am Operator: Nefliu
Sample : sample Inst : 105
Misc : Multiplr: 1.00
IntFile Signal #1: EVENTS.E IntFile Signal #2: events2.e
Quant Time: Dec 12 14:13 2001 Quant Results File: PCB1242.RES

Quant Method : C:\HPCHEM\1\METHODS\508105\PCB1242.M (Chemstation Integrator)
Title : Arochlor 1242
Last Update : Wed Dec 12 14:08:13 2001
Response via : Initial Calibration
DataAcq Meth : PESTPCB.M

Volume Inj. :
Signal #1 Phase : Signal #2 Phase:
Signal #1 Info : Signal #2 Info :

Compound	RT#1	RT#2	Resp#1	Resp#2	ug/l	ug/l

System Monitoring Compounds						
6) S 2,4-DDD Surrogat	21.61	21.90	3389895	254174	0.200m	0.206m
Spiked Amount	0.200		Recovery	, =	100.00%	103.00%
Target Compounds						
1) 1242-1	14.30	14.75	103504	6490	0.147	0.221 #
2) 1242-2	15.47	16.07	59992	4886	0.042	0.049
3) 1242-3	16.87	17.19	4979	3046	0.002	0.031 #
5) 1242-5	18.42	20.65	1075040	12088	0.623	0.350 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

26679.D PCB1242.M Wed Dec 12 14:13:09 2001


Page 1

Quantitation Report

Signal #1 : C:\HPCHEM\1\DATA\508105\113001\26679.D\ECD1A.CH Vial: 62
Signal #2 : C:\HPCHEM\1\DATA\508105\113001\26679.D\ECD2B.CH
Acq On : 2 Dec 2001 5:12 am Operator: Nefliu
Sample : sample Inst : 105
Misc : Multiplr: 1.00
IntFile Signal #1: EVENTS.E IntFile Signal #2: events2.e
Quant Time: Dec 12 14:13 2001 Quant Results File: PCB1242.RES

Quant Method : C:\HPCHEM\1\METHODS\508105\PCB1242.M (Chemstation Integrator)
Title : Arochlor 1242
Last Update : Wed Dec 12 14:08:13 2001
Response via : Multiple Level Calibration
DataAcq Meth : PESTPCB.M

Volume Inj. :
Signal #1 Phase : Signal #2 Phase:
Signal #1 Info : Signal #2 Info : ,



• **Comments for Sample AD26679 - Analysis \$508**

Westchester Dept. of Labs & Research

User: Vinci, David L

Date: 12-13-2001 Time: 11:36:15

Recoveries for 9 of the 10 compounds in the LCS Standard were below their acceptance ranges.