

Method 508Date Extracted 11/2/01Page

Lab #	Description	Sample	Drip rate	Surr	Spk	IS	Final	T.V.	pH	Count		
		vol/wt	ml/min	add	add	add	vol.	pos.	= 7			
	LB	500ML		1mL			2mL					
	Ref chL/Ref											
	Ref TOX											
	Ref chL/Ref											
	PLL 1016											
	26079									1		
	26080									2		
	26127									3		
	26128									4		
	26129									5		
	26130									6		
	26131									7		
	26132									8		
	26133									9		
	26133									10		
26079	SPK chL											
26079	S chL dup											
26080	SPK TOX											
26080	SPK TOX dup											
26127	SPK Pest											
26127	S Pest dup											
Remarks												
SPK dupe chL / chL												
SPK dupe TOX / TOX												
Reviewed by: <u>JB</u> 1/13/02												
drip rate must be between 5-15 ml/min for liq-liq extractors												
drip rate -												
Extraction type - <u>SEP Pannel</u> Solv. lot# <u>V42E17</u> NaCl lot# <u>V1599</u> Na2SO4 lot# <u>V32603</u>												
STANDARD	INT STD LOG#											
CALIBRATION												
REF/SPIKE												
MDL/MDRF												
Q. CHECK												
INT. STD.												
SURROGATE												
LPC												
ENDRIN/DDT												

Analyst ML/CM/AD
revision 4/12/01Verified by Reviewed by DW
Date 12/15/11

