

Quantitation Report (QT Reviewed)

Signal #1 : C:\HPCHEM\1\DATA\508F\011130D\2PESTC2.D\ECD1A.CH Vial: 78
 Signal #2 : C:\HPCHEM\1\DATA\508F\011130D\2PESTC2.D\ECD2B.CH
 Acq On : 2 Dec 2001 3:55 am Operator: Nefliu
 Sample : pest ccv1 Inst : 104
 Misc : Multiplr: 1.00
 IntFile Signal #1: events.e IntFile Signal #2: events2.e
 Quant Time: Dec 11 8:27 2001 Quant Results File: 011204P.RES

Quant Method : C:\HPCHEM\1\METHODS\508\011204P.M (Chemstation Integrator)
 Title : 508 CHLORINATED PESTICIDES 9/6/00
 Last Update : Sun Oct 21 15:26:52 2001
 Response via : Initial Calibration
 DataAcq Meth : 508F2.M

Volume Inj. :
 Signal #1 Phase : Signal #2 Phase:
 Signal #1 Info : Signal #2 Info :

Compound	RT#1	RT#2	Resp#1	Resp#2	ug/L	ug/L

System Monitoring Compounds						
10) S 2,4-DDD SURROGAT	36.60	33.62	998363	1033859	0.209m	0.197m
Spiked Amount 0.200			Recovery	=	104.50%	98.50%
Target Compounds						
1) PROPACHLOR	16.36	15.23	198359	202641	0.238 ¹¹⁵	0.215m 108%.
2) g-BHC	22.45	20.11	2855652	3637498	0.327 ¹¹⁴	0.297 106%.
3) HEPTACHLOR	25.10	21.49	638532	954615	0.110 ¹¹⁵	0.096 100%.
4) ALDRIN	27.78	23.53	1375744	1523781	0.133 ¹¹¹	0.123 103%.
5) HEPTACHLOR EPOXI	31.77	28.81	2316537	2608856	0.273 ¹¹⁴	0.241 100%.
6) 4,4'-DDE	35.25	31.98	3694209	3887171	0.478m ¹²⁰	0.378 95%.
7) DIELDRIN	35.52	32.64	696140	815495	0.095m ¹¹⁹	0.076 95%.
8) ENDRIN	37.50	33.69	1110996	1536622	0.341 ¹²¹	0.208m# 104
9) METHOXYCHLOR	48.37	40.58	758060	987899	0.514 ¹²⁹	0.324m# 81%.

ok
 mid ccv?

Quantitation Report

Signal #1 : C:\HPCHEM\1\DATA\508F\011130D\2PESTC2.D\ECD1A.CH Vial: 78
Signal #2 : C:\HPCHEM\1\DATA\508F\011130D\2PESTC2.D\ECD2B.CH
Acq On : 2 Dec 2001 3:55 am Operator: Nefliu
Sample : pest ccv1 Inst : 104
Misc : Multiplr: 1.00
IntFile Signal #1: events.e IntFile Signal #2: events2.e
Quant Time: Dec 11 8:27 2001 Quant Results File: 011204P.RES

Quant Method : C:\HPCHEM\1\METHODS\508\011204P.M (Chemstation Integrator)
Title : 508 CHLORINATED PESTICIDES 9/6/00
Last Update : Sun Oct 21 15:26:52 2001
Response via : Multiple Level Calibration
DataAcq Meth : 508F2.M

Volume Inj. :
Signal #1 Phase :
Signal #1 Info :
Signal #2 Phase :
Signal #2 Info :

