

p.14

P.11/12

212 679 9392

JOSE JIMENEZ 201-943-9256
SCIENTIFIC LABS

Monday, December 10, 2001 5:42 PM
NDU-07-2001 09:42

SciLab Job #: 201111618

Client Name: Applied Technology Services, Inc.

Table I

Summary of Transmission Electron Microscopy (TEM) Results for Asbestos (air)

A01200-C; Apartment 4N

SciLab sample #	Client Sample #	Dilution Factor	Air Filtered (liters)	Area Analyzed (sq. mm.)	* Analytical Sensitivity (struc/cc air)	Asbestos Structures Detected (Microns)			Structure Density (struc/sq. mm.)		Structure Concentration (struc/cc air)		Type of Asbestos
						0.5-5.0	>5.0	Total	>5.0	Total	>5.0	Total	
01	29 11/6/01- 01 4th Floor, South West Corner		600	.1	0.0064	0.0	0.0	0.0	<10	<10	<0.0064	<0.0064	NSD

concentration represented by the detection of 1 structure
* not analyzed

NSD: No Asbestos Structures Detected

Reviewed By: _____

; Analyzed By: _____

Aleksandr Barengolts

Date: 11/7/01

NVLAP#: 200546-0