

10/02/2002 07:44 FAX

010

**ATHENICA**ENVIRONMENTAL SERVICES INC.
Tel: (718) 704-7490
Fax: (718) 704-4085**AIRBORNE ASBESTOS ANALYSIS BY TRANSMISSION ELECTRON MICROSCOPY**

CLIENT: Warren & Panzer Engineers, P.C.

LABORATORY ID #: T02-09-114

ANALYT. METHODOLOGY: AHERA

FILTER TYPE: M.C.B.

AVERAGE G.O. AREA (mm²): 0.013838

PROJECT: NYC DEP, 90/92 Fulton Street

DATE OF REPORT: 09/30/02

DATE OF ANALYSIS: 09/30/02

EFFECTIVE FILTER AREA (mm²): 385**LABORATORY RESULTS**

CLIENT #	LAB. ID	LOCATION	VOLUME (ft ³)	G.O.'s ANALYZED	AREA ANALYZED (mm ²)	ANALYT. SENSITIVITY (fibre/cm ²)	TOTAL ASBESTOS AIR CONC. (fibre/cm ³)	TOTAL ASBESTOS FILTER CONC. (fibre/cm ²)	AIR CONC. FOR STRUCT. U.S.S.-CLIM	AIR CONC. FOR STRUCT. SC-CLIM	ASBEST. TYPE CHRYS.	AMPHIB. ASBEST. TYPE SPECIFY
01	T02-09-114-2660	East side of facade	1680	4	0.0554	0.0041	<0.0041	<18.07	NA	NA	NA	NA
02	T02-09-114-2661	Center 1	1680	4	0.0554	0.0041	<0.0041	<18.07	NA	NA	NA	NA
03	T02-09-114-2662	Center 2	1680	4	0.0554	0.0041	<0.0041	<18.07	NA	NA	NA	NA
04	T02-09-114-2663	Center 3	1680	4	0.0554	0.0041	<0.0041	<18.07	NA	NA	NA	NA
05	T02-09-114-2664	West side of facade	1680	4	0.0554	0.0041	<0.0041	<18.07	NA	NA	NA	NA

*E. Sioulaki*ANALYST
E. Sioulaki*Spino Domingos*LABORATORY DIRECTOR
Spino Domingos**LABORATORY CERTIFICATION NUMBERS: NVLAP 101958, ELAP 10955**

- Athenica Environmental Services Inc. (AES), is responsible only for information pertaining to samples taken by its employees.
- Air sampling cassettes will be stored for sixty (60) days. AES Inc., should be notified within this time frame for a duplicate analysis.
- If AES Inc. did not conduct the sampling, the reported airborne concentration was derived based on information provided by the client. Under these circumstances, the air concentration data is supplied solely for the convenience of the client.
- The report relates only to tickets tested. This report must not be used as claim product endorsement by NVLAP or any other agency of the U.S. Government. Test reports may not be reproduced except in full and with prior approval of AES Inc.
- The liability of Athenica Environmental Services Inc. with respect to the services charged, shall in no event exceed the amount of the invoice.

10/02/2002 07:44 FAX

011

226 East 48th Street
New York, NY 10017
(212) 922-0077
Fax (212) 922-0630

WARREN & PANZER ENGINEERS, P. C.

AIR MONITORING LOG SHEET/CHAIN-OF-CUSTODY FORM

PAGE ____ OF ____

P.O. _____

24 hr⁺/a 12 hr⁺/a 6 hr⁺/a 3 hr⁺/a Immediate ⁺/a

CLIENT: NYC DEP

W&P FILE #: 1079-01-02

CONTACT: _____

PROJECT: FLORIT PACADSE

LABORATORY: _____

LOCATION: 90/92 FULTON ST

ANALYSIS METHOD: PCM TEM TEM
circle one 7400 AHERM EPA LEVEL II

TECHNICIAN: FESTY DEWELLE

SAMPLE NUMBER	SAMPLE DESCRIPTION/LOCATION	SAMPLE TYPE	PUMP #	FLOWRATE START STOP	ON	TIME OFF	TOTAL TIME	VOLUME (liters)	CONCENTR. (lb./cc.)
01	EAST SIDE OF FULTON	3		8 8	0940	1310	210	1680	
02	CENTER 1	3		8 8	0941	1311	210	1680	
03	CENTER 2	3		8 8	0942	1312	210	1680	
04	CENTER 3	3		8 8	0943	1313	210	1680	
05	WEST SIDE OF FULTON	3		8 8	0944	1314	210	1680	
06	REAR								
07	11								
08	11								

COMMENTS:

102-09-114

SAMPLE BY: FESTY DEWELLE DATE: 09-29-02

SIGNATURE: [Signature]

RELEASED BY: FESTY DEWELLE DATE: 09-29-02

SIGNATURE: [Signature]

RECEIVED BY: ASABOAL DATE: 9/30/02

SIGNATURE: [Signature]

SAMPLE TYPE CODES	
1 - Blank	5 - Final Clearance (inside)
2 - Pre-Abatement	6 - Final Clearance (outside)
3 - During Abatement	7 - Personal
4 - During Glove Bag Removal	8 - Background

SEE REVERSE SIDE FOR ANALYSIS INFORMATION