

Aug 17 04 02:22p

IBM USER

732 6018

P. 5

RJ Lee Group, Inc. 100 Hochberg Road

Date: 08/12/2004 Time: 11:41:02 AM

Page 2 of 3

8/10

RJ LeeGroup, Inc.

350 Hochberg Road, Monroeville, PA 15146

Phone: (724) 325-1776 Fax: (724) 733-1799

Final Laboratory Report**TEM Air Analysis**

RJ Lee Group, Inc.
350 Hochberg Road
Monroeville, PA 15146
USA

Report Date: 8/12/2004
Sample Receipt Date: 8/11/2004
RJ Lee Group Job No.: ATH306366-989
Samples Received: 0
Building: 130 Liberty

Method: Federal Register, 40 CFR, Part 763, October 30, 1987

RJL Sample Number	Sample Type	Date Collected Start Stop	Floor	Grid	Sub Location	Filter Area (mm ²)	Volume (liters)	Area Analyzed (mm ²)	Asbestos Structures Chry Amph	Concentration	
										Analytical Sensitivity (S/mm ²)	(S/cc)
228912.HT	Area	8/10/2004 8/10/2004	99	Not Acquired	(SouthWest) - Washington and Albany	335	1924.00	0.04466	0 0	< 22.4 22.4	< 0.004 0.004
228913.HT	Area	8/10/2004 8/10/2004	99	Not Acquired	(SouthEast) - Greenwich and Albany	335	1932.00	0.04466	0 0	< 22.4 22.4	< 0.004 0.004
228914.HT	Area	8/10/2004 8/10/2004	99	N/A	(East) - Greenwich	335	1948.00	0.04466	0 0	< 22.4 22.4	< 0.004 0.004
228915.HT	Area	8/10/2004 8/10/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	335	1968.00	0.04466	0 0	< 22.4 22.4	< 0.004 0.004
228911.HT	Field Blank	8/10/2004 8/10/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	335	0.000	0.08932	0 0	< 11.2 11.2	N/A N/A

Authorized Signature.



/s/b

Kimberly A. Allison, Manager - TEM Analysis

Notes:

- Volumes provided by Ambient were used to calculate analytical results and sensitivities
- "-" indicates results less than analytical sensitivity. "N/A" indicates that sample was not analyzed.
- Sample(s) for this project were analyzed at our Monroeville, PA (NVLAP #101208-0, NY ELAP #10844) facility.
- If RJ Lee Group, Inc. did not collect the samples analyzed, the verifiability of the laboratory's results are limited to the reported values.
- Abbreviations: N/A-Not Applicable, Chry-Chrysotile Asbestos, Amph-Amphibole Asbestos, S-Structures.
- Samples will be held for 90 days and then disposed of per Federal regulations.
- These results are submitted pursuant to RJ Lee Group's current terms and conditions of sale, including the company's standard warranty and limitation of liability provisions. No responsibility or liability is assumed for the manner in which these results are used or interpreted.

Aug 17 04 02:25p

IBM USER

732 6018

p. 1

RJ LeeGroup, Inc.

350 Hochberg Road, Monroeville, PA 15146
Phone: (724) 325-1776 Fax: (724) 733-1799

Final Laboratory Report TEM Air Analysis

RJ Lee Group, Inc.
350 Hochberg Road
Monroeville, PA 15146
NJ

Report Date: 8/13/2004
Sample Receipt Date: 8/12/2004
RJ Lee Group Job No.: ATH306365-993
Samples Received: 10
Building: 130 Liberty

United: Federal Register, 40 CFR, Part 763, October 30, 1987

RJLG Sample Number	Sample Type	Date Collected		Floor	Grid	Sub Location	Filter Area (nm ²)	Volume (liter)	Area Analyzed (nm ²)	Asbestos Structures		Concentration Analytical Sensitivity	
		Start	Stop							Chry	Amph	(S/nm ²)	(S/cc)
228543 HT	Field Blank	8/10/2004	8/10/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	335	0.000	0.08949	0	0	11.2 11.2	N/A N/A
228544 HT	Area	8/10/2004	8/10/2004	99	Not Acquired	(SouthWest) - Washington and Albany	335	1820.00	0.04475	0	0	22.3 22.3	< 0.005 0.005
228545 HT	Area	8/10/2004	8/10/2004	99	Not Acquired	(SouthEast) - Greenwich and Albany	335	1812.00	0.04475	0	0	22.3 22.3	< 0.005 0.005
228546 HT	Area	8/10/2004	8/10/2004	99	N/A	(East) - Greenwich	335	1796.00	0.04475	0	0	22.3 22.3	< 0.005 0.005
228547 HT	Area	8/10/2004	8/10/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	335	1768.00	0.04475	0	0	22.3 22.3	< 0.005 0.005
228548 HT	Field Blank	8/10/2004	8/10/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	335	0.000	0.08949	0	0	11.2 11.2	N/A N/A
228549 HT	Area	8/10/2004	8/11/2004	99	N/A	(East) - Greenwich	335	2024.00	0.04475	0	0	22.3 22.3	< 0.004 0.004
228550 HT	Area	8/10/2004	8/11/2004	99	Not Acquired	(SouthEast) - Greenwich and Albany	335	1898.75	0.04475	0	0	22.3 22.3	< 0.005 0.005
228558 HT	Area	8/10/2004	8/11/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	335	2032.00	0.04475	0	0	22.3 22.3	< 0.004 0.004
228555 HT	Area	8/10/2004	8/11/2004	99	Not Acquired	(SouthWest) - Washington and Albany	335	2024.00	0.04475	0	0	22.3 22.3	< 0.004 0.004

Notes:

Volumes provided by Ambient were used to calculate analytical results and sensitivities.

"<" indicates results less than analytical sensitivity. "N/A" indicates that sample was not analyzed.

Sample(s) for this project were analyzed at our Monroeville, PA (NYLAP #101208-0, NY ELAP #10844) facility.

1. If RJ Lee Group, Inc. did not collect the samples analyzed, the verifiability of the laboratory's results are limited to the reported values.

Abbreviations: N/A-Not Applicable, Chry-Chrysotile Asbestos, Amph-Amphibole Asbestos, S-Structures.

2. Samples will be held for 90 days and then disposed of per Federal regulations.

These results are submitted pursuant to RJ Lee Group's current terms and conditions of sale, including the company's standard warranty and limitation of liability provisions. No responsibility or liability is assumed for the manner in which these results are used or interpreted.

Aug 17 04 02:25p IBM USER
From: RJ Lee Group, Inc. To: Mr. Noel Drago

732 6018
Date: 8/13/2004 Time: 10:12:04 AM

p. 2

RJ Lee Group, Inc.

Laboratory Report (cont.)

J Lee Group Job No: ATH306366-993
Building: 130 Liberty

RJ Lee Group, Inc.
Report Date: 8/13/2004

RJLG Sample Number	Sample Type	Date Collected		Floor	Grid	Sub Location	Filter Area (mm ²)	Volume (liter)	Area Analyzed (mm ²)	Asbestos Structures Chry - Amph	Concentration	
		Start	Stop								Analytical Sensitivity (S/mm ²)	(S/cc)

Authorized Signature: _____

 /slb
Kimberly A. Allison, Manager - TEM Analysis

Notes:

- Volumes provided by Ambient were used to calculate analytical results and sensitivities.
- "--" indicates results less than analytical sensitivity. "N/A" indicates that sample was not analyzed.
- Samples for this project were analyzed at our: Monroeville, PA (NYLAP #101208-0, NY ELAP #10844) facility.
- If RJ Lee Group, Inc. did not collect the samples analyzed, the verifiability of the laboratory's results are limited to the reported values.
- Abbreviations: N/A-Not Applicable, Chry-Chrysotile Asbestos, Amph-Amphibole Asbestos, S-Structures.
- Samples will be held for 90 days and then disposed of per Federal regulations.
- These results are submitted pursuant to RJ Lee Group's current terms and conditions of sale, including the company's standard warranty and limitation of liability provisions. No responsibility or liability is assumed for the manner in which these results are used or interpreted.