

Jun 24 04 03:18p IBM USER
From: RJ Lee Group, Inc. To: Mr. Noel Drago

732 6018
Date: 6/19/2004 Time: 9:10:22 PM

Page 2 of 4 P. 2

RJ LeeGroup, Inc.

350 Hochberg Road, Monroeville, PA 15146
Phone: (724) 325-1776 Fax: (724) 733-1799

Final Laboratory Report TEM Air Analysis

RJ Lee Group, Inc.
350 Hochberg Road
Monroeville, PA 15146
USA

Report Date: 6/19/2004
Sample Receipt Date: 6/19/2004
RJ Lee Group Job No.: ATH306366-797
Samples Received: 11
Building: 130 Liberty

Method: Federal Register, 40 CFR, Part 763, October 30, 1987

RJLG Sample Number	Sample Type	Date Collected		Floor	Grid	Sub Location	Filter Area (mm ²)	Volume (liter)	Area Analyzed (mm ²)	Asbestos Structures		Concentration Analytical Sensitivity	
		Start	Stop							Chry	Amph	(Sf/mm ²)	(S/cc)
5221551.HT	Area	6/18/2004	6/18/2004	0	Z	(South) - Decon Exit	385	1800.00	0.04315	0	0	< 23.2 23.2	< 0.005 0.005
5221552.HT	Area	6/18/2004	6/18/2004	0	FF	(South) - Shanty Area	385	1804.00	0.04315	0	0	< 23.2 23.2	< 0.005 0.005
5221553.HT	Area	6/18/2004	6/18/2004	0	Y	(South) - Decon Entrance	385	1808.00	0.04315	0	0	< 23.2 23.2	< 0.005 0.005
5221554.HT	Area	6/18/2004	6/18/2004	0	X	(South) - Ramp	385	1812.00	0.04315	0	0	< 23.2 23.2	< 0.005 0.005
5221555.HT	Field Blank	6/18/2004	6/18/2004	0	X	(South) - Ramp	385	0.000	0.08630	0	0	< 11.6 11.6	N/A N/A
5221556.HT	Field Blank	6/18/2004	6/18/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	385	0.000	0.08630	0	0	< 11.6 11.6	N/A N/A
5221557.HT	Area	6/18/2004	6/18/2004	99	Not Acquired	(SouthWest) - Washington and Albany	385	1812.00	0.04315	0	0	< 23.2 23.2	< 0.005 0.005
5221558.HT	Area	6/18/2004	6/18/2004	99	Not Acquired	(SouthEast) - Greenwich and Albany	385	1820.00	0.04315	1	0	23.2 23.2	0.005 0.005
5221559.HT	Area	6/18/2004	6/18/2004	99	N/A	(East) - Greenwich	385	1828.00	0.04315	0	0	< 23.2 23.2	< 0.005 0.005
5221560.HT	Area	6/18/2004	6/18/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	385	1832.00	0.04315	0	0	< 23.2 23.2	< 0.005 0.005

Notes:

1. Volumes provided by Ambient were used to calculate analytical results and sensitivities.
2. "<" indicates results less than analytical sensitivity. "..." indicates that sample was not analyzed.
3. Sample(s) for this project were analyzed at our: Monroeville, PA (NVLAP #101208-0, NY ELAP #10844) facility.
4. If RJ Lee Group, Inc. did not collect the samples analyzed, the verifiability of the laboratory's results are limited to the reported values.
5. Abbreviations: N/A-Not Applicable, Chry-Chrysotile Asbestos, Amph-Amphibole Asbestos, S-Structures.
6. Samples will be held for 90 days and then disposed of per Federal regulations.
7. These results are submitted pursuant to RJ Lee Group's current terms and conditions of sale, including the company's standard warranty and limitation of liability provisions. No responsibility or liability is assumed for the manner in which these results are used or interpreted.

See Appendix A for fiber data.

Page 2 of 4

Jun 24 04 03:19p IBM USER

From: RJ Lee Group, Inc. To: Mr. Noel Drago

732 6018
Date: 6/19/2004 Time: 9:10:22 PM

Page 3 of 4 P. 3

RJ Lee Group, Inc.**Laboratory Report (cont.)**RJ Lee Group Job No: ATH306366-797
Building: 130 LibertyRJ Lee Group, Inc.
Report Date: 6/19/2004

RJLG Sample Number	Sample Type	Date Collected		Floor	Grid	Sub Location	Filter Area (mm ²)	Volume (liter)	Area Analyzed (mm ²)	Asbestos Structures		Concentration	
		Start	Stop							Chry	Amph	Analytical Sensitivity (S/mm ²)	(S/cc)
5221561.HT	Area	6/18/2004	6/18/2004	1	K	(NorthWest) - Cedar Decon	385	1840.00	0.04315	0	0	< 23.2 23.2	< 0.005 0.005

Authorized Signature: _____



/s/b

Kimberly A. Allison, Manager - TEM Analysis

Notes:

1. Volumes provided by Ambient were used to calculate analytical results and sensitivities.
2. "<" indicates results less than analytical sensitivity. "—" indicates that sample was not analyzed.
3. Sample(s) for this project were analyzed at our: Monroeville, PA (NVLAP #101208-0, NY ELAP #10844) facility.
4. If RJ Lee Group, Inc. did not collect the samples analyzed, the verifiability of the laboratory's results are limited to the reported values.
5. Abbreviations: N/A-Not Applicable, Chry-Chrysotile Asbestos, Amph-Amphibole Asbestos, S-Structures.
6. Samples will be held for 90 days and then disposed of per Federal regulations.
7. These results are submitted pursuant to RJ Lee Group's current terms and conditions of sale, including the company's standard warranty and limitation of liability provisions. No responsibility or liability is assumed for the manner in which these results are used or interpreted.

See Appendix A for fiber data.

Page 3 of 4

Jun 24 04 03:19p

IBM USER

732 6018

P. 4

From: RJ Lee Group, Inc. To: Mr. Noel Drago

LEAD: 04/24/04P 11:00:11 AM 04/24/04

RJ LeeGroup, Inc.350 Hochberg Road, Monroeville, PA 15146
Phone: (724) 325-1776 Fax: (724) 733-1799**Final Laboratory Report**
TEM Air AnalysisRJ Lee Group, Inc.
350 Hochberg Road
Monroeville, PA 15146
USAReport Date: 6/22/2004
Sample Receipt Date: 6/21/2004
RJ Lee Group Job No.: ATH306366-799
Samples Received: 0
Building: 130 Liberty

Method: Federal Register, 40 CFR, Part 763, October 30, 1987

RJLG Sample Number	Sample Type	Date Collected Start Stop	Floor	Grid	Sub Location	Filter Area (mm ²)	Volume (liter)	Area Analyzed (mm ²)	Asbestos Structures Chry Amph	Concentration	
										Analytical Sensitivity (S/mm ²)	Analytical Sensitivity (S/cc)
5223769.HT	Area	6/18/2004 6/18/2004	0	Z	(South) - Decon Exit	385	1952.00	0.04315	0 0	< 23.2 23.2	< 0.005 0.005
5223778.HT	Area	6/18/2004 6/18/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	385	1916.00	0.04315	0 0	< 23.2 23.2	< 0.005 0.005
5223771.HT	Area	6/18/2004 6/18/2004	0	Y	(South) - Decon Entrance	385	1826.25	0.04315	0 0	< 23.2 23.2	< 0.005 0.005
5223770.HT	Area	6/18/2004 6/18/2004	0	FF	(South) - Shanty Area	385	1952.00	0.04315	0 0	< 23.2 23.2	< 0.005 0.005
5223773.HT	Field Blank	6/18/2004 6/18/2004	0	X	(South) - Ramp	385	0.000	0.08630	0 0	< 11.6 11.6	N/A N/A
5223772.HT	Area	6/18/2004 6/18/2004	0	X	(South) - Ramp	385	1948.00	0.04315	0 0	< 23.2 23.2	< 0.005 0.005
5223774.HT	Field Blank	6/18/2004 6/18/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	385	0.000	0.08630	0 0	< 11.6 11.6	N/A N/A
5223775.HT	Area	6/18/2004 6/18/2004	99	Not Acquired	(SouthWest) - Washington and Albany	385	1936.00	0.04315	0 0	< 23.2 23.2	< 0.005 0.005
5223776.HT	Area	6/18/2004 6/18/2004	99	Not Acquired	(SouthEast) - Greenwich and Albany	385	1803.75	0.04315	0 0	< 23.2 23.2	< 0.005 0.005
5223777.HT	Area	6/18/2004 6/18/2004	99	N/A	(East) - Greenwich	385	1916.00	0.04315	0 0	< 23.2 23.2	< 0.005 0.005

Notes:

1. Volumes provided by Ambient were used to calculate analytical results and sensitivities.
2. "<" indicates results less than analytical sensitivity. "N/A" indicates that sample was not analyzed.
3. Sample(s) for this project were analyzed at our: Monroeville, PA (NVLAP #101208-0, NY ELAP #10844) facility.
4. If RJ Lee Group, Inc. did not collect the samples analyzed, the verifiability of the laboratory's results are limited to the reported values.
5. Abbreviations: N/A-Not Applicable, Chry-Chrysotile Asbestos, Amph-Amphibole Asbestos, S-Structures.
6. Samples will be held for 90 days and then disposed of per Federal regulations.
7. These results are submitted pursuant to RJ Lee Group's current terms and conditions of sale, including the company's standard warranty and limitation of liability provisions. No responsibility or liability is assumed for the manner in which these results are used or interpreted.

See Appendix A for fiber data.

Page 2 of 5

Jun 24 04 03:19p

IBM USER

732 6018

P.5

From: RJ Lee Group, Inc. To: Mr. Noel Drago

L000: 07/24/2004 11:00:11 AM

RJ Lee Group, Inc.**Laboratory Report (cont.)**RJ Lee Group Job No: ATH306366-799
Building: 130 LibertyRJ Lee Group, Inc.
Report Date: 6/22/2004

RJLG Sample Number	Sample Type	Date Collected		Floor	Grid	Sub Location	Filter Area (mm ²)	Volume (liter)	Area Analyzed (mm ²)	Asbestos Structures		Concentration Analytical Sensitivity	
		Start	Stop							Chry	Amph	(8mm ²)	(S/cc)
5223779.HT	Area	6/18/2004	6/18/2004	1	K	(NorthWest) - Cedar Decon	385	1904.00	0.04315	0	0	< 23.2 23.2	< 0.005 0.005
5223169.HT	Area	6/18/2004	6/19/2004	0	Z	(South) - Decon Exit	385	1936.00	0.04315	0	0	< 23.2 23.2	< 0.005 0.005
5223170.HT	Area	6/18/2004	6/19/2004	0	FF	(South) - Shanty Area	385	1932.00	0.04315	0	0	< 23.2 23.2	< 0.005 0.005
5223171.HT	Area	6/18/2004	6/19/2004	0	Y	(South) - Decon Entrance	385	1932.00	0.04315	0	0	< 23.2 23.2	< 0.005 0.005
5223172.HT	Area	6/18/2004	6/19/2004	0	X	(South) - Ramp	385	1928.00	0.04315	0	0	< 23.2 23.2	< 0.005 0.005
5223173.HT	Field Blank	6/18/2004	6/18/2004	0	X	(South) - Ramp	385	0.000	0.08630	0	0	< 11.6 11.6	N/A N/A
5223174.HT	Field Blank	6/18/2004	6/18/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	385	0.000	0.08630	0	0	< 11.6 11.6	N/A N/A
5223175.HT	Area	6/18/2004	6/19/2004	99	Not Acquired	(SouthWest) - Washington and Albany	385	1928.00	0.04315	0	0	< 23.2 23.2	< 0.005 0.005
5223177.HT	Area	6/18/2004	6/19/2004	99	N/A	(East) - Greenwich	385	1932.00	0.04315	0	0	< 23.2 23.2	< 0.005 0.005
5223178.HT	Area	6/18/2004	6/19/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	385	1932.00	0.04315	0	0	< 23.2 23.2	< 0.005 0.005
5223176.HT	Area	6/18/2004	6/19/2004	99	Not Acquired	(SouthEast) - Greenwich and Albany	385	2052.75	0.04315	0	0	< 23.2 23.2	< 0.004 0.004
5223179.HT	Area	6/18/2004	6/19/2004	1	K	(NorthWest) - Cedar Decon	385	1936.00	0.04315	0	0	< 23.2 23.2	< 0.005 0.005

Notes:

1. Volumes provided by Ambient were used to calculate analytical results and sensitivities.
2. "<" indicates results less than analytical sensitivity. "--" indicates that sample was not analyzed.
3. Sample(s) for this project were analyzed at our: Monroeville, PA (NVLAP #101208-0, NY ELAP #10844) facility.
4. If RJ Lee Group, Inc. did not collect the samples analyzed, the verifiability of the laboratory's results are limited to the reported values.
5. Abbreviations: N/A-Not Applicable, Chry-Chrysotile Asbestos, Amph-Amphibole Asbestos, S-Structures.
6. Samples will be held for 90 days and then disposed of per Federal regulations.
7. These results are submitted pursuant to RJ Lee Group's current terms and conditions of sale, including the company's standard warranty and limitation of liability provisions. No responsibility or liability is assumed for the manner in which these results are used or interpreted.

See Appendix A for fiber data.

Page 3 of 5

Jun 24 04 03:20p IBM USER
From: RJ Lee Group, Inc. To: Mr. Noel Drago

732 6018
USER: 07/24/2004 11:08:11 AM

P. 6

RJ Lee Group, Inc.

Laboratory Report (cont.)

RJ Lee Group Job No: ATH306366-799
Building: 130 Liberty

RJ Lee Group, Inc.
Report Date:

RJLG Sample Number	Sample Type	Date Collected		Floor	Grid	Sub Location	Filter Area (mm ²)	Volume (liter)	Area Analyzed (mm ²)	Asbestos Structures		Concentration Analytical Sensitivity	
		Start	Stop							Chry	Amph	(S/m ²)	(S/cc)
5223176.HT	Area	6/18/2004	6/19/2004	99	Not Acquired	(SouthEast) - Greenwich and Albany	385	2052.75	0.04398	0	0	< 22.7 22.7	< 0.004 0.004

Authorized Signature: _____

 /s/b
Kimberly A. Allison, Manager - TEM Analysis

Notes:

1. Volumes provided by Ambient were used to calculate analytical results and sensitivities.
2. "<" indicates results less than analytical sensitivity. "---" indicates that sample was not analyzed.
3. Sample(s) for this project were analyzed at our: Monroeville, PA (NVLAP #101208-0, NY ELAP #10344) facility.
4. If RJ Lee Group, Inc. did not collect the samples analyzed, the verifiability of the laboratory's results are limited to the reported values.
5. Abbreviations: N/A-Not Applicable, Chry-Chrysotile Asbestos, Amph-Amphibole Asbestos, S-Structures.
6. Samples will be held for 90 days and then disposed of per Federal regulations.
7. These results are submitted pursuant to RJ Lee Group's current terms and conditions of sale, including the company's standard warranty and limitation of liability provisions. No responsibility or liability is assumed for the manner in which these results are used or interpreted.

See Appendix A for fiber data.

Page 4 of 5