

Jul 02 04 02:21p

IBM USER

732 6018

P. 2

6/20

RJ LeeGroup, Inc.

350 Hochberg Road, Monroeville, PA 15146

Phone: (724) 325-1776 Fax: (724) 733-1799

Final Laboratory Report
TEM Air AnalysisRJ Lee Group, Inc.
350 Hochberg Road
Monroeville, PA 15146
USAReport Date: 6/23/2004
Sample Receipt Date: 6/22/2004
RJ Lee Group Job No.: ATH306366-802
Samples Received: 23
Building: 130 Liberty

Method: Federal Register, 40 CFR, Part 763, October 30, 1987

RJLG Sample Number	Sample Type	Date Collected		Floor	Grid	Sub Location	Filter Area (mm ²)	Volume (liter)	Area Analyzed (mm ²)	Asbestos Structures		Concentration	
		Start	Stop							Chry	Amph	Analytical Sensitivity (S/mm ²)	(Stc)
5223810.HT	Area	6/20/2004	6/20/2004	0	Z	(South) - Decon Exit	385	1924.00	0.04317	0	0	< 23.2 23.2	< 0.005 0.005
5223811.HT	Area	6/20/2004	6/20/2004	0	FF	(South) - Shanty Area	385	1916.00	0.04317	0	0	< 23.2 23.2	< 0.005 0.005
5223812.HT	Area	6/20/2004	6/20/2004	0	Y	(South) - Decon Entrance	385	1940.00	0.04317	0	0	< 23.2 23.2	< 0.005 0.005
5223813.HT	Area	6/20/2004	6/20/2004	0	X	(South) - Ramp	385	1932.00	0.04317	0	0	< 23.2 23.2	< 0.005 0.005
5223814.HT	Field Blank	6/20/2004	6/20/2004	0	X	(South) - Ramp	385	0.000	0.08633	0	0	< 11.6 11.6	N/A N/A
5223815.HT	Field Blank	6/20/2004	6/20/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	385	0.000	0.08633	0	0	< 11.6 11.6	N/A N/A
5223816.HT	Area	6/20/2004	6/20/2004	99	Not Acquired	(SouthWest) - Washington and Albany	385	1936.00	0.04317	0	0	< 23.2 23.2	< 0.005 0.005
5223817.HT	Area	6/20/2004	6/20/2004	99	Not Acquired	(SouthEast) - Greenwich and Albany	385	1948.00	0.04317	0	0	< 23.2 23.2	< 0.005 0.005
5223818.HT	Area	6/20/2004	6/20/2004	99	N/A	(East) - Greenwich	385	1964.00	0.04317	0	0	< 23.2 23.2	< 0.005 0.005
5223819.HT	Area	6/20/2004	6/20/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	385	1980.00	0.04317	0	0	< 23.2 23.2	< 0.005 0.005

Notes:

1. Volumes provided by Ambient were used to calculate analytical results and sensitivities.
2. "<" indicates results less than analytical sensitivity. "N/A" indicates that sample was not analyzed.
3. Sample(s) for this project were analyzed at our: Monroeville, PA (NVLAP #101208-0, NY ELAP #10844) facility.
4. If RJ Lee Group, Inc. did not collect the samples analyzed, the verifiability of the laboratory's results are limited to the reported values.
5. Abbreviations: N/A-Not Applicable, Chry-Chrysotile Asbestos, Amph-Amphibole Asbestos, S-Structures.
6. Samples will be held for 90 days and then disposed of per Federal regulations.
7. These results are submitted pursuant to RJ Lee Group's current terms and conditions of sale, including the company's standard warranty and limitation of liability provisions. No responsibility or liability is assumed for the manner in which these results are used or interpreted.

See Appendix A for fiber data.

Page 2 of 5

RJ Lee Group, Inc.**Laboratory Report (cont.)**RJ Lee Group Job No: ATH306366-802
Building: 130 LibertyRJ Lee Group, Inc.
Report Date: 6/23/2004

RJLG Sample Number	Sample Type	Date Collected Start Stop	Floor	Grid	Sub Location	Filter Area (mm ²)	Volume (liter)	Area Analyzed (mm ²)	Asbestos Structures		Concentration	
									Chry	Amph	S/mmf	S/Sec
5223820.HT	Area	6/20/2004 6/20/2004	1	K	(NorthWest) - Cedar Decon	385	1984.00	0.04317	0	0	< 23.2 23.2	< 0.004 0.004
5223821.HT	Area	6/20/2004 6/20/2004	0	Z	(South) - Decon Exit	385	1924.00	0.04317	0	0	< 23.2 23.2	< 0.005 0.005
5223822.HT	Area	6/20/2004 6/20/2004	0	FF	(South) - Shanty Area	385	1924.00	0.04317	0	0	< 23.2 23.2	< 0.005 0.005
5223823.HT	Area	6/20/2004 6/20/2004	0	Y	(South) - Decon Entrance	385	1916.00	0.04317	0	0	< 23.2 23.2	< 0.005 0.005
5223824.HT	Area	6/20/2004 6/20/2004	0	X	(South) - Ramp	385	1912.00	0.04317	0	0	< 23.2 23.2	< 0.005 0.005
5223824.HT	Area	6/20/2004 6/20/2004	0	X	(South) - Ramp	385	1912.00	0.04261	0	0	< 23.5 23.5	< 0.005 0.005
5223825.HT	Field Blank	6/20/2004 6/20/2004	0	X	(South) - Ramp	385	0.000	0.08633	0	0	< 11.6 11.6	N/A N/A
5223826.HT	Field Blank	6/20/2004 6/20/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	385	0.000	0.08633	0	0	< 11.6 11.6	N/A N/A
5223827.HT	Area	6/20/2004 6/20/2004	99	Not Acquired	(SouthWest) - Washington and Albany	385	1912.00	0.04317	0	0	< 23.2 23.2	< 0.005 0.005
5223828.HT	Area	6/20/2004 6/20/2004	99	Not Acquired	(SouthEast) - Greenwich and Albany	385	1900.00	0.04317	0	0	< 23.2 23.2	< 0.005 0.005
5223829.HT	Area	6/20/2004 6/20/2004	99	N/A	(East) - Greenwich	385	1880.00	0.04317	0	0	< 23.2 23.2	< 0.005 0.005
5223830.HT	Area	6/20/2004 6/20/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	385	1868.00	0.04317	0	0	< 23.2 23.2	< 0.005 0.005

Notes:

- Volumes provided by Ambient were used to calculate analytical results and sensitivities.
- "<" indicates results less than analytical sensitivity. "—" indicates that sample was not analyzed.
- Sample(s) for this project were analyzed at our: Monroeville, PA (NVLAP #101208-0, NY ELAP #10844) facility.
- If RJ Lee Group, Inc. did not collect the samples analyzed, the verifiability of the laboratory's results are limited to the reported values.
- Abbreviations: N/A-Not Applicable, Chry-Chrysotile Asbestos, Amph-Amphibole Asbestos, S-Structures
- Samples will be held for 90 days and then disposed of per Federal regulations.
- These results are submitted pursuant to RJ Lee Group's current terms and conditions of sale, including the company's standard warranty and limitation of liability provisions. No responsibility or liability is assumed for the manner in which these results are used or interpreted.

See Appendix A for fiber data.

Page 3 of 5

Jul 02 04 02:21p


IBM USER

732 6018

P. 4

RJLG Sample Number	Sample Type	Collected Start Stop	Floor	Grid	Sub Location	Area (mm ²)	Volume (liter)	Analyzed (mm ²)	Structures Chry Amph		Analytical Sensitivity (S/mm ²) (S/cc)	
5223831 HT	Area	6/20/2004 6/20/2004	1	K	(NorthWest) - Cedar Decon	385	1872.00	0.04317	0	0	< 23.2 23.2	< 0.005 0.005

Authorized Signature:



/sib

Kimberly A. Allison, Manager - TEM Analysis

Notes:

1. Volumes provided by Ambient were used to calculate analytical results and sensitivities.
2. "<" indicates results less than analytical sensitivity. "-" indicates that sample was not analyzed.
3. Sample(s) for this project were analyzed at our: Monroeville, PA (NVLAP #101208-0, NY ELAP #10844) facility.
4. If RJ Lee Group, Inc. did not collect the samples analyzed, the verifiability of the laboratory's results are limited to the reported values.
5. Abbreviations: N/A-Not Applicable, Chry-Chrysotile Asbestos, Amph-Amphibole Asbestos, S-Structures.
6. Samples will be held for 90 days and then disposed of per Federal regulations.
7. These results are submitted pursuant to RJ Lee Group's current terms and conditions of sale, including the company's standard warranty and limitation of liability provisions. No responsibility or liability is assumed for the manner in which these results are used or interpreted.

See Appendix A for fiber data.

Page 4 of 5

Jul 02 04 02:21p

IBM USER

732 6018

p. 5

From: RJ Lee Group, Inc. To: Mr. Noel Drago

Date: 07/02/2004 Time: 0:00:00 PM

RJ LeeGroup, Inc.350 Hochberg Road, Monroeville, PA 15146
Phone: (724) 325-1776 Fax: (724) 733-1799**Final Laboratory Report**
TEM Air AnalysisRJ Lee Group, Inc.
350 Hochberg Road
Monroeville, PA 15146
USAReport Date: 6/23/2004
Sample Receipt Date: 6/22/2004
RJ Lee Group Job No.: ATH306366-800
Samples Received: 11
Building: 130 Liberty

Method: Federal Register, 40 CFR, Part 763, October 30, 1987

RJLG Sample Number	Sample Type	Date Collected Start Stop	Floor	Grid	Sub Location	Filter Area (mm ²)	Volume (liter)	Area Analyzed (mm ²)	Asbestos Structures Chry Amph	Concentration	
										Analytical Sensitivity (S/mn ²)	(S/cc)
5223842.HT	Area	6/20/2004 6/21/2004	0	Z	(South) - Decon Exit	385	1904.00	0.04317	0 0	< 23.2 23.2	< 0.005 0.005
5223843.HT	Area	6/20/2004 6/21/2004	0	FF	(South) - Shanty Area	385	1904.00	0.04317	0 0	< 23.2 23.2	< 0.005 0.005
5223844.HT	Area	6/20/2004 6/21/2004	0	Y	(South) - Decon Entrance	385	1904.00	0.04317	0 0	< 23.2 23.2	< 0.005 0.005
5223845.HT	Area	6/20/2004 6/21/2004	0	X	(South) - Ramp	385	1785.00	0.04317	0 0	< 23.2 23.2	< 0.005 0.005
5223846.HT	Field Blank	6/20/2004 6/20/2004	0	X	(South) - Ramp	385	0.000	0.08633	0 0	< 11.6 11.6	N/A N/A
5223847.HT	Field Blank	6/20/2004 6/20/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	385	0.000	0.08633	0 0	< 11.6 11.6	N/A N/A
5223848.HT	Area	6/20/2004 6/21/2004	99	Not Acquired	(SouthWest) - Washington and Albany	385	1781.25	0.05180	0 0	< 19.3 19.3	< 0.004 0.004
5223849.HT	Area	6/20/2004 6/21/2004	99	Not Acquired	(SouthEast) - Greenwich and Albany	385	1900.00	0.04317	0 0	< 23.2 23.2	< 0.005 0.005
5223851.HT	Area	6/20/2004 6/21/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	385	1900.00	0.04317	0 0	< 23.2 23.2	< 0.005 0.005
5223852.HT	Area	6/20/2004 6/21/2004	1	K	(NorthWest) - Cedar Decon	385	1888.00	0.04317	0 0	< 23.2 23.2	< 0.005 0.005

Notes:

1. Volumes provided by Ambient were used to calculate analytical results and sensitivities.
2. "<" indicates results less than analytical sensitivity. "N/A" indicates that sample was not analyzed.
3. Sample(s) for this project were analyzed at our Monroeville, PA (NVLAP #101208-0, NY ELAP #10844) facility.
4. If RJ Lee Group, Inc. did not collect the samples analyzed, the verifiability of the laboratory's results are limited to the reported values.
5. Abbreviations: N/A-Not Applicable, Chry-Chrysotile Asbestos, Amph-Amphibole Asbestos, S-Structures.
6. Samples will be held for 90 days and then disposed of per Federal regulations.
7. These results are submitted pursuant to RJ Lee Group's current terms and conditions of sale, including the company's standard warranty and limitation of liability provisions. No responsibility or liability is assumed for the manner in which these results are used or interpreted.

See Appendix A for fiber data.

Page 2 of 4

Jul 02 04 02:22p

IBM USER

732 6018

p.6

RJ Lee Group, Inc.**Laboratory Report (cont.)**

RJ Lee Group Job No: ATH306366-800
 Building: 130 Liberty

RJ Lee Group, Inc.
 Report Date: 6/23/2004

RJLG Sample Number	Sample Type	Date Collected		Floor	Grid	Sub Location	Filter Area (mm ²)	Volume (liter)	Area Analyzed (mm ²)	Asbestos Structures		Concentration	
		Start	Stop							Chry	Amph	Sensitivity (S/mm ²)	Analytical Sensitivity (S/cc)
5223850.HT	Area	6/20/2004	6/21/2004	99	N/A	(East) - Greenwich	385	1904.00	0.04317	0	0	< 23.2 23.2	< 0.005 0.005

Authorized Signature: _____



/s/b

Kimberly A. Allison, Manager - TEM Analysis

Notes:

1. Volumes provided by Ambient were used to calculate analytical results and sensitivities.
2. "<" indicates results less than analytical sensitivity. "----" indicates that sample was not analyzed.
3. Sample(s) for this project were analyzed at our: Monroeville, PA (NVLAP #101208-0, NY ELAP #10644) facility.
4. If RJ Lee Group, Inc. did not collect the samples analyzed, the verifiability of the laboratory's results are limited to the reported values.
5. Abbreviations: N/A-Not Applicable, Chry-Chrysotile Asbestos, Amph-Amphibole Asbestos, S-Structures.
6. Samples will be held for 90 days and then disposed of per Federal regulations.
7. These results are submitted pursuant to RJ Lee Group's current terms and conditions of sale, including the company's standard warranty and limitation of liability provisions. No responsibility or liability is assumed for the manner in which these results are used or interpreted.

See Appendix A for fiber data.

Page 3 of 4