

Jul 13 04 02:16p

IBM USER

732 6018

P.5

1/1

RJ LeeGroup, Inc.

350 Hochberg Road, Monroeville, PA 15146
Phone: (724) 325-1776 Fax: (724) 733-1799

Final Laboratory Report TEM Air Analysis

RJ Lee Group, Inc.
350 Hochberg Road
Monroeville, PA 15146
USA

Report Date: 7/9/2004
Sample Receipt Date: 7/8/2004
RJ Lee Group Job No.: ATH306366-856
Samples Received: 5
Building: 130 Liberty

Method: Federal Register, 40 CFR, Part 763, October 30, 1987

RJLG Sample Number	Sample Type	Date Collected Start Stop	Floor	Grid	Sub Location	Filter Area (mm ²)	Volume (liter)	Area Analyzed (mm ²)	Asbestos Structures Chry Amph	Concentration	
										Analytical Sensitivity (S/mm ²)	(S/cc)
5225674.HT	Field Blank	7/7/2004 7/7/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	385	0.000	0.08730	0 0	< 11.5 11.5	N/A N/A
5225675.HT	Area	7/7/2004 7/7/2004	99	Not Acquired	(SouthWest) - Washington and Albany	385	1792.00	0.04365	0 0	< 22.9 22.9	< 0.005 0.005
5225676.HT	Area	7/7/2004 7/7/2004	99	Not Acquired	(SouthEast) - Greenwich and Albany	385	1796.00	0.04365	0 0	< 22.9 22.9	< 0.005 0.005
5225677.HT	Area	7/7/2004 7/7/2004	99	N/A	(East) - Greenwich	385	1816.00	0.04365	0 0	< 22.9 22.9	< 0.005 0.005
5225678.HT	Area	7/7/2004 7/7/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	385	1812.00	0.04365	0 0	< 22.9 22.9	< 0.005 0.005

Authorized Signature: _____



/s/lb

Kimberly A. Allison, Manager - TEM Analysis

Notes:

- Volumes provided by Ambient were used to calculate analytical results and sensitivities.
- "<" indicates results less than analytical sensitivity. "..." indicates that sample was not analyzed.
- Sample(s) for this project were analyzed at our: Monroeville, PA (NVLAP #101208-0, NY ELAP #10844) facility.
- If RJ Lee Group, Inc. did not collect the samples analyzed, the verifiability of the laboratory's results are limited to the reported values.
- Abbreviations: N/A-Not Applicable, Chry-Chrysotile Asbestos, Amph-Amphibole Asbestos, S-Structures.
- Samples will be held for 90 days and then disposed of per Federal regulations.
- These results are submitted pursuant to RJ Lee Group's current terms and conditions of sale, including the company's standard warranty and limitation of liability provisions. No responsibility or liability is assumed for the manner in which these results are used or interpreted.

See Appendix A for fiber data.

Page 2 of 3

Jul 13 04 02:17p

IBM USER

732 6018

P. 6

RJ LeeGroup, Inc.

350 Hochberg Road, Monroeville, PA 15146
Phone: (724) 325-1776 Fax: (724) 733-1799

Final Laboratory Report TEM Air Analysis

RJ Lee Group, Inc.
350 Hochberg Road
Monroeville, PA 15146
USA

Report Date: 7/10/2004
Sample Receipt Date: 7/9/2004
RJ Lee Group Job No.: ATH306366-862
Samples Received: 0
Building: 130 Liberty

Method: Federal Register, 40 CFR, Part 763, October 30, 1987

RJLG Sample Number	Sample Type	Date Collected Start Stop	Floor	Grid	Sub Location	Filter Area (mm ²)	Volume (liter)	Area Analyzed (mm ²)	Asbestos Structures Chry Amph	Concentration	
										Analytical Sensitivity (S/mm ²)	Analytical Sensitivity (S/cc)
5225685.HT	Field Blank	7/7/2004 7/7/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	385	0.000	0.08773	0 0	< 11.4 11.4	N/A N/A
5225686.HT	Area	7/7/2004 7/7/2004	99	Not Acquired	(SouthWest) - Washington and Albany	385	2048.00	0.04386	0 0	< 22.8 22.8	< 0.004 0.004
5225687.HT	Area	7/7/2004 7/7/2004	99	Not Acquired	(SouthEast) - Greenwich and Albany	385	2040.00	0.04386	0 0	< 22.8 22.8	< 0.004 0.004
5225688.HT	Area	7/7/2004 7/7/2004	99	N/A	(East) - Greenwich	385	2024.00	0.04386	0 0	< 22.8 22.8	< 0.004 0.004
5225689.HT	Area	7/7/2004 7/7/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	385	2024.00	0.04386	0 0	< 22.8 22.8	< 0.004 0.004
5225726.HT	Field Blank	7/7/2004 7/7/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	385	0.000	0.08773	0 0	< 11.4 11.4	N/A N/A
5225727.HT	Area	7/7/2004 7/8/2004	99	Not Acquired	(SouthWest) - Washington and Albany	385	1892.00	0.04386	0 0	< 22.8 22.8	< 0.005 0.005
5225728.HT	Area	7/7/2004 7/8/2004	99	Not Acquired	(SouthEast) - Greenwich and Albany	385	1896.00	0.04386	0 0	< 22.8 22.8	< 0.005 0.005
5225729.HT	Area	7/7/2004 7/8/2004	99	N/A	(East) - Greenwich	385	1892.00	0.04386	0 0	< 22.8 22.8	< 0.005 0.005
5225730.HT	Area	7/7/2004 7/8/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	385	1884.00	0.04386	0 0	< 22.8 22.8	< 0.005 0.005

Notes:

1. Volumes provided by Ambientx were used to calculate analytical results and sensitivities.
2. "<" indicates results less than analytical sensitivity. "N/A" indicates that sample was not analyzed.
3. Sample(s) for this project were analyzed at our Monroeville, PA (NVLAP #101208-0, NY ELAP #10844) facility.
4. If RJ Lee Group, Inc. did not collect the samples analyzed, the verifiability of the laboratory's results are limited to the reported values.
5. Abbreviations: N/A-Not Applicable, Chry-Chrysotile Asbestos, Amph-Amphibole Asbestos, S-Structures.
6. Samples will be held for 90 days and then disposed of per Federal regulations.
7. These results are submitted pursuant to RJ Lee Group's current terms and conditions of sale, including the company's standard warranty and limitation of liability provisions. No responsibility or liability is assumed for the manner in which these results are used or interpreted.

See Appendix A for fiber data.

Page 2 of 4

Jul 13 04 02:18p

IBM USER

732 6018

Page 7 of 7

RJ Lee Group, Inc.**Laboratory Report (cont.)**RJ Lee Group Job No: ATH306366-862
Building: 130 LibertyRJ Lee Group, Inc.
Report Date 7/10/2004

RJLG Sample Number	Sample Type	Date Collected		Floor	Grid	Sub Location	Filter Area (mm ²)	Volume (liter)	Area Analyzed (mm ²)	Asbestos Structures		Concentration	
		Start	Stop							Chry	Amph	Sensitivity (S/mm ²)	Sensitivity (S/cc)

Authorized Signature _____ /sib

Notes:

1. Volumes provided by Ambient were used to calculate analytical results and sensitivities.
2. "<" indicates results less than analytical sensitivity. "—" indicates that sample was not analyzed.
3. Sample(s) for this project were analyzed at our: Monroeville, PA (NVLAP #101208-0, NY ELAP #10844) facility.
4. If RJ Lee Group, Inc. did not collect the samples analyzed, the verifiability of the laboratory's results are limited to the reported values.
5. Abbreviations: N/A-Not Applicable, Chry-Chrysotile Asbestos, Amph-Amphibole Asbestos, S-Structures.
6. Samples will be held for 90 days and then disposed of per Federal regulations.
7. These results are submitted pursuant to RJ Lee Group's current terms and conditions of sale, including the company's standard warranty and limitation of liability provisions. No responsibility or liability is assumed for the manner in which these results are used or interpreted.

See Appendix A for fiber data.

Page 3 of 4