

Jul 27 04 11:55a

IBM USER

732 6018

P. 5

RJ LeeGroup, Inc.

350 Hochberg Road, Monroeville, PA 15146
Phone: (724) 325-1776 Fax: (724) 733-1799

Final Laboratory Report

TEM Air Analysis

RJ Lee Group, Inc.
350 Hochberg Road
Monroeville, PA 15146
USA

Report Date: 7/23/2004
Sample Receipt Date: 7/23/2004
RJ Lee Group Job No.: ATH306366-911
Samples Received: 10
Building: 130 Liberty

Method: Federal Register, 40 CFR, Part 763, October 30, 1987

RJLG Sample Number	Sample Type	Date Collected Start Stop	Floor	Grid	Sub Location	Filter Area (mm ²)	Volume (liter)	Area Analyzed (mm ²)	Asbestos Structures Chry Amph	Concentration	
										Analytical Sensitivity (S/mm ²)	(S/cc)
5225937.HT	Field Blank	7/21/2004 7/21/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	385	0.000	0.08772	0 0	< 11.4 11.4	N/A N/A
5225938.HT	Area	7/21/2004 7/21/2004	99	Not Acquired	(SouthWest) - Washington and Albany	385	2004.00	0.05263	0 0	< 19.0 19.0	< 0.004 0.004
5225939.HT	Area	7/21/2004 7/21/2004	99	Not Acquired	(SouthEast) - Greenwich and Albany	385	1980.00	0.04386	0 0	< 22.8 22.8	< 0.004 0.004
5225940.HT	Area	7/21/2004 7/21/2004	99	N/A	(East) - Greenwich	385	1964.00	0.04386	0 0	< 22.8 22.8	< 0.004 0.004
5225941.HT	Area	7/21/2004 7/21/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	385	1964.00	0.04386	0 0	< 22.8 22.8	< 0.004 0.004
5227774.HT	Field Blank	7/21/2004 7/21/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	385	0.000	0.08772	0 0	< 11.4 11.4	N/A N/A
5227775.HT	Area	7/21/2004 7/22/2004	99	Not Acquired	(SouthWest) - Washington and Albany	385	1830.00	0.04386	0 0	< 22.8 22.8	< 0.005 0.005
5227777.HT	Area	7/21/2004 7/22/2004	99	N/A	(East) - Greenwich	385	1960.00	0.04386	0 0	-----	< 0.002 0.002
5227776.HT	Area	7/21/2004 7/22/2004	99	Not Acquired	(SouthEast) - Greenwich and Albany	385	1960.00	0.04386	0 0	< 22.8 22.8	< 0.004 0.004
5227778.HT	Area	7/21/2004 7/22/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	385	1952.00	0.04386	0 0	< 22.8 22.8	< 0.004 0.004

Notes:

1. Volumes provided by Ambient were used to calculate analytical results and sensitivities.
2. "<" indicates results less than analytical sensitivity. "-----" indicates that sample was not analyzed.
3. Sample(s) for this project were analyzed at our: Monroeville, PA (NVLAP #101208-0, NY ELAP #10844) facility.
4. If RJ Lee Group, Inc. did not collect the samples analyzed, the verifiability of the laboratory's results are limited to the reported values.
5. Abbreviations: N/A-Not Applicable, Chry-Chrysotile Asbestos, Amph-Amphibole Asbestos, S-Structures.
6. Samples will be held for 90 days and then disposed of per Federal regulations.
7. These results are submitted pursuant to RJ Lee Group's current terms and conditions of sale, including the company's standard warranty and limitation of liability provisions. No responsibility or liability is assumed for the manner in which these results are used or interpreted.

See Appendix A for fiber data.

Page 2 of 4

Jul 27 04 11:56a IBM USER

Date: 7/27/2004 Time: 9:23:10 AM 732 6018

Page 3 of 4 P. 6

RJ Lee Group, Inc.**Laboratory Report (cont.)**RJ Lee Group Job No. ATH306366-911
Building: 130 LibertyRJ Lee Group, Inc.
Report Date:

RJLG Sample Number	Sample Type	Date Collected		Floor	Grid	Sub Location	Filter Area (mm ²)	Volume (liter)	Area Analyzed (mm ²)	Asbestos Structures		Concentration	
		Start	Stop							Chry	Amph	Analytical Sensitivity (S/mm ²)	(S/cc)

Authorized Signature:



/s/b

Kimberly A. Allison, Manager - TEM Analysis

Notes:

1. Volumes provided by Ambient were used to calculate analytical results and sensitivities.
2. "<" indicates results less than analytical sensitivity. "-" indicates that sample was not analyzed.
3. Sample(s) for this project were analyzed at our: Monroeville, PA (NVLAP #101208-0, NY ELAP #10844) facility.
4. If RJ Lee Group, Inc. did not collect the samples analyzed, the verifiability of the laboratory's results are limited to the reported values.
5. Abbreviations: N/A-Not Applicable, Chry-Chrysotile Asbestos, Amph-Amphibole Asbestos, S-Structures.
6. Samples will be held for 90 days and then disposed of per Federal regulations.
7. These results are submitted pursuant to RJ Lee Group's current terms and conditions of sale, including the company's standard warranty and limitation of liability provisions. No responsibility or liability is assumed for the manner in which these results are used or interpreted.

See Appendix A for fiber data.


Page 3 of 4

RJ LeeGroup, Inc.350 Hochberg Road, Monroeville, PA 15146
Phone: (724) 325-1776 Fax: (724) 733-1799**Final Laboratory Report**
TEM Air AnalysisRJ Lee Group, Inc.
350 Hochberg Road
Monroeville, PA 15146
USAReport Date: 7/23/2004
Sample Receipt Date: 7/22/2004
RJ Lee Group Job No.: ATH306366-909
Samples Received: 0
Building: 130 Liberty

Method: Federal Register, 40 CFR, Part 763, October 30, 1987

RJLG Sample Number	Sample Type	Date Collected Start Stop	Floor	Grid	Sub Location	Filter Area (mm ²)	Volume (liter)	Area Analyzed (mm ²)	Asbestos Structures Chry Amph	Concentration	
										Analytical Sensitivity (S/mm ²)	Analytical Sensitivity (S/cc)
5227759.HT	Area	7/21/2004 7/21/2004	99	Not Acquired	(SouthWest) - Washington and Albany	385	1796.00	0.04386	0 0	< 22.8 22.8	< 0.005 0.005
5227758.HT	Field Blank	7/21/2004 7/21/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	385	0.000	0.08772	0 0	< 11.4 11.4	N/A N/A
5227762.HT	Area	7/21/2004 7/21/2004	99	Not Acquired	(NorthWest) - Washington and Cedar	385	1840.00	0.04386	0 0	< 22.8 22.8	< 0.005 0.005

Authorized Signature:



/s/lb

Kimberly A. Allison, Manager - TEM Analysis

Notes:

1. Volumes provided by Ambient were used to calculate analytical results and sensitivities.
2. "<" indicates results less than analytical sensitivity. "—" indicates that sample was not analyzed.
3. Sample(s) for this project were analyzed at our: Monroeville, PA (NVLAP #101208-0, NY ELAP #10844) facility.
4. If RJ Lee Group, Inc. did not collect the samples analyzed, the verifiability of the laboratory's results are limited to the reported values.
5. Abbreviations: N/A-Not Applicable, Chry-Chrysotile Asbestos, Amph-Amphibole Asbestos, S-Structures.
6. Samples will be held for 90 days and then disposed of per Federal regulations.
7. These results are submitted pursuant to RJ Lee Group's current terms and conditions of sale, including the company's standard warranty and limitation of liability provisions. No responsibility or liability is assumed for the manner in which these results are used or interpreted.

See Appendix A for fiber data.

Page 2 of 3