

ASBESTOS STANDARDS

- POLARIZED LIGHT MICROSCOPY (PLM)
 - For bulk samples & the ^{def.}std is greater than 1%
- PHASE CONTRAST MICROSCOPY (PCM)
 - For air samples & the re-occupancy ^{std} is 0.01 fibers/cc
- Transmission Electron Microscopy (TEM)
 - For air samples & the re-occupancy ^astd. Is 70 structures/ square millimeter

There is no indoor or outdoor std for asbestos.