NYC 9/11 Public Portal Document

EMSL Analytical, Inc.
307 West 38th Street, New York, NY 10018
Phone: (212) 290-0051 Fax: (212) 290-0058 Email: manhattanlab@emsl.com

Attn: TRC Environmental Consultants
1430 Broadway
10th Floor
New York, NY 10018
Fax: (212) 221-7840 Phone: 212-221-7822
Project: 35052-0000-00017 / New York City Law Department / 100 Church St.

Customer ID: TRCE51
Customer PO:
Received: 03/03/03 6:38 PM
EMSL Order: 030303330
EMSL Project ID:
Analysis Date: 3/4/2003

Phase Contrast Microscopy (PCM) Fiber Count by NIOSH 7400, Fourth Edition, Issue 2, 8/15/94

Sample	Location	Sample Date	Volume	Fibers	Fields	LOD (fib/cc)	Fibers/ mm²	Fibers/ cc	Notes
ZD030303ASBAIR2001 *030303330-0001*	20th Floor - Entrance to North Corridor	3/3/2003	1527.02	25.0	100	0.002	31.85	0.008	
ZD030303ASBAIR2002 *030303330-0002*	20th Floor - Entrance to South Corridor	3/3/2003	1533.33	20.0	100	0.002	25.48	0.006	
ZD030303ASBAIR2003 *030303330-0003*	20th Floor - Entrance to North Corridor	3/3/2003	1527.02	<5.5	100	0.002	<7.0	<0.002	
ZD030303ASBAIR2004 *030303330-0004*	20th Floor - Entrance to South Corridor	3/3/2003	1520.71	<5.5	100	0.002	<7.0	<0.002	
ZD030303ASBAIRFB05 *030303330-0005*	Blank	3/3/2003	0.00	<5.5	100		<7.0		Field Blank

Rey Cross
Analyst

José Quiroga
Laboratory Manager
or other approved signatory

Limit of detection is 7 fibers/mm². The laboratory is not responsible for data reported in fibers/cc, which is dependent on volume collected by non-laboratory personnel. This report relates only to the samples reported above. This report may not be reproduced, except in full, without written approval by EMSL.
Analysis performed by EMSL Manhattan (NY State ELAP #11506, AIHA #102581)

PCM-2

CITY_21MC100_012558160

NYC-WTC_000174576