

EMSL Analytical, Inc.

307 West 38th Street, New York, NY 10018

Phone: (212) 290-0051 Fax: (212) 290-0058 Email: manhattanlab@emsl.com



Attn: TRC Environmental Consultants
1430 Broadway
10th Floor
New York, NY 10018

Fax: (212) 221-7840 Phone: 212-221-7822

Project: 35052-0000-00017 / New York City Law Department / 100
Church St. / Ambient

Customer ID: TRCE51

Customer PO:

Received: 03/08/03 12:18 PM

EMSL Order: 030303661

EMSL Project ID:

Analysis Date: 3/8/2003

**Phase Contrast Microscopy (PCM) Fiber Count by NIOSH 7400, Fourth Edition, Issue 2,
8/15/94**

Sample	Location	Sample Date	Volume	Fibers	Fields	LOD (fib/cc)	Fibers/ mm ²	Fibers/ cc	Notes
NA030703ASBAI R2001	20th Floor	3/7/2003	2019.00	<5.5	100	0.001	<7.0	<0.001	
030303661-0001									
NA030703ASBAI R2002	20th Floor	3/7/2003	1615.00	<5.5	100	0.002	<7.0	<0.002	
030303661-0002									
NA030703ASBAI R2003	20th Floor - Blank	3/7/2003	0.00	<5.5	100		<7.0		Field Blank
030303661-0003									

George Bobie

Analyst

Laboratory Manager
or other approved signatory

Limit of detection is 7 fibers/mm². The laboratory is not responsible for data reported in fiber/cc, which is dependant on volume collected by non-laboratory personnel. This report relates only to the samples reported above. This report may not be reproduced, except in full, without written approval by EMSL.
Analysis performed by EMSL Manhattan (NY State ELAP #11506, AIHA #102581)

PCM-2

1