


TEM Asbestos (air) Count Sheet

sample area
analyzed



Client Name: HP ENVIRONMENTAL, INC.			Volume (liters): 0.0
Job #: 101101001			Filter Type / Filter Area: MCE 385 mm ²
Lab Sample #: 08			Grid Opening Size: 0.01003
Client Sample #: ET-TRL-0929-FB2			Area Examined: 0.10030 mm ²
Received: 10/01/2001		02:00:17	Magnification: 20,000
Date Analyzed: 10/01/2001			Accelerating Voltage: 100 KeV
Scope #: 4			

Analysis Performed by: Thomas R. McKee
Thomas R. McKee

Location	Grid Opening	Fiber	Length μ M	Width μ M	Fiber Type	Morphology	EDS	Orient.	SAED	Photo
D7-C3/1	1	NSD							☐	
D7-E3/1	2	NSD							☐	
D7-F3/1	3	NSD							☐	
D7-G3/1	4	NSD							☐	
D7-H3/1	5	NSD							☐	
D6-C2/6	6	NSD							☐	
D6-E2/6	7	NSD							☐	
D6-F2/6	8	NSD							☐	
D6-G2/6	9	NSD							☐	
D6-H2/6	10	NSD							☐	

NSD: No Asbestos Structures Detected

Comments

	Structure Density (str/mm ²)	Concentration (str/cc air)	Grid Evaluation
Total Grid Openings: 10			
Chrysotile Asbestos Structures: 0	<10		☒ Grid Openings Covered > 50%
Amphibole Asbestos Structures: 0	<10		☒ Intact Grid Opening > 50%
Asbestos Structures $\geq$ 5 microns: 0	<10		☒ Undissolved Filter < 10%
Total Non-Asbestos Structures:			☒ Folded Replica < 50%
Total Asbestos Structures: 0	<10		☒ Filter Loading < 10%
Analytical Sensitivity: 10		.0000	☒ Particulate Even

06.002252

CONFIDENTIAL,

CITY_21MC100_013442374

SUBJECT TO PROTECTIVE ORDER NO. 2, SEPT. 2008