



INDUSTRIAL HYGIENE

ENVIRONMENTAL TESTING

• EPANVLAP 101262-0
• AIHA ACCREDITATION NO. 100439
• NC DENR 589

• NY DOH/NELAC 10903
• PA DEP 06-353

• NJ DEP PA020
• CT DPH PH-0239

ANALYTICAL SERVICES

Date: 21-May-02

CLIENT: ATC Associates
Project: 19125-0246 PCB Wipes

Lab Order: R02050509

Lab ID: R02050509-01
Client Sample ID: PCB1 ATC-0508-STY

Collection Date: 5/8/2002
Matrix: WIPE

Analyses	Result	Limit	Qual	Units	DF	Date Analyzed
POLYCHLORINATED BIPHENYLS		SW8082B		Analyst: TH		
Aroclor 1016	BRL	2.0		µg, Total	1	5/21/2002 12:29:00 PM
Aroclor 1221	BRL	2.0		µg, Total	1	5/21/2002 12:29:00 PM
Aroclor 1232	BRL	2.0		µg, Total	1	5/21/2002 12:29:00 PM
Aroclor 1242	BRL	2.0		µg, Total	1	5/21/2002 12:29:00 PM
Aroclor 1248	BRL	2.0		µg, Total	1	5/21/2002 12:29:00 PM
Aroclor 1254	BRL	2.0		µg, Total	1	5/21/2002 12:29:00 PM
Aroclor 1260	BRL	2.0		µg, Total	1	5/21/2002 12:29:00 PM

Lab ID: R02050509-02
Client Sample ID: PCB2 ATC-0508-STY

Collection Date: 5/8/2002
Matrix: WIPE

Analyses	Result	Limit	Qual	Units	DF	Date Analyzed
POLYCHLORINATED BIPHENYLS		SW8082B		Analyst: TH		
Aroclor 1016	BRL	2.0		µg, Total	1	5/21/2002 1:02:00 AM
Aroclor 1221	BRL	2.0		µg, Total	1	5/21/2002 1:02:00 AM
Aroclor 1232	BRL	2.0		µg, Total	1	5/21/2002 1:02:00 AM
Aroclor 1242	BRL	2.0		µg, Total	1	5/21/2002 1:02:00 AM
Aroclor 1248	BRL	2.0		µg, Total	1	5/21/2002 1:02:00 AM
Aroclor 1254	BRL	2.0		µg, Total	1	5/21/2002 1:02:00 AM
Aroclor 1260	BRL	2.0		µg, Total	1	5/21/2002 1:02:00 AM

Lab ID: R02050509-03
Client Sample ID: PCB3 ATC-0508-STY

Collection Date: 5/8/2002
Matrix: WIPE

Analyses	Result	Limit	Qual	Units	DF	Date Analyzed
POLYCHLORINATED BIPHENYLS		SW8082B		Analyst: TH		
Aroclor 1016	BRL	2.0		µg, Total	1	5/21/2002 1:36:00 AM
Aroclor 1221	BRL	2.0		µg, Total	1	5/21/2002 1:36:00 AM
Aroclor 1232	BRL	2.0		µg, Total	1	5/21/2002 1:36:00 AM
Aroclor 1242	BRL	2.0		µg, Total	1	5/21/2002 1:36:00 AM
Aroclor 1248	BRL	2.0		µg, Total	1	5/21/2002 1:36:00 AM
Aroclor 1254	BRL	2.0		µg, Total	1	5/21/2002 1:36:00 AM
Aroclor 1260	BRL	2.0		µg, Total	1	5/21/2002 1:36:00 AM

Qualifiers:
BRL - Below Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level

S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range

Page 1 of 1

CONFIDENTIAL,

CITY_21MC100_003432451

SUBJECT TO PROTECTIVE ORDER NO. 2, SEPT. 2008