

12/12/2001 06:33 212-290-2538

EMSL ANALYTICAL INC

New York Lab

307 West 38th Street, New York, NY 10018

Phone: 212-290-0051 Fax: 212-290-0058 Email:

EMSL

Attn: U.S. EPA - Residential
HQ Financial Services Division
1200 Pennsylvania Ave. NW Maildrop 2734R
Washington, DC 20460

Fax: Phone:

Project: USPHS MISSION #12 RESIDENTIAL SAMPLING/ BLDG
#32/161 EAST 88TH ST

Customer ID: USEPA44
Customer PO: 2W-1103-NNLX
Received: 12/11/01 4:44 PM

EMSL Order: 030109589
EMSL Project ID:

Analysis Date: 12/12/2001

Fiber Analysis of Air Samples via NIOSH 7400, Revision 3, Issue 2, 8/15/94

Sample	Location	Sample Date	Volume	Fibers	Fields	LOD (fib/cc)	Fibers/ mm ³	Fibers/ cc	Notes
[R]A-R1-X-32	RESIDENCE 1	12/8/2001	3674.34	8.0	100	0.001	10.19	0.001	
030109589-0001									

Charmel Dozier

Analyst

Laboratory Manager
or other approved signatory

Limit of detection is 7 fibers/mm³. The laboratory is not responsible for data reported in fibers/cc, which is dependent on volume collected by non-laboratory personnel. This report relates only to the samples reported above. This report may not be reproduced, except in full, without written approval by EMSL.

Analysis performed by EMSL Manhattan (NY State ELAP #11506)

1CM-1

1

CONFIDENTIAL,

CITY_21MC100_003423402

SUBJECT TO PROTECTIVE ORDER NO. 2, SEPT. 2008