NYC 9/11 Public Portal Document

New York Lab
307 West 38th Street, New York, NY 10018
Phone: 212-290-0051 Fax: 2122900058 Email:

Attn: U.S. EPA - Residential
HQ Financial Services Division
1200 Pennsylvania Ave. NW Maildrop 2734R
Washington, DC 20460

Fax: Phone:

Project: USPHS MISSION 12 Residential Sampling 66 Frankfurt

Customer ID: USEPA44
Customer PO: 2W-1103-NNLX
Received: 11/20/01 7:30 PM

EMSL Order: 030108405
EMSL Project ID:

Analysis Date: 11/21/01

Asbestos Analysis of Bulk Materials by PLM via the NY State ELAP 198.1 Method

Sample	Location	Appearance	Treatment	Non-Asbestos %	Fibrous	Non-Asbestos %	Non-Fibrous	Asbestos % Type
[R]VA-C-11 *030108405-0001*	COMMON AREA	White/Tan Fibrous Heterogeneous	Teased Dissolved	5.00%	Cellulose	95.00%	Non-fibrous (other)	None Detected
[R]VA-R1-11 *030108405-0002*	RESIDENCE 1	Tan Fibrous Heterogeneous	Teased Dissolved	3.00% 5.00%	Synthetic Cellulose	92.00%	Non-fibrous (other)	None Detected
[R]VA-R2-11 *030108405-0003*	RESIDENCE 2	Tan Fibrous Heterogeneous	Teased Dissolved	15.00% 5.00%	Synthetic Cellulose	80.00%	Non-fibrous (other)	None Detected

Dan Osborne
Analyst

Laboratory Manager
or other approved signatory

PLM has been known to miss asbestos in a small percentage of samples which contain asbestos. Negative PLM results cannot be guaranteed. Samples reported as <1% or none detected should be tested with TEM. The above test report relates only to the items tested. This report may not be reproduced, except in full, without written approval by EMSL Analytical, Inc. The above test must not be used by the client to claim product endorsement by NVLAP nor any agency of the United States Government.
nalysis performed by EMSL Manhattan (NVLAP #101048-0), NY ELAP 11508

PLM PointCount-1

CONFIDENTIAL, CITY_21MC100_003423500
SUBJECT TO PROTECTIVE ORDER NO. 2, SEPT. 2008

NYC-WTC_000176103