

TEM Asbestos (air) Count Sheet

<u>Client Name:</u>	NYC BOE	<u>Volume (liters):</u>	1840
<u>Lab Group #:</u>	7540	<u>Filter Type / Filter Area:</u>	MCE 385 mm ²
<u>Lab Sample #:</u>	9	<u>Grid Opening Size:</u>	0.01124
<u>Client Sample #:</u>	09	<u>Area Examined:</u>	0.04496 mm ²
<u>GRID BOX:</u>	1025	<u>Magnification:</u>	20,000
<u>GRID ID:</u>	2640	<u>Accelerating Voltage:</u>	100 KeV
<u># Grid Openings:</u>	4	<u>Analysis Date:</u>	4/27/02

A handwritten signature consisting of stylized cursive letters, possibly reading "M. J." or similar, followed by a long horizontal flourish. The signature is written above a solid black horizontal line.

MARK PEYSAKHOV

204	1 N3	N50
—	2 V3	—
20	3 P1	—
	4 V1	—

Struct = $> 0.5\mu$ $> 5:1$ Aspect Ratio

√ = Positive diffraction pattern properties coniforming to laboratory identification criteria observed

✓ = EDS spectrum conforming to laboratory identification criteria observed

F = Single Fiber

B = Bundle (<3 fibers <1 fiber diameter apart)

C = Cluster (>3 Intersections)

M = Matrix

NSD = No Structures Detected

Grid acceptance:

<50% undissolved filter, overlapped or folded
<10% particulate coverage >50% coverage at
least 20 unfolded or overlapped grid squares <5%
holes or opaque areas

AMPHIBOLES

Amosite

Crocidolite

Anthrophyllite

Tremolite

Actinolite

Totals

	F	B	C	M	>50
Chrysotile					
Amphibole					
Ambiguous					
Non-Asbestos					
Non-Identified					

CONFIDENTIAL,

CITY 21MC100 003471301

SUBJECT TO PROTECTIVE ORDER NO. 2, SEPT. 2008