TEM Asbestos (air) Count Sheet

Client Name:	NYC BOE	Volume (liters):	1960
Lab Group #:	7412	Filter Type / Filter Area:	MCE 385 mm²
Lab Sample #:	14	Grid Opening Size:	0.01124
Client Sample #:	14	Area Examined:	0.04496 mm²
GRID BOX:	1036	Magnification:	20,000
GRID ID:	E4 H0	Accelerating Voltage:	100 KeV
# Grid Openings:	4	Analysis Date:	4/14/02

Analysis Performed By: MARK PEYSAKHOV

GO		Structure No	Structure Type				Length		ID of Structure N or √				EDS (Nonadb Elements)
No	Coord		F	B	C	M	>=05 µ <5µ	>=5µ	Chrys	Amphibole	Nonasb	Ambig/ND	
1	C7	NSD											
2	H7	— "											
3	D1	— "											
4	F C1	— "											

Struct = > 0.5µ > 5:1 Aspect Ratio

√= Positive diffraction pattern properties conforming to laboratory identification criteria observed

√= EDS spectrum conforming to laboratory identification criteria observed

F = Single Fiber
B = Bundle (<3 fibers <1 fiber diameter apart
C = Cluster (>3 Intersections)
M = Matrix
NSD = No Structures Detected

Grid acceptance:

<50% undissolved filter, overlapped or folded <10% particulate coverage >50% coverage ate least 20 unfolded or overlapped grid squares <5% holes or opaque areas

AMPHIBOLES
Amosite
Crocidolite
Anthophyllite
Tremolite
Actinolite

Totals

	F	B	C	M	>5µ
Chrysotile					
Amphibole					
Ambiguous					
Non-Asbestos					
Non-Identified					