

ATC Associates

104 East 25 Street, New York, NY 10010

Phone: (212) 353-8280

Fax: (212) 353-8306



Attn: Benjamin Sallera
ATC Associates Inc.
104 East 25th Street
New York

NY 10010

Fax: (212) 353-3599

Phone: (212) 353-8280

Project: DEP Clean-up ATC Project #: 22906-0002

125 Cedar Street, New York, NY

Received: 10/15/02 9:25:00 PM

ATC Group #: 8991

Analysis Date: 10/16/02

Basement

EPA PCMe Method Via Transmission Electron Microscopy

Sample	Volume	Non Asbestos Fibers	PCM F/cc	Asbestos Type(s)	Asbestos Fibers	Asbestos % of total	EPA PCMe (TEM) F/cc	Notes
03 8991 -1	3700.00	10	0.001	None Detected	0	0	<0.0002	Cellulose - 6 Structures, Gypsum - 4 Structures
04 8991 -2	3690.00	9	0.001	None Detected	0	0	<0.0002	Cellulose - 6 Structures, Gypsum - 3 Structures
05 8991 -3	3680.00	12	0.001	None Detected	0	0	<0.0002	Cellulose - 5 Structures, Gypsum - 7 Structures
06 8991 -4	3670.00	24	0.001	None Detected	0	0	<0.0002	Cellulose - 8 Structures, Gypsum - 16 Structures
07 8991 -5	3660.00	26	0.001	None Detected	0	0	<0.0002	Cellulose - 4 Structures, Gypsum - 20 Structures
08 8991 -6	3650.00	10	0.001	None Detected	0	0	<0.0002	Cellulose - 7 Structures, Gypsum - 3 Structures
09 8991 -7	3640.00	10	0.001	None Detected	0	0	<0.0002	Cellulose - 7 Structures, Gypsum - 3 Structures

ROMAN PEYSAKHOV

Analyzed by:

RPey

MICHAEL A. GORYCKI, Ph.D.

Approved by:

ATC is not responsible for data reported in fibers/cc, which is dependent on volume collected by non-laboratory personnel. The above report relates only to the items tested. This report may not be reproduced, except in full, without written approval by ATC Associates Inc.

Analysis performed by ATC Associates Inc. (NY ELAP #10879).

Forty Grid Openings from each of the samples were examined for structures greater than 5µm length and having a length-to-width ratio greater than 3:1.

Wednesday, October 16, 2002

Page 1 of 2

CITY_21MC100_009369714

Attn: Benjamin Sallento
ATC Associates Inc.
104 East 25th Street
New York NY 10010

Received: 10/15/02 9:25:00 PM
ATC Group #: 8991
Analysis Date: 10/16/02

Fax: (212) 353-3599 Phone: (212) 353-8280


Project: DEP Clean-up ATC Project # 22906-0002
125 Cedar Street, New York, NY

EPA PCMe Method Via Transmission Electron Microscopy

Sample	Volume	Non Asbestos Fibers	PCM F/cc	Asbestos Type(s)	Asbestos Fibers	Asbestos % of total	EPA PCMe (TEM) F/cc	Notes
10 8991 -R	3630.00	12	0.001	None Detected	0	0	<0.0002	Cellulose - 7 Structures, Gypsum - 5 Structures

ROMAN PEYSAKHOV

Analyzed by:



MICHAEL A. GORYCKI, Ph.D.

Approved by:

ATC is not responsible for data reported in fibers/cc, which is dependent on volume collected by non-laboratory personnel. The above report relates only to the items tested. This report may not be reproduced, except in full, without written approval by ATC Associates Inc.

Analysis performed by ATC Associates Inc (NY ELAP #10873).

Fifty Grid Openings from each of the samples were examined for structures greater than 5µm length and having a length-to-width ratio greater than 3:1

Wednesday, October 16, 2002

Page 2 of 2

CITY_21MC100_009369715