NYC 9/11 Public Portal Document

TEM Asbestos (air) Count Sheet

Client Name:	NYC BOE	**Volume (liters):**	1944
Lab Group #:	6229	**Filter Type / Filter Area:**	MCE 385 mm²
Lab Sample #:	17	**Grid Opening Size:**	0.01124
Client Sample #:	17	**Area Examined:**	0.04496 mm²
GRID BOX:	7201	**Magnification:**	20.000
GRID ID:	149	**Accelerating Voltage:**	100 KeV
# Grid Openings:	4	**Analysis Date:**	12/30/01

Analysis Performed By: RP
ROMAN PEYSAKHOV

GO No	GO Coord	Structure No	Structure Type F	B	C	M	Length >=.05 µ <5µ	Length >=5µ	ID of Structure N or Y Chrys	Amphibole	Nonasb	Ambig/ND	EDS (Nonadb Elements)
1A 1	J2	NSD											
2	P2	4											
3	P2	11											
4	Q2	11											
1B													

Struct = > 0.5µ > 5:1 Aspect Ratio

Y = Positive diffraction pattern properties conforming to laboratory identification criteria observed

Y = EDS spectrum conforming to laboratory identification criteria observed

F = Single Fiber
B = Bundle (<3 fibers <1 fiber diameter apart
C = Cluster (>3 Intersections)
M = Matrix
NSD = No Structures Detected

Grid acceptance:

<50% undissolved filter, overlapped or folded
<10% particulate coverage >50% coverage ate least 20 unfolded or overlapped grid squares <5% holes or opaque areas

AMPHIBOLES
Amosite
Crocidolite
Anthophyllite
Tremolite
Actinolite

Totals

	F	B	C	M	>5µ
Chrysotile					
Amphibole					
Ambiguous					
Non-Asbestos					
Non-Identified					

NYC-WTC_000184496