NYC 9/11 Public Portal Document

TEM Asbestos (air) Count Sheet

Client Name:	NYC BOE	**Volume (liters):**	1960
Lab Group #:	5904	**Filter Type / Filter Area:**	MCE 385 mm²
Lab Sample #:	21	**Grid Opening Size:**	0.01124
Client Sample #:	21 -	**Area Examined:**	0.04496 mm²
GRID BOX:	546	**Magnification:**	20.000
GRID ID:	2DE	**Accelerating Voltage:**	100 KeV
# Grid Openings:	4	**Analysis Date:**	11/17/01

Analysis Performed By: MG
MICHAEL GORYCKI

GO No	GO Coord	Structure No	Structure Type F	B	C	M	Length >=05 µ <5µ	>=5µ	ID of Structure N or √ Chrys	Amphibole	Nonasb	Ambig/ND	EDS (Nonadb Elements)
1	D9	NSD											
2	C9	"											
3	P9	"											
4	V9	"											

Struct = > 0.5µ > 5:1 Aspect Ratio

√= Positive diffraction pattern properties conforming to laboratory identification criteria observed

√= EDS spectrum conforming to laboratory identification criteria observed

F = Single Fiber
B = Bundle (<3 fibers <1 fiber diameter apart
C = Cluster (>3 Intersections)
M = Matrix
NSD = No Structures Detected

Grid acceptance:

<50% undissolved filter, overlapped or folded <10% particulate coverage >50% coverage ate least 20 unfolded or overlapped grid squares <5% holes or opaque areas

AMPHIBOLES
Amosite
Crocidolite
Anthophyllite
Tremolite
Actinolite

Totals

	F	B	C	M	>5µ
Chrysotile					
Amphibole					
Ambiguous					
Non-Asbestos					
Non-Identified					

NYC-WTC_000185410