

Tel. (718) 784-7490
Fax (718) 784-4085

TEM AHERA COUNTSHEET

CLIENT: <i>wamen & powder</i>	DATE: <i>2/4/03</i>
PROJECT: <i>MCDEP/15 Park Row</i>	ANALYST: <i>ES</i>
BATCH #: <i>T03-02-012</i>	CLIENT ID #: <i>04</i>
SAMPLE ID #: <i>T03-02-012-179</i>	VOLUME (lit): <i>1800</i>
G.O. AREA (mm ²): <i>0.012820</i>	GO's TO ANALYZE: <i>4</i>
MAGNIFICATION: <i>194</i>	ACC. VOLTAGE (KeV): <i>100</i>
INSTRUMENT: JEOL 100CX II	FILTER AREA (mm ²): <i>385</i>
ANALYTICAL METHODOLOGY: EPA FEDERAL REGISTER, VOL. 52, No 210, OCT. 30, 1987, #40 CFR, PART 763, APPENDIX A TO SUBPART E	

[illegible]

Analytical Sensitivity (str/cm ²):	0.0042	Eucentricity check:	
Concentration (str/cm ²):	40.0042	*Grid acceptance criteria check:	(A)..... (B).....
Concentration (str/mm ²):	419.50	**CHR. # STRUCT.:	F M C B
Number of Asbestos Structures:	0.5 ≤ S < 5 μm 0... S ≥ 5 μm 0...	**AMPH. # STRUCT.:	F M C B
		Specify AMPH. Type:	AM. ACT. TRE.
Number of Non-asbestos Structures:	0...	CRO. ANTH.	

*Replica covers >50%, at least 20 GO's with no overlapping or folded replica, intact GO's have >50% coverage, <50% Undissolved filter, <5% holes or opaque areas, <10% particulate coverage.

** CHR.=chrysotile, AMPH.=amphibole, AM.=amosite, ACT.=actinolite, TRE.=tremolite, CRO.=crocidolite, ANTH.=anthophyllite.