



ATHENICA

ENVIRONMENTAL SERVICES INC.

45-09 Greenpoint Avenue
Long Island City, N.Y. 11104

TEM DEPARTMENT

Tel. (718) 784-7490
Fax (718) 784-4085

TEM AHERA COUNTSHEET

CLIENT: WARREN & PANZER	DATE: 1/27/03
PROJECT: MYCDEP/15 PARK ROW	ANALYST: EMD
BATCH #: TD3-01-022	CLIENT ID #: 09
SAMPLE ID #: TD3-01-022-097	VOLUME (lt): 1719
G.O. AREA (mm ²): 0.012820	GO's TO ANALYZE: 4
MAGNIFICATION: 19K	ACC. VOLTAGE (KeV): 100
INSTRUMENT: JEOL 100CX II	FILTER AREA (mm ²): 285
ANALYTICAL METHODOLOGY: EPA FEDERAL REGISTER, VOL. 52, No 210, OCT. 30, 1987, #40 CFR, PART 763, APPENDIX A TO SUBPART E	

[illegible]

Analytical Sensitivity (str/cm ³): 0.0044	Eucentricity check: ✓
Concentration (str/cm ³): 0.0044	*Grid acceptance criteria check: (A)..... (B).....
Concentration (str/mm ²): 19.50	**CHR. # STRUCT.: F M C B
Number of Asbestos Structures: 0.5 ≤ S < 5 μm 0...	**AMPH. # STRUCT.: F 0 M 0 C 0 B 0
S ≥ 5 μm 01..	Specify AMPH. Type: AM. ACT. (IRE) 01
Number of Non-asbestos Structures: 0...	CRO. ANTH.

* Replica covers >50%, at least 20 GO's with no overlapping or folded replica, intact GO's have >50% coverage, <50% Undissolved filter, <5% holes or opaque areas, <10% particulate coverage.

** CHR.=chrysotile, AMPH.=amphibole, AM.=amosite, ACT.=actinolite, TRE.=tremolite, CRO.=crocidolite, ANTH.=anthophyllite.

CITY_21MC100_002964458