

Tel. (718) 784-7480
Fax (718) 784-4085

CLIENT: WARREN & PANZER	DATE: 10/15/02
PROJECT: MCDEP/62 WILLIAMS ST.	ANALYST: END
BATCH #: T02-10-097	CLIENT ID #: 02
SAMPLE ID #: T02-10-097-3172	VOLUME (l): 1320
G.O. AREA (mm ²): 0.013838	GO's TO ANALYZE: 5
MAGNIFICATION: 19K	ACC. VOLTAGE (KeV): 100
INSTRUMENT: JEOL 100CX II	FILTER AREA (mm ²): 385
ANALYTICAL METHODOLOGY: EPA FEDERAL REGISTER, VOL. 52, No 210, OCT. 30, 1987, #40 CFR, PART 763, APPENDIX A TO SUBPART E	

[illegible]

Analytical Sensitivity (str/cm ²):	0.0042	Eucentricity check:	
Concentration (str/cm ²):	< 0.0042	*Grid acceptance criteria check:	(A)..... (B).....
Concentration (str/mm ²):	< 14.45	**CHR. # STRUCT.:	F M C B
Number of Asbestos Structures:	0.5 ≤ S < 5 μm 0	**AMPH. # STRUCT.:	F M C B
	S ≥ 5 μm 0	Specify AMPH. Type:	AM ACT TRE
Number of Non-asbestos Structures:	0	CRO. ANIH.	

*Replica covers >50%, at least 20 GO's with no overlapping or folded replica, intact GO's have >50% coverage, <50% Undissolved filler, <5% holes or opaque areas, <10% particulate coverage.

** CHR.=chrysotile, AMPH.=amphibole, AM.=amosite, ACT.=actinolite, TRE.=tremolite, CRO.=crocidolite, ANTH.=anthophyllite.