

ENVIRONMENTAL SERVICES INC.

Tel. (716) 784-7480
Fax (716) 784-4085

TEM AHERA COUNTSHEET

CLIENT: WARREN & PANZER	DATE: 10/23/02
PROJECT: MCDER/62 WILLIAM ST.	ANALYST: ES
BATCH #: T02-10-154	CLIENT ID #: 3
SAMPLE ID #: T02-10-154-3475	VOLUME (lt): 1300
G.O. AREA (mm ²): 0.013838	GO's TO ANALYZE: 5
MAGNIFICATION: 19K	ACC. VOLTAGE (KeV): 100
INSTRUMENT: JEOL 100CX II	FILTER AREA (mm ²): 385
ANALYTICAL METHODOLOGY: EPA FEDERAL REGISTER, VOL. 52, No 210, OCT. 30, 1987, #40 CFR, PART 763, APPENDIX A TO SUBPART E	

[illegible]

Analytical Sensitivity (str/cm ³):	0.0043	Eucentricity check :	
Concentration (str/cm ³) :	20.0043	*Grid acceptance criteria check :	(A)...✓ (B)...✓
Concentration (str/mm ²) :	<14.45	**CHR. # STRUCT.:	F M C B
Number of Asbestos Structures :	0.5 ≤ S < 5 μm ... 0..	**AMPH. # STRUCT.:	F M C B
	S ≥ 5 μm 0..	Specify AMPH. Type :	AM. ACT. TRE.
Number of Non-asbestos Structures:	 0..	CRO. ANTH.	

** CHR.=chrysotile, AMPH.=amphibole, AM.=amosite, ACT.=actinolite, TRE.=tremolite, CRO.=crocidolite, ANTH.=anthophyllite