

Tel. (718) 784-7480
Fax (718) 784-4085

CLIENT: Warren & Pauer	DATE: 10/14/02
PROJECT: MCDEP/123 Williams St	ANALYST: END
BATCH #: T02-10-083	CLIENT ID #: 04
SAMPLE ID #: T02-10-083 - 3104	VOLUME (lit): 1620
G.O. AREA (mm ²): 0.013838	GO's TO ANALYZE: 4
MAGNIFICATION: 194	ACC. VOLTAGE (KeV): 100
INSTRUMENT: JEOL 100CX II	FILTER AREA (mm ²): 385
ANALYTICAL METHODOLOGY: EPA FEDERAL REGISTER, VOL. 52, No 210, OCT. 30, 1987, #40 CFR, PART 763, APPENDIX A TO SUBPART E	

[illegible]

Analytical Sensitivity (str/cm ²):	0.0043	Eucentricity check:	
Concentration (str/cm ²):	<0.0043	*Grid acceptance criteria check:	(A)✓... (B)....✓
Concentration (str/mm ²):	<18.07	**CHR. # STRUCT.:	F M C B
Number of Asbestos Structures:	0.5 ≤ S < 5 μm ... 0...	**AMPH. # STRUCT.:	F M C B
	S ≥ 5 μm 0...	Specify AMPH. Type:	AM ACT TRE
Number of Non-asbestos Structures:	 0.....	CRO. ANTH.	

* Replica covers >50%, at least 20 GO's with no overlapping or folded replica, intact GO's have >50% coverage, <50% Undissolved filter, <5% holes or opaque areas, <10% particulate coverage.